

LB120399


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

52890



A= 0.0000e+000

B= 1.9036e-004

C= 3.0413e-002

Rho= 0.9996922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.022	0.022	-42	0.000	-42					-
0.2	0.200	0.235	0.035	1073	0.0 %	1073					18
2.5	2.500	2.566	0.066	13319	0.0 %	13319					3
5.0	5.000	4.819	-0.181	25153	0.0 %	25153					-4
7.5	7.500	7.460	-0.040	39029	0.0 %	39029					-1
10.0	10.000	10.099	0.099	52890	0.0 %	52890					1

LB120399 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 May 2022 07:51:14

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-42	-	0.0000	7471B	PPB	28 May 2022 07:55:56	S	Std
0.2 - 1	50.2	1073	-	0.2000	7471B	PPB	28 May 2022 07:58:13	S	Std
2.5 - 1	52.5	13319	-	2.5000	7471B	PPB	28 May 2022 08:00:30	S	Std
5.0 - 1	55	25153	-	5.0000	7471B	PPB	28 May 2022 08:02:49	S	Std
7.5 - 1	57.5	39029	-	7.5000	7471B	PPB	28 May 2022 08:07:39	S	Std
10.0 - 1	111	52890	-	10.0000	7471B	PPB	28 May 2022 08:09:54	S	Std
ICV49 - 1	ICV49	21042	4.0360	-	7471B	PPB	28 May 2022 08:12:44	U	SMPL
ICB49 - 1	ICB49	-68	0.0175	-	7471B	PPB	28 May 2022 08:14:58	U	SMPL
CCV76 - 1	CCV76	25900	4.9607	-	7471B	PPB	28 May 2022 08:17:15	U	SMPL
CCB76 - 1	CCB76	-165	-0.0010	-	7471B	PPB	28 May 2022 08:19:30	U	SMPL
CRA - 1	CRA	970	0.2151	-	7471B	PPB	28 May 2022 08:21:46	U	SMPL
HighStd - 1	HighStd	52937	10.1075	-	7471B	PPB	28 May 2022 08:24:01	U	SMPL
ChkStd - 1	ChkStd	36051	6.8931	-	7471B	PPB	28 May 2022 08:26:16	U	SMPL
PB145165BL - 1	PBS	-76	0.0159	-	7471B	PPB	28 May 2022 08:28:34	U	SMPL
PB145165BS - 1	LCSS	21458	4.1152	-	7471B	PPB	28 May 2022 08:33:52	U	SMPL
N3032-01 - 1	SU-406-ROLL-OFF	2118	0.4336	-	7471B	PPB	28 May 2022 08:36:08	U	SMPL
N3041-01 - 1	2147	338124	64.3958	-	7471B	PPB	28 May 2022 08:38:24	U	SMPL
N3042-05 - 1	303	-322	-0.0309	-	7471B	PPB	28 May 2022 08:40:41	U	SMPL
N3051-02 - 1	72-11944	14772	2.8424	-	7471B	PPB	28 May 2022 08:43:08	U	SMPL
N3053-01 - 1	52322-OIL	5565	1.0898	-	7471B	PPB	28 May 2022 08:45:23	U	SMPL
CCV77 - 1	CCV77	25444	4.8739	-	7471B	PPB	28 May 2022 09:01:21	U	SMPL
CCB77 - 1	CCB77	-57	0.0196	-	7471B	PPB	28 May 2022 09:03:36	U	SMPL
N3055-03 - 1	ETGI-329	181151	34.5144	-	7471B	PPB	28 May 2022 09:05:53	U	SMPL
N3057-01 - 1	72-1193	4447	0.8769	-	7471B	PPB	28 May 2022 09:08:08	U	SMPL
N3062-01 - 1	FS-SOIL-0526	1811	0.3752	-	7471B	PPB	28 May 2022 09:10:32	U	SMPL
N3067-01 - 1	33686	3632	0.7218	-	7471B	PPB	28 May 2022 09:12:47	U	SMPL
N3070-01 - 1	SC1	14650	2.8192	-	7471B	PPB	28 May 2022 09:15:03	U	SMPL
N3072-01 - 1	RT-4206	3660	0.7271	-	7471B	PPB	28 May 2022 09:17:19	U	SMPL
N3072-03 - 1	PSC-124035	29040	5.5585	-	7471B	PPB	28 May 2022 09:19:36	U	SMPL
N3072-05 - 1	PSC-124030	2344	0.4766	-	7471B	PPB	28 May 2022 09:21:53	U	SMPL
N3076-02 - 1	RBR-671	824	0.1873	-	7471B	PPB	28 May 2022 09:24:11	U	SMPL
N3077-01 - 1	CL-02-052622	7324	1.4246	-	7471B	PPB	28 May 2022 09:26:27	U	SMPL
CCV78 - 1	CCV78	26324	5.0415	-	7471B	PPB	28 May 2022 09:31:23	U	SMPL
CCB78 - 1	CCB78	-108	0.0099	-	7471B	PPB	28 May 2022 09:33:39	U	SMPL
N3080-01 - 1	WC-4	1761	0.3656	-	7471B	PPB	28 May 2022 09:35:56	U	SMPL
N3080-01DUP - 1	WC-4DUP	1613	0.3375	-	7471B	PPB	28 May 2022 09:38:11	U	SMPL
N3080-01MS - 1	WC-4MS	20414	3.9164	-	7471B	PPB	28 May 2022 09:43:50	U	SMPL
N3080-01MSD - 1	WC-4MSD	20254	3.8860	-	7471B	PPB	28 May 2022 09:46:05	U	SMPL
N3080-04 - 1	WC-5	1568	0.3289	-	7471B	PPB	28 May 2022 09:48:22	U	SMPL
PB145166BL - 1	PBS	8	0.0319	-	7471B	PPB	28 May 2022 09:50:40	U	SMPL
PB145166BS - 1	LCSS	20572	3.9465	-	7471B	PPB	28 May 2022 09:55:07	U	SMPL
N3054-01 - 1	OL-VC-80071-04	6784	1.3218	-	7471B	PPB	28 May 2022 09:57:23	U	SMPL
N3054-01DUP - 1	OL-VC-80071-04DUP	7695	1.4952	-	7471B	PPB	28 May 2022 09:59:41	U	SMPL
N3054-01MS - 1	OL-VC-80071-04MS	24648	4.7224	-	7471B	PPB	28 May 2022 10:01:57	U	SMPL
CCV79 - 1	CCV79	26091	4.9971	-	7471B	PPB	28 May 2022 10:04:12	U	SMPL
CCB79 - 1	CCB79	-117	0.0081	-	7471B	PPB	28 May 2022 10:06:28	U	SMPL
N3054-01MSD - 1	OL-VC-80071-04MSD	26204	5.0186	-	7471B	PPB	28 May 2022 10:08:45	U	SMPL
N3054-02 - 1	OL-VC-80071-410	22144	4.2457	-	7471B	PPB	28 May 2022 10:11:00	U	SMPL
N3054-03 - 1	OL-VC-80040-04	4012	0.7941	-	7471B	PPB	28 May 2022 10:13:17	U	SMPL
N3054-04 - 1	OL-VC-80040-410	5804	1.1353	-	7471B	PPB	28 May 2022 10:15:34	U	SMPL
N3054-05 - 1	OL-STA-80238-04	7967	1.5470	-	7471B	PPB	28 May 2022 10:17:50	U	SMPL
N3054-06 - 1	OL-STA-80238-410	40004	7.6456	-	7471B	PPB	28 May 2022 10:20:06	U	SMPL
N3054-07 - 1	ST-51-04	10476	2.0246	-	7471B	PPB	28 May 2022 10:22:22	U	SMPL
N3054-08 - 1	ST-51-410	19652	3.7714	-	7471B	PPB	28 May 2022 10:24:41	U	SMPL
N3054-09 - 1	OL-VC-80079-04	8186	1.5887	-	7471B	PPB	28 May 2022 10:26:57	U	SMPL
N3054-10 - 1	OL-VC-80079-410	15180	2.9201	-	7471B	PPB	28 May 2022 10:29:14	U	SMPL
CCV80 - 1	CCV80	26393	5.0546	-	7471B	PPB	28 May 2022 10:31:30	U	SMPL
CCB80 - 1	CCB80	-93	0.0127	-	7471B	PPB	28 May 2022 10:33:46	U	SMPL
N3080-01LX5 - 1		335	0.0942	-	7471B	PPB	28 May 2022 10:36:02	U	SMPL
N3080-01A - 1		21300	4.0851	-	7471B	PPB	28 May 2022 10:38:18	U	SMPL
N3054-01LX5 - 1		1150	0.2493	-	7471B	PPB	28 May 2022 10:40:33	U	SMPL
N3054-01A - 1		29340	5.6156	-	7471B	PPB	28 May 2022 10:42:50	U	SMPL
N3041-01DLX10 - 1		36952	7.0646	-	7471B	PPB	28 May 2022 10:45:06	U	SMPL
N3053-03DLX10 - 1		18681	3.5865	-	7471B	PPB	28 May 2022 10:47:23	U	SMPL
CCV81 - 1	CCV81	27239	5.2156	-	7471B	PPB	28 May 2022 10:49:42	U	SMPL
CCB81 - 1	CCB81	-152	0.0015	-	7471B	PPB	28 May 2022 10:51:59	U	SMPL