

LB120400


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41593



A= 0.0000e+000

B= 2.4290e-004

C= 2.9836e-002

Rho= 0.9997088

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.024	0.024	-22	0.000	-22					16
0.2	0.200	0.231	0.031	828	0.0 %	828					0
2.5	2.500	2.512	0.012	10220	0.0 %	10220					1
5.0	5.000	4.963	-0.037	20311	0.0 %	20311					2
7.5	7.500	7.336	-0.164	30079	0.0 %	30079					1
10.0	10.000	10.133	0.133	41593	0.0 %	41593					

LB120400

Method: 7470A

Operator: Admin

Date of Analysis: 28 May 2022 11:55:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-22	-	0.0000	7470A	PPB	28 May 2022 11:57:41	S	Std
0.2 - 1	51.2	828	-	0.2000	7470A	PPB	28 May 2022 11:59:58	S	Std
2.5 - 1	52.5	10220	-	2.5000	7470A	PPB	28 May 2022 12:02:15	S	Std
5.0 - 1	55	20311	-	5.0000	7470A	PPB	28 May 2022 12:04:32	S	Std
7.5 - 1	57.5	30079	-	7.5000	7470A	PPB	28 May 2022 12:06:48	S	Std
10.0 - 1	51	41593	-	10.0000	7470A	PPB	28 May 2022 12:09:03	S	Std
ICV50 - 1	ICV50	15186	3.7185	-7470A	PPB	PPB	28 May 2022 12:11:54	U	SMPL
ICB50 - 1	ICB50	-68	0.0133	-7470A	PPB	PPB	28 May 2022 12:14:08	U	SMPL
CCV82 - 1	CCV82	18913	4.6238	-7470A	PPB	PPB	28 May 2022 12:16:23	U	SMPL
CCB82 - 1	CCB82	-79	0.0106	-7470A	PPB	PPB	28 May 2022 12:18:39	U	SMPL
CRA - 1	CRA	772	0.2174	-7470A	PPB	PPB	28 May 2022 12:20:54	U	SMPL
HighStd - 1	HighStd	39993	9.7442	-7470A	PPB	PPB	28 May 2022 12:23:08	U	SMPL
ChkStd - 1	ChkStd	28743	7.0115	-7470A	PPB	PPB	28 May 2022 12:25:23	U	SMPL
PB145168BL - 1	PBW	-133	-0.0025	-7470A	PPB	PPB	28 May 2022 12:27:40	U	SMPL
PB145168BS - 1	LCSW	14944	3.6597	-7470A	PPB	PPB	28 May 2022 12:29:57	U	SMPL
N3070-02 - 1	SC1	-68	0.0133	-7470A	PPB	PPB	28 May 2022 12:32:13	U	SMPL
N3072-02 - 1	RT-4206	0	0.0298	-7470A	PPB	PPB	28 May 2022 12:34:30	U	SMPL
N3072-04 - 1	PSC-124035	253	0.0913	-7470A	PPB	PPB	28 May 2022 12:36:46	U	SMPL
N3072-06 - 1	PSC-124030	6	0.0313	-7470A	PPB	PPB	28 May 2022 12:39:03	U	SMPL
N3076-03 - 1	RBR-671	9	0.0320	-7470A	PPB	PPB	28 May 2022 12:41:18	U	SMPL
CCV83 - 1	CCV83	20001	4.8881	-7470A	PPB	PPB	28 May 2022 12:43:33	U	SMPL
CCB83 - 1	CCB83	-87	0.0087	-7470A	PPB	PPB	28 May 2022 12:45:48	U	SMPL
N3080-02 - 1	WC-4	-12	0.0269	-7470A	PPB	PPB	28 May 2022 12:48:04	U	SMPL
N3080-05 - 1	WC-5	3	0.0306	-7470A	PPB	PPB	28 May 2022 12:50:20	U	SMPL
N3080-05DUP - 1	WC-5DUP	8	0.0318	-7470A	PPB	PPB	28 May 2022 12:52:35	U	SMPL
N3080-05MS - 1	WC-5MS	15819	3.8723	-7470A	PPB	PPB	28 May 2022 12:54:51	U	SMPL
N3080-05MSD - 1	WC-5MSD	16020	3.9211	-7470A	PPB	PPB	28 May 2022 12:57:06	U	SMPL
PB145169BL - 1	PBW	-76	0.0114	-7470A	PPB	PPB	28 May 2022 12:59:23	U	SMPL
PB145169BS - 1	LCSW	16274	3.9828	-7470A	PPB	PPB	28 May 2022 13:01:39	U	SMPL
N3042-08 - 1	304	1237	0.3303	-7470A	PPB	PPB	28 May 2022 13:03:56	U	SMPL
N3057-03 - 1	WATER-DRUM	140	0.0638	-7470A	PPB	PPB	28 May 2022 13:06:12	U	SMPL
N3064-01 - 1	MW-1	30	0.0371	-7470A	PPB	PPB	28 May 2022 13:08:28	U	SMPL
CCV84 - 1	CCV84	22113	5.4011	-7470A	PPB	PPB	28 May 2022 13:10:43	U	SMPL
CCB84 - 1	CCB84	-100	0.0055	-7470A	PPB	PPB	28 May 2022 13:12:58	U	SMPL
N3064-01DUP - 1	MW-1DUP	40	0.0396	-7470A	PPB	PPB	28 May 2022 13:15:15	U	SMPL
N3064-02 - 1	N3064-01MS	15324	3.7521	-7470A	PPB	PPB	28 May 2022 13:17:30	U	SMPL
N3064-03 - 1	N3064-01MSD	15533	3.8028	-7470A	PPB	PPB	28 May 2022 13:19:45	U	SMPL
N3064-06 - 1	FIELD-BLANK	-102	0.0051	-7470A	PPB	PPB	28 May 2022 13:22:01	U	SMPL
N3089-01 - 1	220517053-02	738	0.2091	-7470A	PPB	PPB	28 May 2022 13:24:17	U	SMPL
PB145143TB - 1	PB145143TB	-11	0.0272	-7470A	PPB	PPB	28 May 2022 13:26:33	U	SMPL
N3080-05LX5 - 1		-36	0.0211	-7470A	PPB	PPB	28 May 2022 13:28:50	U	SMPL
N3080-05A - 1		15575	3.8130	-7470A	PPB	PPB	28 May 2022 13:31:06	U	SMPL
N3064-01LX5 - 1		-107	0.0038	-7470A	PPB	PPB	28 May 2022 13:33:23	U	SMPL
N3064-01A - 1		15215	3.7256	-7470A	PPB	PPB	28 May 2022 13:35:39	U	SMPL
CCV85 - 1	CCV85	19348	4.7295	-7470A	PPB	PPB	28 May 2022 13:37:56	U	SMPL
CCB85 - 1	CCB85	-87	0.0087	-7470A	PPB	PPB	28 May 2022 13:40:12	U	SMPL