

LB120401


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

56286



A= 0.0000e+000

B= 1.7855e-004

C= -2.9128e-004

Rho= 0.9998493

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.001	0.001	10	0.000	10					-
0.2	0.200	0.208	0.008	1164	0.0 %	1164					4
2.5	2.500	2.473	-0.027	13854	0.0 %	13854					-1
5.0	5.000	5.086	0.086	28488	0.0 %	28488					2
7.5	7.500	7.382	-0.118	41345	0.0 %	41345					-2
10.0	10.000	10.050	0.050	56286	0.0 %	56286					1

LB120401 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 May 2022 14:31:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7471B	PPB	28 May 2022 14:34:03	S	Std
0.2 - 1	50.2	1164	-	0.2000	7471B	PPB	28 May 2022 14:36:19	S	Std
2.5 - 1	52.5	13854	-	2.5000	7471B	PPB	28 May 2022 14:38:36	S	Std
5.0 - 1	55	28488	-	5.0000	7471B	PPB	28 May 2022 14:40:53	S	Std
7.5 - 1	57.5	41345	-	7.5000	7471B	PPB	28 May 2022 14:43:10	S	Std
10.0 - 1	510	56286	-	10.0000	7471B	PPB	28 May 2022 14:45:27	S	Std
ICV51 - 1	ICV51	22307	3.9826	-	7471B	PPB	28 May 2022 14:48:14	U	SMPL
ICB51 - 1	ICB51	-162	-0.0292	-	7471B	PPB	28 May 2022 14:50:28	U	SMPL
CCV86 - 1	CCV86	27695	4.9446	-	7471B	PPB	28 May 2022 14:52:45	U	SMPL
CCB86 - 1	CCB86	-271	-0.0487	-	7471B	PPB	28 May 2022 14:54:59	U	SMPL
CRA - 1	CRA	1086	0.1936	-	7471B	PPB	28 May 2022 14:59:16	U	SMPL
HighStd - 1	HighStd	56042	10.0060	-	7471B	PPB	28 May 2022 15:01:31	U	SMPL
ChkStd - 1	ChkStd	39215	7.0015	-	7471B	PPB	28 May 2022 15:03:46	U	SMPL
PB145198BL - 1	PBS	-187	-0.0337	-	7471B	PPB	28 May 2022 15:06:04	U	SMPL
PB145198BS - 1	LCSS	20944	3.7393	-	7471B	PPB	28 May 2022 15:08:22	U	SMPL
N2951-01 - 1	P035-SS001-0006-01	6361	1.1355	-	7471B	PPB	28 May 2022 15:10:39	U	SMPL
N2951-01DUP - 1	P035-SS001-0006-01DUP	6704	1.1967	-	7471B	PPB	28 May 2022 15:12:56	U	SMPL
N2951-02 - 1	N2951-01MS	27445	4.9000	-	7471B	PPB	28 May 2022 15:15:13	U	SMPL
N2951-03 - 1	N2951-01MSD	26949	4.8114	-	7471B	PPB	28 May 2022 15:17:30	U	SMPL
N2951-04 - 1	P035-SS001-0006-02	6648	1.1867	-	7471B	PPB	28 May 2022 15:19:46	U	SMPL
CCV87 - 1	CCV87	27850	4.9723	-	7471B	PPB	28 May 2022 15:22:03	U	SMPL
CCB87 - 1	CCB87	-71	-0.0130	-	7471B	PPB	28 May 2022 15:24:18	U	SMPL
N2951-05 - 1	P035-SS001-0612-01	2767	0.4938	-	7471B	PPB	28 May 2022 15:26:35	U	SMPL
N2951-06 - 1	P035-SS001-1218-01	1554	0.2772	-	7471B	PPB	28 May 2022 15:28:50	U	SMPL
N2951-07 - 1	N2951-06MS	22700	4.0528	-	7471B	PPB	28 May 2022 15:31:06	U	SMPL
N2951-08 - 1	N2951-06MSD	22747	4.0612	-	7471B	PPB	28 May 2022 15:33:21	U	SMPL
N2951-09 - 1	P035-SS001-1218-02	1435	0.2559	-	7471B	PPB	28 May 2022 15:35:38	U	SMPL
N2951-10 - 1	P035-SS001-1824-01	3306	0.5900	-	7471B	PPB	28 May 2022 15:37:56	U	SMPL
N2951-11 - 1	P035-SS001-2430-01	4966	0.8864	-	7471B	PPB	28 May 2022 15:40:13	U	SMPL
N2951-12 - 1	P035-SS002-0006-01	2338	0.4172	-	7471B	PPB	28 May 2022 15:42:29	U	SMPL
N2951-17 - 1	P036-SS001-0006-01	21414	3.8232	-	7471B	PPB	28 May 2022 15:44:46	U	SMPL
N2951-18 - 1	P036-SS001-0612-01	30328	5.4148	-	7471B	PPB	28 May 2022 15:47:01	U	SMPL
CCV88 - 1	CCV88	27883	4.9782	-	7471B	PPB	28 May 2022 15:49:18	U	SMPL
CCB88 - 1	CCB88	-83	-0.0151	-	7471B	PPB	28 May 2022 15:51:35	U	SMPL
N2951-19 - 1	N2951-18MS	56455	10.0797	-	7471B	PPB	28 May 2022 15:53:53	U	SMPL
N2951-20 - 1	N2951-18MSD	59812	10.6791	-	7471B	PPB	28 May 2022 15:56:08	U	SMPL
N2951-21 - 1	P036-SS001-0612-02	24679	4.4061	-	7471B	PPB	28 May 2022 15:58:26	U	SMPL
N2951-22 - 1	P036-SS001-1218-01	18545	3.3109	-	7471B	PPB	28 May 2022 16:00:44	U	SMPL
N2951-23 - 1	P036-SS001-1824-01	21494	3.8375	-	7471B	PPB	28 May 2022 16:03:01	U	SMPL
N2951-24 - 1	P036-SS001-2430-01	5923	1.0573	-	7471B	PPB	28 May 2022 16:05:18	U	SMPL
N2951-25 - 1	P036-SS002-0006-01	5226	0.9328	-	7471B	PPB	28 May 2022 16:07:35	U	SMPL
PB145199BL - 1	PBS	-58	-0.0106	-	7471B	PPB	28 May 2022 16:09:52	U	SMPL
PB145199BS - 1	LCSS	20987	3.7469	-	7471B	PPB	28 May 2022 16:12:09	U	SMPL
N2952-07 - 1	P037-SS001-0006-01	2747	0.4902	-	7471B	PPB	28 May 2022 16:14:24	U	SMPL
CCV89 - 1	CCV89	28613	5.1085	-	7471B	PPB	28 May 2022 16:16:40	U	SMPL
CCB89 - 1	CCB89	-161	-0.0290	-	7471B	PPB	28 May 2022 16:18:56	U	SMPL
N2952-08 - 1	P037-SS001-0612-01	3473	0.6198	-	7471B	PPB	28 May 2022 16:21:13	U	SMPL
N2952-09 - 1	P037-SS001-1218-01	3602	0.6428	-	7471B	PPB	28 May 2022 16:23:28	U	SMPL
N2952-09DUP - 1	P037-SS001-1218-01DUP	3171	0.5659	-	7471B	PPB	28 May 2022 16:25:44	U	SMPL
N2952-10 - 1	N2952-09MS	24886	4.4431	-	7471B	PPB	28 May 2022 16:27:59	U	SMPL
N2952-11 - 1	N2952-09MSD	24983	4.4604	-	7471B	PPB	28 May 2022 16:30:15	U	SMPL
N2952-12 - 1	P037-SS001-1218-02	4150	0.7407	-	7471B	PPB	28 May 2022 16:32:32	U	SMPL
N2952-13 - 1	P037-SS001-1824-01	3611	0.6445	-	7471B	PPB	28 May 2022 16:34:50	U	SMPL
N2952-14 - 1	P037-SS001-2430-01	1481	0.2641	-	7471B	PPB	28 May 2022 16:37:06	U	SMPL
N2952-15 - 1	P037-SS002-0006-01	3356	0.5989	-	7471B	PPB	28 May 2022 16:39:23	U	SMPL
N2952-26 - 1	P043-SS001-0006-01	12413	2.2160	-	7471B	PPB	28 May 2022 16:41:39	U	SMPL
CCV90 - 1	CCV90	28273	5.0478	-	7471B	PPB	28 May 2022 16:43:54	U	SMPL
CCB90 - 1	CCB90	-203	-0.0365	-	7471B	PPB	28 May 2022 16:46:10	U	SMPL
N2952-27 - 1	P043-SS001-0612-01	6733	1.2019	-	7471B	PPB	28 May 2022 16:48:28	U	SMPL
N2952-28 - 1	N2952-27MS	29816	5.3233	-	7471B	PPB	28 May 2022 16:50:43	U	SMPL
N2952-29 - 1	N2952-27MSD	31548	5.6326	-	7471B	PPB	28 May 2022 16:52:58	U	SMPL
N2952-30 - 1	P043-SS001-0612-02	10240	1.8281	-	7471B	PPB	28 May 2022 16:55:15	U	SMPL
N2954-05 - 1	P044-SS004-0006-01	8441	1.5068	-	7471B	PPB	28 May 2022 16:57:33	U	SMPL
N2954-10 - 1	P044-SS005-0006-01	3462	0.6178	-	7471B	PPB	28 May 2022 16:59:49	U	SMPL
PB145200BL - 1	PBS	-36	-0.0067	-	7471B	PPB	28 May 2022 17:02:04	U	SMPL
PB145200BS - 1	LCSS	21228	3.7900	-	7471B	PPB	28 May 2022 17:04:21	U	SMPL
N2953-01 - 1	P043-SS001-1218-01	2380	0.4247	-	7471B	PPB	28 May 2022 17:06:38	U	SMPL
N2953-02 - 1	P043-SS001-1824-01	1538	0.2743	-	7471B	PPB	28 May 2022 17:08:55	U	SMPL
CCV91 - 1	CCV91	28410	5.0723	-	7471B	PPB	28 May 2022 17:11:12	U	SMPL
CCB91 - 1	CCB91	-139	-0.0251	-	7471B	PPB	28 May 2022 17:13:28	U	SMPL
N2953-03 - 1	P043-SS001-2430-01	1366	0.2436	-	7471B	PPB	28 May 2022 17:15:46	U	SMPL
N2953-04 - 1	P044-SS001-0006-01	1859	0.3316	-	7471B	PPB	28 May 2022 17:18:01	U	SMPL
N2953-05 - 1	P044-SS001-0612-01	3214	0.5736	-	7471B	PPB	28 May 2022 17:20:17	U	SMPL
N2953-05DUP - 1	P044-SS001-0612-01DUP	3032	0.5411	-	7471B	PPB	28 May 2022 17:22:33	U	SMPL
N2953-06 - 1	N2953-05MS	25871	4.6190	-	7471B	PPB	28 May 2022 17:24:49	U	SMPL
N2953-07 - 1	N2953-05MSD	23737	4.2379	-	7471B	PPB	28 May 2022 17:27:05	U	SMPL
N2953-08 - 1	P044-SS001-0612-02	2936	0.5239	-	7471B	PPB	28 May 2022 17:29:23	U	SMPL
N2953-09 - 1	P044-SS001-1218-01	2739	0.4888	-	7471B	PPB	28 May 2022 17:31:41	U	SMPL
N2953-10 - 1	P044-SS001-1824-01	1422	0.2536	-	7471B	PPB	28 May 2022 17:33:59	U	SMPL
N2953-11 - 1	P044-SS001-2430-01	1062	0.1893	-	7471B	PPB	28 May 2022 17:36:15	U	SMPL
CCV92 - 1	CCV92	27607	4.9289	-	7471B	PPB	28 May 2022 17:38:31	U	SMPL

LB120401 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 May 2022 14:31:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB92 - 1	CCB92	-63	-0.0115		-7471B	PPB	28 May 2022 17:40:47	U	SMPL
N2953-12 - 1	P044-SS002-0006-01	6875	1.2272		-7471B	PPB	28 May 2022 17:43:05	U	SMPL
N2953-13 - 1	N2953-12MS	28010	5.0009		-7471B	PPB	28 May 2022 17:45:22	U	SMPL
N2953-14 - 1	N2953-12MSD	28064	5.0105		-7471B	PPB	28 May 2022 17:47:41	U	SMPL
N2953-15 - 1	P044-SS002-0006-02	8050	1.4370		-7471B	PPB	28 May 2022 17:49:59	U	SMPL
N2953-26 - 1	P044-SS003-0006-01	12308	2.1973		-7471B	PPB	28 May 2022 17:52:17	U	SMPL
N2951-18DLX2 - 1		15757	2.8131		-7471B	PPB	28 May 2022 17:54:33	U	SMPL
N2951-19DLX2 - 1		28098	5.0166		-7471B	PPB	28 May 2022 17:59:07	U	SMPL
N2951-20DLX2 - 1		30253	5.4014		-7471B	PPB	28 May 2022 18:01:23	U	SMPL
CCV93 - 1	CCV93	28117	5.0200		-7471B	PPB	28 May 2022 18:03:42	U	SMPL
CCB93 - 1	CCB93	44	0.0076		-7471B	PPB	28 May 2022 18:06:00	U	SMPL
N2951-01LX5 - 1		1269	0.2263		-7471B	PPB	28 May 2022 18:08:17	U	SMPL
N2951-01A - 1		28376	5.0662		-7471B	PPB	28 May 2022 18:10:34	U	SMPL
N2952-09LX5 - 1		663	0.1181		-7471B	PPB	28 May 2022 18:12:50	U	SMPL
N2952-09A - 1		26288	4.6934		-7471B	PPB	28 May 2022 18:15:07	U	SMPL
N2953-05LX5 - 1		556	0.0990		-7471B	PPB	28 May 2022 18:17:23	U	SMPL
N2953-05A - 1		25223	4.5033		-7471B	PPB	28 May 2022 18:19:41	U	SMPL
CCV94 - 1	CCV94	27676	4.9412		-7471B	PPB	28 May 2022 18:21:57	U	SMPL
CCB94 - 1	CCB94	-139	-0.0251		-7471B	PPB	28 May 2022 18:24:14	U	SMPL