



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB120402

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	05/28/2022	10:50
2	N3042-08		1	12	100.00	>212.0	>212.0	05/28/2022	11:10
3	N3042-08DUP		1	12	100.00	>212.0	>212.0	05/28/2022	11:35
4	N3053-07		1	15	100.00	>212.0	>212.0	05/28/2022	11:55
5	N3057-03		1	18	100.00	>212.0	>212.0	05/28/2022	12:20
6	N3059-01		1	17	100.00	>212.0	>212.0	05/28/2022	12:45
7	N3067-02		1	15	100.00	>212.0	>212.0	05/28/2022	13:05
8	N3068-01		1	14	100.00	>212.0	>212.0	05/28/2022	13:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB120402

WorkList Name : FP52822

WorkList ID : 160084

Department : Wet-Chemistry

Date : 05-28-2022 10:26:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/31/2022	Water	N3042-08	Flash Point	Cool 4 deg C	PSEG03	N11	304	05/25/2022	1010B
05/30/2022	Water	N3053-07	Flash Point	Cool 4 deg C	PSEG03	N11	52322-WATER	05/25/2022	1010B
06/01/2022	Water	N3057-03	Flash Point	Cool 4 deg C	PSEG03	N11	WATER-DRUM	05/25/2022	1010B
05/30/2022	Water	N3059-01	Flash Point	Cool 4 deg C	PSEG03	N11	4922A	05/25/2022	1010B
05/31/2022	Water	N3067-02	Flash Point	Cool 4 deg C	PSEG03	N11	SUM-WATER-0526	05/26/2022	1010B
06/02/2022	Water	N3068-01	Flash Point	Cool 4 deg C	TULL02	N11	WATER-DRUM	05/26/2022	1010B

Date/Time 05/28/22 10:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 05/28/22 13:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)