



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 5/31/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 15:00

In Date: 05/27/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:10

Out Date: 05/28/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB120409

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N2920-01	P015-SS006-0006-01	1	0.9	8.84	9.74	9.25	94.5	
N2920-02	P015-SS007-0006-01	2	0.9	8.93	9.83	7.96	79.1	
N2920-03	P015-SS008-0006-01	3	0.85	9.07	9.92	8.8	87.7	
N2920-04	P015-SS009-0006-01	4	0.89	8.62	9.51	9.00	94.1	
N2920-05	P015-SS010-0006-01	5	0.91	8.98	9.89	9.26	93.0	
N2920-06	P024-SS001-0006-01	6	0.98	8.49	9.47	8.48	88.3	
N2920-07	P024-SS001-0612-01	7	0.92	8.46	9.38	8.52	89.8	
N2920-08	P024-SS001-1218-01	8	0.91	8.67	9.58	8.32	85.5	
N2920-09	P024-SS001-1824-01	9	0.94	8.48	9.42	8.52	89.4	
N2920-10	P024-SS001-2430-01	10	0.9	8.66	9.56	8.67	89.7	
N2920-11	N2920-10MS	11	0.9	8.66	9.56	8.67	89.7	
N2920-12	N2920-10MSD	12	0.9	8.66	9.56	8.67	89.7	
N2920-13	P024-SS001-2430-02	13	0.92	8.72	9.64	8.71	89.3	
N2920-14	P024-SS002-0006-01	14	0.88	8.68	9.56	8.61	89.1	
N2920-15	P024-SS003-0006-01	15	0.91	8.47	9.38	7.96	83.2	
N2920-16	P024-SS004-0006-01	16	0.92	8.75	9.67	8.5	86.6	
N2920-17	P024-SS005-0006-01	17	0.93	8.62	9.55	8.62	89.2	
N2920-18	P024-SS006-0006-01	18	0.94	9.02	9.96	9.3	92.7	
N2920-19	P024-SS007-0006-01	19	0.93	8.99	9.92	9.24	92.4	
N2920-20	P024-SS008-0006-01	20	0.93	8.84	9.77	8.75	88.5	
N2920-21	P024-SS009-0006-01	21	0.88	8.80	9.68	8.41	85.6	
N2920-22	P024-SS010-0006-01	22	0.88	8.84	9.72	8.64	87.8	
N2920-23	P028-SS001-0006-01	23	0.91	8.94	9.85	9.14	92.1	
N2920-24	P028-SS001-1218-01	24	0.9	8.61	9.51	8.54	88.7	
N2920-25	P028-SS002-0006-01	25	0.92	8.76	9.68	8.81	90.1	
N2920-26	P028-SS003-0006-01	26	0.93	9.01	9.94	9.39	93.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$