

LB 120411


7471B

INSTRUMENT 1D CV2

Linear

 μ Abs.:

55536



A= 0.0000e+000

B= 1.8099e-004

C= -1.7834e-002

Rho= 0.9997713

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.018	-0.018	-3	0.000	-3					-
0.2	0.200	0.187	-0.013	1129	0.0 %	1129					-7
2.5	2.500	2.507	0.007	13950	0.0 %	13950					0
5.0	5.000	5.129	0.129	28438	0.0 %	28438					3
7.5	7.500	7.362	-0.138	40776	0.0 %	40776					-2
10.0	10.000	10.034	0.034	55536	0.0 %	55536					0

LB120411 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 31 May 2022 08:49:04

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.00007471B	PPB		31 May 2022 08:55:31	S	Std
0.2 - 1	50.2	1129	-	0.20007471B	PPB		31 May 2022 08:57:48	S	Std
2.5 - 1	52.5	13950	-	2.50007471B	PPB		31 May 2022 09:00:05	S	Std
5.0 - 1	55	28438	-	5.00007471B	PPB		31 May 2022 09:02:23	S	Std
7.5 - 1	57.5	40776	-	7.50007471B	PPB		31 May 2022 09:04:40	S	Std
10.0 - 1	570	55536	-	10.00007471B	PPB		31 May 2022 09:06:57	S	Std
ICV52 - 1	ICV52	22702	4.0910	-7471B	PPB		31 May 2022 09:09:39	U	SMPL
ICB52 - 1	ICB52	-77	-0.0318	-7471B	PPB		31 May 2022 09:11:53	U	SMPL
CCV95 - 1	CCV95	27999	5.0497	-7471B	PPB		31 May 2022 09:14:09	U	SMPL
CCB95 - 1	CCB95	-72	-0.0309	-7471B	PPB		31 May 2022 09:16:24	U	SMPL
CRA - 1	CRA	1092	0.1798	-7471B	PPB		31 May 2022 09:18:41	U	SMPL
HighStd - 1	HighStd	55030	9.9420	-7471B	PPB		31 May 2022 09:20:56	U	SMPL
ChkStd - 1	ChkStd	38237	6.9027	-7471B	PPB		31 May 2022 09:23:11	U	SMPL
PB145208BL - 1	PBS	-118	-0.0392	-7471B	PPB		31 May 2022 09:25:28	U	SMPL
PB145208BS - 1	LCSS	20136	3.6266	-7471B	PPB		31 May 2022 09:27:45	U	SMPL
N3088-01 - 1	JSW-COMP-01	3417	0.6006	-7471B	PPB		31 May 2022 09:35:20	U	SMPL
N3088-01DUP - 1	JSW-COMP-01DUP	3389	0.5955	-7471B	PPB		31 May 2022 09:40:11	U	SMPL
N3088-01MS - 1	JSW-COMP-01MS	25664	4.6271	-7471B	PPB		31 May 2022 09:42:27	U	SMPL
N3088-01MSD - 1	JSW-COMP-01MSD	27698	4.9952	-7471B	PPB		31 May 2022 09:44:44	U	SMPL
N3090-01 - 1	220517028-04	5499	0.9774	-7471B	PPB		31 May 2022 09:47:01	U	SMPL
CCV96 - 1	CCV96	28731	5.1822	-7471B	PPB		31 May 2022 09:49:18	U	SMPL
CCB96 - 1	CCB96	-68	-0.0301	-7471B	PPB		31 May 2022 09:51:32	U	SMPL
N3091-01 - 1	220516067-01	1908	0.3275	-7471B	PPB		31 May 2022 09:53:49	U	SMPL
N3099-01 - 1	20220417-COMPOSITE-F	20231	3.6438	-7471B	PPB		31 May 2022 09:56:05	U	SMPL
N3099-02 - 1	20220420-COMPOSITE-I	33644	6.0714	-7471B	PPB		31 May 2022 09:58:20	U	SMPL
N3099-03 - 1	20220426-AMENDED-COMPOSITE-F	22559	4.0651	-7471B	PPB		31 May 2022 10:00:37	U	SMPL
N3099-04 - 1	20220429-AMENDED-COMPOSITE-I	27822	5.0176	-7471B	PPB		31 May 2022 10:02:54	U	SMPL
N3104-01 - 1	WC-6	1614	0.2743	-7471B	PPB		31 May 2022 10:05:12	U	SMPL
N3104-04 - 1	WC-7	38727	6.9913	-7471B	PPB		31 May 2022 10:07:30	U	SMPL
N3088-01LX5 - 1		571	0.0855	-7471B	PPB		31 May 2022 10:09:47	U	SMPL
N3088-01A - 1		26641	4.8039	-7471B	PPB		31 May 2022 10:12:06	U	SMPL
CCV97 - 1	CCV97	28634	5.1646	-7471B	PPB		31 May 2022 10:14:21	U	SMPL
CCB97 - 1	CCB97	-61	-0.0289	-7471B	PPB		31 May 2022 10:16:38	U	SMPL
N2951-18DLX2 - 1		15654	2.8154	-7471B	PPB		31 May 2022 10:18:55	U	SMPL
N2951-19DLX2 - 1		28566	5.1523	-7471B	PPB		31 May 2022 10:26:52	U	SMPL
N2951-20DLX2 - 1		29699	5.3574	-7471B	PPB		31 May 2022 10:29:07	U	SMPL
N2951-18ADLX2 - 1		28738	5.1834	-7471B	PPB		31 May 2022 10:31:24	U	SMPL
CCV98 - 1	CCV98	28698	5.1762	-7471B	PPB		31 May 2022 10:33:41	U	SMPL
CCB98 - 1	CCB98	-48	-0.0265	-7471B	PPB		31 May 2022 10:35:58	U	SMPL