

LB 120425


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43677



A= 0.0000e+000

B= 2.2898e-004

C= 4.8482e-002

Rho= 0.9997074

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/0/1
0.0	0.000	0.057	0.057	38	0.000	38					
0.2	0.200	0.240	0.040	837	0.0 %	837					20
2.5	2.500	2.343	-0.157	10021	0.0 %	10021					-6
5.0	5.000	5.094	0.094	22034	0.0 %	22034					2
7.5	7.500	7.416	-0.084	32176	0.0 %	32176					-1
10.0	10.000	10.050	0.050	43677	0.0 %	43677					1

LB120425 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 May 2022 13:21:43

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	38	-	0.00007470A	PPB	31 May 2022 13:23:56	S	Std
0.2 - 1	50.2	837	-	0.20007470A	PPB	31 May 2022 13:26:13	S	Std
2.5 - 1	50.5	10021	-	2.50007470A	PPB	31 May 2022 13:28:30	S	Std
5.0 - 1	55	22034	-	5.00007470A	PPB	31 May 2022 13:30:46	S	Std
7.5 - 1	57.5	32176	-	7.50007470A	PPB	31 May 2022 13:33:03	S	Std
10.0 - 1	57.5	43677	-	10.00007470A	PPB	31 May 2022 13:40:31	S	Std
ICV53 - 1	ICV53	17693	4.0998	-7470A	PPB	31 May 2022 13:45:48	U	SMPL
ICB53 - 1	ICB53	-53	0.0363	-7470A	PPB	31 May 2022 13:48:02	U	SMPL
CCV01 - 1	CCV01	22174	5.1259	-7470A	PPB	31 May 2022 13:50:18	U	SMPL
CCB01 - 1	CCB01	-57	0.0354	-7470A	PPB	31 May 2022 13:52:33	U	SMPL
CRA - 1	CRA	822	0.2367	-7470A	PPB	31 May 2022 13:54:49	U	SMPL
HighStd - 1	HighStd	43557	10.0222	-7470A	PPB	31 May 2022 13:57:04	U	SMPL
ChkStd - 1	ChkStd	30112	6.9435	-7470A	PPB	31 May 2022 13:59:19	U	SMPL
PB145205BL - 1	PBW	-63	0.0341	-7470A	PPB	31 May 2022 14:01:36	U	SMPL
PB145205BS - 1	LCSW	16761	3.8864	-7470A	PPB	31 May 2022 14:03:53	U	SMPL
N3099-05 - 1	20220426-SPLP-AMENDED-COMPOSITE-F	2369	0.5909	-7470A	PPB	31 May 2022 14:06:09	U	SMPL
N3099-05DUP - 1	20220426-SPLP-AMENDED-COMPOSITE-FDUP	2446	0.6086	-7470A	PPB	31 May 2022 14:08:25	U	SMPL
N3099-05MS - 1	20220426-SPLP-AMENDED-COMPOSITE-FMS	18349	4.2500	-7470A	PPB	31 May 2022 14:10:42	U	SMPL
N3099-05MSD - 1	20220426-SPLP-AMENDED-COMPOSITE-FMSD	17411	4.0352	-7470A	PPB	31 May 2022 14:13:00	U	SMPL
N3099-06 - 1	20220429-SPLP-AMENDED-COMPOSITE-I	1487	0.3890	-7470A	PPB	31 May 2022 14:15:15	U	SMPL
CCV02 - 1	CCV02	21999	5.0858	-7470A	PPB	31 May 2022 14:17:31	U	SMPL
CCB02 - 1	CCB02	-56	0.0357	-7470A	PPB	31 May 2022 14:19:46	U	SMPL
N3099-05LX5 - 1		496	0.1621	-7470A	PPB	31 May 2022 14:22:02	U	SMPL
N3099-05A - 1		18214	4.2191	-7470A	PPB	31 May 2022 14:24:18	U	SMPL
PB145205TB - 1	PB145164TB	-27	0.0423	-7470A	PPB	31 May 2022 14:26:33	U	SMPL
CCV03 - 1	CCV03	22045	5.0963	-7470A	PPB	31 May 2022 14:28:49	U	SMPL
CCB03 - 1	CCB03	-52	0.0366	-7470A	PPB	31 May 2022 14:31:05	U	SMPL