



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB120428

Slope : 99.3

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	05/31/2022	11:25
2	CAL2	1	Water	NA	NA	20.9	7.00	05/31/2022	11:26
3	CAL3	1	Water	NA	NA	20.9	10.03	05/31/2022	11:27
4	ICV	1	Water	NA	NA	20.8	6.99	05/31/2022	11:30
5	CCV1	1	Water	NA	NA	20.9	2.01	05/31/2022	11:32
6	N3067-01	1	Solid	20.02	20	21.9	8.84	05/31/2022	11:35
7	N3108-01	1	Solid	20.03	20	24.1	6.32	05/31/2022	11:39
8	N3108-02	1	Solid	20.02	20	23.9	5.53	05/31/2022	11:40
9	N3108-03	1	Solid	20.01	20	23.7	5.60	05/31/2022	11:44
10	N3108-03DUP	1	Solid	20.02	20	23.8	5.61	05/31/2022	11:45
11	CCV2	1	Water	NA	NA	20.8	12.03	05/31/2022	11:47

WORKLIST(Hardcopy Internal Chain)

VB 120628

WorkList Name : PH N3108

WorkList ID : 160116

Department : Wet-Chemistry

Date : 05-31-2022 11:19:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/26/2022	Solid	N3067-01	pH	Cool 4 deg C	PSEG03	N11	33686	05/26/2022	9045D
05/27/2022	Solid	N3108-01	pH	Cool 4 deg C	EARTH03	N11	NP-1	05/27/2022	9045D
05/27/2022	Solid	N3108-02	pH	Cool 4 deg C	EARTH03	N11	NP-2	05/27/2022	9045D
05/27/2022	Solid	N3108-03	pH	Cool 4 deg C	EARTH03	N11	NP-3	05/27/2022	9045D

Date/Time 05/31/22 11:20
 Raw Sample Received by: JY (SM)
 Raw Sample Relinquished by: JY (SM)

Date/Time 05/31/22 13:00
 Raw Sample Received by: JY (SM)
 Raw Sample Relinquished by: JY (SM)