

LB120433


245.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

38944



A= 0.0000e+000

B= 2.5477e-004

C= -4.3758e-003

Rho= 0.9998700

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	gnd
0.0	0.000	0.012	0.012	66	0.000	66					
0.05	0.050	0.045	-0.005	192	0.0 %	192	0				-10
0.2	0.200	0.195	-0.005	781	0.0 %	781					-3
2.5	2.500	2.424	-0.076	9533	0.0 %	9533					-3
5.0	5.000	5.073	0.073	19931	0.0 %	19931					-1
7.5	7.500	7.583	0.083	29782	0.0 %	29782					-1
10.0	10.000	9.917	-0.083	38944	0.0 %	38944					-1

LB120433 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 31 May 2022 15:45:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	66	-	0.0000	245.1	PPB	31 May 2022 15:46:59	S	Std
0.05 - 1	50.05	192	-	0.0500	245.1	PPB	31 May 2022 15:49:16	S	Std
0.2 - 1	50.2	781	-	0.2000	245.1	PPB	31 May 2022 15:51:33	S	Std
2.5 - 1	52.5	9533	-	2.5000	245.1	PPB	31 May 2022 15:53:50	S	Std
5.0 - 1	55	19931	-	5.0000	245.1	PPB	31 May 2022 15:56:06	S	Std
7.5 - 1	57.5	29782	-	7.5000	245.1	PPB	31 May 2022 15:58:22	S	Std
10.0 - 1	51.0	38944	-	10.0000	245.1	PPB	31 May 2022 16:11:00	S	Std
ICV54 - 1	ICV54	15802	4.0215	-245.1	PPB	PPB	31 May 2022 16:13:43	U	SMPL
ICB54 - 1	ICB54	-4	-0.0054	-245.1	PPB	PPB	31 May 2022 16:15:57	U	SMPL
CCV04 - 1	CCV04	19892	5.0635	-245.1	PPB	PPB	31 May 2022 16:18:13	U	SMPL
CCB04 - 1	CCB04	-3	-0.0051	-245.1	PPB	PPB	31 May 2022 16:20:28	U	SMPL
CRA - 1	CRA	172	0.0394	-245.1	PPB	PPB	31 May 2022 16:22:43	U	SMPL
HighStd - 1	HighStd	38900	9.9062	-245.1	PPB	PPB	31 May 2022 16:24:58	U	SMPL
ChkStd - 1	ChkStd	27738	7.0625	-245.1	PPB	PPB	31 May 2022 16:27:13	U	SMPL
PB145229BL - 1	PBW	-119	-0.0347	-245.1	PPB	PPB	31 May 2022 16:29:31	U	SMPL
PB145229BS - 1	LCSW	16081	4.0926	-245.1	PPB	PPB	31 May 2022 16:31:48	U	SMPL
N2946-03 - 1	OB-722	-82	-0.0253	-245.1	PPB	PPB	31 May 2022 16:34:04	U	SMPL
N2946-03DUP - 1	OB-722DUP	-11	-0.0072	-245.1	PPB	PPB	31 May 2022 16:36:20	U	SMPL
N2946-03MS - 1	OB-722MS	15675	3.9892	-245.1	PPB	PPB	31 May 2022 16:38:37	U	SMPL
N2946-03MSD - 1	OB-722MSD	15424	3.9252	-245.1	PPB	PPB	31 May 2022 16:40:53	U	SMPL
N3068-01 - 1	WATER-DRUM	-318	-0.0854	-245.1	PPB	PPB	31 May 2022 16:43:11	U	SMPL
CCV05 - 1	CCV05	19889	5.0628	-245.1	PPB	PPB	31 May 2022 16:45:28	U	SMPL
CCB05 - 1	CCB05	-8	-0.0064	-245.1	PPB	PPB	31 May 2022 16:47:44	U	SMPL
N2946-03LX5 - 1		-40	-0.0146	-245.1	PPB	PPB	31 May 2022 16:50:00	U	SMPL
N2946-03A - 1		16462	4.1897	-245.1	PPB	PPB	31 May 2022 16:52:15	U	SMPL
CCV06 - 1	CCV06	19725	5.0210	-245.1	PPB	PPB	31 May 2022 16:54:30	U	SMPL
CCB06 - 1	CCB06	-12	-0.0074	-245.1	PPB	PPB	31 May 2022 16:56:46	U	SMPL