


LB120442

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.8138e-004

C= 2.3063e-002

Rho= 0.9999663

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.029	0.029	33	0.000	33					
0.2	0.200	0.209	0.009	1027	0.0 %	1027					4
2.5	2.500	2.454	-0.046	13400	0.0 %	13400					-2
5.0	5.000	4.974	-0.026	27295	0.0 %	27295					-1
7.5	7.500	7.541	0.041	41449	0.0 %	41449					1
10.0	10.000	9.993	-0.007	54968	0.0 %	54968					0

LB120442 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Jun 2022 13:20:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	33	-	0.0000	7471B	PPB	01 Jun 2022 14:36:34	S	Std
0.2 - 1	51.2	1027	-	0.2000	7471B	PPB	01 Jun 2022 14:38:51	S	Std
2.5 - 1	52.5	13400	-	2.5000	7471B	PPB	01 Jun 2022 14:41:08	S	Std
5.0 - 1	55	27295	-	5.0000	7471B	PPB	01 Jun 2022 14:43:25	S	Std
7.5 - 1	57.5	41449	-	7.5000	7471B	PPB	01 Jun 2022 14:45:41	S	Std
10.0 - 1	51V	54968	-	10.0000	7471B	PPB	01 Jun 2022 14:50:37	S	Std
ICV56 - 1	ICV56	21963	4.0067	-	7471B	PPB	01 Jun 2022 14:53:19	U	SMPL
ICB56 - 1	ICB56	-108	0.0035	-	7471B	PPB	01 Jun 2022 14:55:33	U	SMPL
CCV11 - 1	CCV11	27436	4.9994	-	7471B	PPB	01 Jun 2022 14:57:50	U	SMPL
CCB11 - 1	CCB11	-217	-0.0163	-	7471B	PPB	01 Jun 2022 15:00:05	U	SMPL
CRA - 1	CRA	959	0.1970	-	7471B	PPB	01 Jun 2022 15:05:49	U	SMPL
HighStd - 1	HighStd	55364	10.0650	-	7471B	PPB	01 Jun 2022 15:08:04	U	SMPL
PB145242BL - 1	PBS	-111	0.0029	-	7471B	PPB	01 Jun 2022 15:12:37	U	SMPL
PB145242BS - 1	LCSS	21927	4.0002	-	7471B	PPB	01 Jun 2022 15:14:53	U	SMPL
N2931-08 - 1	P009-SS003-2430-02	3210	0.6053	-	7471B	PPB	01 Jun 2022 15:17:09	U	SMPL
N2931-09 - 1	P009-SS004-0006-01	5965	1.1050	-	7471B	PPB	01 Jun 2022 15:19:29	U	SMPL
N3108-01 - 1	NP-1	2016	0.3887	-	7471B	PPB	01 Jun 2022 15:21:48	U	SMPL
N3108-02 - 1	NP-2	1623	0.3174	-	7471B	PPB	01 Jun 2022 15:24:06	U	SMPL
N3108-03 - 1	NP-3	751	0.1593	-	7471B	PPB	01 Jun 2022 15:26:22	U	SMPL
ChkStd - 1	ChkStd	39222	7.1371	-	7471B	PPB	01 Jun 2022 15:28:54	U	SMPL
CCV12 - 1	CCV12	27711	5.0493	-	7471B	PPB	01 Jun 2022 15:31:26	U	SMPL
CCB12 - 1	CCB12	-98	0.0053	-	7471B	PPB	01 Jun 2022 15:33:41	U	SMPL
N3109-01 - 1	BB-8	12973	2.3761	-	7471B	PPB	01 Jun 2022 15:35:58	U	SMPL
N3109-03 - 1	BB-9	8421	1.5505	-	7471B	PPB	01 Jun 2022 15:38:13	U	SMPL
N3109-05 - 1	BB-10	19362	3.5349	-	7471B	PPB	01 Jun 2022 15:40:29	U	SMPL
N3109-07 - 1	BB-11	2509	0.4781	-	7471B	PPB	01 Jun 2022 15:42:44	U	SMPL
N3112-15 - 1	OF-3-20220527	2848	0.5396	-	7471B	PPB	01 Jun 2022 15:49:53	U	SMPL
N3112-15DUP - 1	OF-3-20220527DUP	2855	0.5409	-	7471B	PPB	01 Jun 2022 15:52:09	U	SMPL
N3112-16 - 1	N3112-15MS	24635	4.4914	-	7471B	PPB	01 Jun 2022 15:54:25	U	SMPL
N3112-17 - 1	N3112-15MSD	26826	4.8888	-	7471B	PPB	01 Jun 2022 15:56:42	U	SMPL
N3112-18 - 1	OF-4-20220527	9563	1.7576	-	7471B	PPB	01 Jun 2022 15:59:00	U	SMPL
N3112-22 - 1	DUP03-20220527	3065	0.5790	-	7471B	PPB	01 Jun 2022 16:01:17	U	SMPL
CCV13 - 1	CCV13	27837	5.0721	-	7471B	PPB	01 Jun 2022 16:03:33	U	SMPL
CCB13 - 1	CCB13	-41	0.0156	-	7471B	PPB	01 Jun 2022 16:05:48	U	SMPL
PB145243BL - 1	PBS	19	0.0265	-	7471B	PPB	01 Jun 2022 16:08:05	U	SMPL
PB145243BS - 1	LCSS	22070	4.0261	-	7471B	PPB	01 Jun 2022 16:10:21	U	SMPL
N3107-01 - 1	MS-5	19160	3.4983	-	7471B	PPB	01 Jun 2022 16:12:36	U	SMPL
N3107-03 - 1	MS-6	10822	1.9860	-	7471B	PPB	01 Jun 2022 16:14:53	U	SMPL
N3113-01 - 1	357	1120	0.2262	-	7471B	PPB	01 Jun 2022 16:17:10	U	SMPL
N3113-03 - 1	RBR-346	2485	0.4738	-	7471B	PPB	01 Jun 2022 16:19:28	U	SMPL
N3124-01 - 1	220518045-01	1421	0.2808	-	7471B	PPB	01 Jun 2022 16:21:44	U	SMPL
N3124-05 - 1	220518045-04	1173	0.2358	-	7471B	PPB	01 Jun 2022 16:24:01	U	SMPL
N3125-01 - 1	220520039-01	16687	3.0497	-	7471B	PPB	01 Jun 2022 16:26:18	U	SMPL
N3138-01 - 1	OR-3-053122	2351	0.4495	-	7471B	PPB	01 Jun 2022 16:28:34	U	SMPL
CCV14 - 1	CCV14	27582	5.0259	-	7471B	PPB	01 Jun 2022 16:30:50	U	SMPL
CCB14 - 1	CCB14	-114	0.0024	-	7471B	PPB	01 Jun 2022 16:33:06	U	SMPL
N3138-01DUP - 1	OR-3-053122DUP	2538	0.4834	-	7471B	PPB	01 Jun 2022 16:35:23	U	SMPL
N3138-01MS - 1	OR-3-053122MS	24497	4.4663	-	7471B	PPB	01 Jun 2022 16:37:38	U	SMPL
N3138-01MSD - 1	OR-3-053122MSD	23117	4.2160	-	7471B	PPB	01 Jun 2022 16:39:54	U	SMPL
N3139-01 - 1	COMP-1	25340	4.6192	-	7471B	PPB	01 Jun 2022 16:42:10	U	SMPL
N3139-02 - 1	COMP-2	96330	17.4954	-	7471B	PPB	01 Jun 2022 16:44:27	U	SMPL
N3139-03 - 1	COMP-3	37121	6.7561	-	7471B	PPB	01 Jun 2022 16:46:45	U	SMPL
N3140-01 - 1	RCA-BACKFILL-MATERIAL	975	0.1999	-	7471B	PPB	01 Jun 2022 16:49:05	U	SMPL
CCV15 - 1	CCV15	27999	5.1015	-	7471B	PPB	01 Jun 2022 16:51:23	U	SMPL
CCB15 - 1	CCB15	-116	0.0020	-	7471B	PPB	01 Jun 2022 16:53:39	U	SMPL
N3112-15LX5 - 1		572	0.1268	-	7471B	PPB	01 Jun 2022 16:55:57	U	SMPL
N3112-15A - 1		25990	4.7371	-	7471B	PPB	01 Jun 2022 16:58:12	U	SMPL
N3138-01LX5 - 1		365	0.0893	-	7471B	PPB	01 Jun 2022 17:00:27	U	SMPL
N3138-01A - 1		24184	4.4096	-	7471B	PPB	01 Jun 2022 17:02:45	U	SMPL
N3139-02DLX5 - 1		19483	3.5569	-	7471B	PPB	01 Jun 2022 17:05:00	U	SMPL
CCV16 - 1	CCV16	27700	5.0473	-	7471B	PPB	01 Jun 2022 17:07:17	U	SMPL
CCB16 - 1	CCB16	-165	-0.0069	-	7471B	PPB	01 Jun 2022 17:09:34	U	SMPL