

LB120459


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42743



A= 0.0000e+000

B= 2.3451e-004 *510j**C= 3.1599e-005 *-1.101e-005*

Rho= 0.9999509

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>g/vb</i>
0.00	0.000	0.048	0.048	205	0.000	205					=
0.05	0.050					185	0				-6
0.20	0.200	0.188	-0.012	801	0.0 %	801					0
2.50	2.500	2.493	-0.007	10632	0.0 %	10632					-1
5.00	5.000	4.931	-0.069	21028	0.0 %	21028					0
7.50	7.500	7.516	0.016	32048	0.0 %	32048					0
10.0	10.000	10.024	0.024	42743	0.0 %	42743					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	205	-		0 SFAM01.1	PPB	6/2/2022 13:10	Std
	0.2 S01	801	-		0.2 SFAM01.1	PPB	6/2/2022 13:12	Std
	2.5 S02	10632	-		2.5 SFAM01.1	PPB	6/2/2022 13:14	Std
	5 S03	21028	-		5 SFAM01.1	PPB	6/2/2022 13:16	Std
	7.5 S04	32048	-		7.5 SFAM01.1	PPB	6/2/2022 13:21	Std
	10 S05	42743	-		10 SFAM01.1	PPB	6/2/2022 13:26	Std
ICV057	ICV057	16739	3.9255	-	SFAM01.1	PPB	6/2/2022 13:29	SMPL
ICB057	ICB507	134	0.0315	-	SFAM01.1	PPB	6/2/2022 13:31	SMPL
CCV017	CCV017	21237	4.9803	-	SFAM01.1	PPB	6/2/2022 13:34	SMPL
CCB017	CCB017	147	0.0345	-	SFAM01.1	PPB	6/2/2022 13:36	SMPL
PB145273BL	PBW273	-47	-0.011	-	SFAM01.1	PPB	6/2/2022 13:41	SMPL
N2890-01	YB7C9	55694	13.0609	-	SFAM01.1	PPB	6/2/2022 13:43	SMPL
N2890-02	YB7D1	186969	43.8463	-	SFAM01.1	PPB	6/2/2022 13:45	SMPL
N2890-03	YB7D3	210373	49.3347	-	SFAM01.1	PPB	6/2/2022 13:47	SMPL
N2890-04	YB7E7	739	0.1733	-	SFAM01.1	PPB	6/2/2022 13:50	SMPL
N2890-05	YB7E9	2089	0.4899	-	SFAM01.1	PPB	6/2/2022 13:52	SMPL
N2890-06	YB7F1	1972	0.4625	-	SFAM01.1	PPB	6/2/2022 13:54	SMPL
N2890-07	YB7F7	493	0.1156	-	SFAM01.1	PPB	6/2/2022 13:57	SMPL
N2890-09	YB7G1	569	0.1335	-	SFAM01.1	PPB	6/2/2022 13:59	SMPL
N2890-10	YB7C4	15694	3.6804	-	SFAM01.1	PPB	6/2/2022 14:01	SMPL
N2890-11	YB7C6	110024	25.8018	-	SFAM01.1	PPB	6/2/2022 14:04	SMPL
N2890-12	YB7D6	367483	HIGH	-	SFAM01.1	PPB	6/2/2022 14:10	
N2890-13	YB7D6D	371055	HIGH	-	SFAM01.1	PPB	6/2/2022 14:12	
N2890-14	YB7D6S	484034	HIGH	-	SFAM01.1	PPB	6/2/2022 14:14	
N2890-15	YB7D8	298143	69.9177	-	SFAM01.1	PPB	6/2/2022 14:16	SMPL
N2890-16	YB7E3	71979	16.8799	-	SFAM01.1	PPB	6/2/2022 14:19	SMPL
N2890-18	YB7G3	638	0.1496	-	SFAM01.1	PPB	6/2/2022 14:22	SMPL
N2892-01	YB7D0	-144	-0.0337	-	SFAM01.1	PPB	6/2/2022 14:27	SMPL
N2892-02	YB7D2	217402	50.9831	-	SFAM01.1	PPB	6/2/2022 14:30	SMPL
CCV018	CCV018	19864	4.6584	-	SFAM01.1	PPB	6/2/2022 14:32	SMPL
CCB018	CCB018	-42	-0.0098	-	SFAM01.1	PPB	6/2/2022 14:34	SMPL
N2892-03	YB7D4	194444	45.5992	-	SFAM01.1	PPB	6/2/2022 14:36	SMPL
N2892-04	YB7E8	-270	-0.0633	-	SFAM01.1	PPB	6/2/2022 14:39	SMPL
N2892-05	YB7F0	394	0.0924	-	SFAM01.1	PPB	6/2/2022 14:41	SMPL
N2892-06	YB7F2	416	0.0976	-	SFAM01.1	PPB	6/2/2022 14:43	SMPL
N2892-07	YB7F8	268	0.0629	-	SFAM01.1	PPB	6/2/2022 14:46	SMPL
N2892-09	YB7C5	4656	1.0919	-	SFAM01.1	PPB	6/2/2022 14:48	SMPL
N2892-10	YB7C7	83063	19.4792	-	SFAM01.1	PPB	6/2/2022 14:50	SMPL
N2892-11	YB7D7	21987	5.1562	-	SFAM01.1	PPB	6/2/2022 14:52	SMPL
N2892-12	YB7D7D	20550	4.8192	-	SFAM01.1	PPB	6/2/2022 14:55	SMPL
N2892-14	YB7D9	263739	61.8496	-	SFAM01.1	PPB	6/2/2022 15:04	SMPL
N2892-15	YB7E4	23589	5.5319	-	SFAM01.1	PPB	6/2/2022 15:06	SMPL
PB145275BL	PBW275	114	0.0268	-	SFAM01.1	PPB	6/2/2022 15:08	SMPL
N3097-01	JNLN0	304	0.0713	-	SFAM01.1	PPB	6/2/2022 15:10	SMPL
N3097-02	JNLN1	327	0.0767	-	SFAM01.1	PPB	6/2/2022 15:13	SMPL
N3097-03	JNLN2	320	0.0751	-	SFAM01.1	PPB	6/2/2022 15:15	SMPL

INSTRUMENT ID : CV1

N3097-04	JNLN3	323	0.0758 -	SFAM01.1	PPB	6/2/2022 15:17	SMPL
N3097-05	JNLN4	269	0.0631 -	SFAM01.1	PPB	6/2/2022 15:19	SMPL
N3097-06	JNLN5	319	0.0748 -	SFAM01.1	PPB	6/2/2022 15:22	SMPL
N3097-07	JNLN6	214	0.0502 -	SFAM01.1	PPB	6/2/2022 15:24	SMPL
CCV019	CCV019	20784	4.8741 -	SFAM01.1	PPB	6/2/2022 15:26	SMPL
CCB019	CCB019	-117	-0.0274 -	SFAM01.1	PPB	6/2/2022 15:29	SMPL
N3097-08	JNLN7	1218	0.2857 -	SFAM01.1	PPB	6/2/2022 15:31	SMPL
N3097-09	JNLN8	1239	0.2906 -	SFAM01.1	PPB	6/2/2022 15:33	SMPL
N3097-10	JNLN9	1230	0.2885 -	SFAM01.1	PPB	6/2/2022 15:35	SMPL
N3097-11	JNLN9D	1311	0.3075 -	SFAM01.1	PPB	6/2/2022 15:38	SMPL
N3097-12	JNLN9S	6283	1.4735 -	SFAM01.1	PPB	6/2/2022 15:40	SMPL
N3100-01	JNLP0	1399	0.3281 -	SFAM01.1	PPB	6/2/2022 15:47	SMPL
N3100-02	JNLP1	1467	0.3441 -	SFAM01.1	PPB	6/2/2022 15:49	SMPL
N3100-03	JNLP2	1701	0.3989 -	SFAM01.1	PPB	6/2/2022 15:52	SMPL
N3100-04	JNLP3	1448	0.3396 -	SFAM01.1	PPB	6/2/2022 15:54	SMPL
N3100-05	JNLP4	288	0.0676 -	SFAM01.1	PPB	6/2/2022 15:56	SMPL
N3100-06	JNLP5	325	0.0762 -	SFAM01.1	PPB	6/2/2022 15:58	SMPL
N3100-07	JNLP6	279	0.0655 -	SFAM01.1	PPB	6/2/2022 16:01	SMPL
N3100-08	JNLP7	332	0.0779 -	SFAM01.1	PPB	6/2/2022 16:03	SMPL
N3100-09	JNLP8	306	0.0718 -	SFAM01.1	PPB	6/2/2022 16:05	SMPL
N3100-10	JNLP9	340	0.0798 -	SFAM01.1	PPB	6/2/2022 16:08	SMPL
N3100-11	JNLP9D	305	0.0716 -	SFAM01.1	PPB	6/2/2022 16:10	SMPL
N3100-12	JNLP9S	4318	1.0126 -	SFAM01.1	PPB	6/2/2022 16:12	SMPL
PB145274BL	PBW274	14	0.0033 -	SFAM01.1	PPB	6/2/2022 16:17	SMPL
PB145276BL	PBW276	177	0.0415 -	SFAM01.1	PPB	6/2/2022 16:19	SMPL
CCV020	CCV020	20990	4.9224 -	SFAM01.1	PPB	6/2/2022 16:21	SMPL
CCB020	CCB020	-117	-0.0274 -	SFAM01.1	PPB	6/2/2022 16:23	SMPL
N2892-13		26056	6.1104 -	SFAM01.1	PPB	6/2/2022 16:30	SMPL
N2890-01DLX2	YB7C9	29968	7.0278 -	SFAM01.1	PPB	6/2/2022 16:32	SMPL
N2890-02DLX10	YB7D1	19776	4.6377 -	SFAM01.1	PPB	6/2/2022 16:35	SMPL
N2890-03DLX10	YB7D3	22082	5.1785 -	SFAM01.1	PPB	6/2/2022 16:37	SMPL
N2890-11DLX5	YB7C6	22202	5.2066 -	SFAM01.1	PPB	6/2/2022 16:39	SMPL
N2890-12DLX10	YB7D6	58307	13.6736 -	SFAM01.1	PPB	6/2/2022 16:42	SMPL
N2890-13DLX10	YB7D6D	57370	13.4539 -	SFAM01.1	PPB	6/2/2022 16:44	SMPL
N2890-14DLX10	YB7D6S	66565	15.6102 -	SFAM01.1	PPB	6/2/2022 16:49	SMPL
N2890-15DLX25	YB7D8	11431	2.6807 -	SFAM01.1	PPB	6/2/2022 16:51	SMPL
N2890-16DLX5	YB7E3	14741	3.457 -	SFAM01.1	PPB	6/2/2022 16:54	SMPL
N2892-02DLX10	YB7D2	22249	5.2177 -	SFAM01.1	PPB	6/2/2022 16:56	SMPL
N2892-10DLX5	YB7C7	16722	3.9215 -	SFAM01.1	PPB	6/2/2022 17:00	SMPL
N2892-03DLX10	YB7D4	21455	5.0315 -	SFAM01.1	PPB	6/2/2022 17:03	SMPL
N2892-14DLX10	YB7D9	28187	6.6102 -	SFAM01.1	PPB	6/2/2022 17:06	SMPL
N2890-12DL2X50	YB7D6	10959	2.57 -	SFAM01.1	PPB	6/2/2022 17:08	SMPL
N2890-13DL2X50	YB7D6D	11301	2.6502 -	SFAM01.1	PPB	6/2/2022 17:10	SMPL
N2890-14DL2X50	YB7D6S	15126	3.5472 -	SFAM01.1	PPB	6/2/2022 17:13	SMPL
CCV021	CCV021	20728	4.861 -	SFAM01.1	PPB	6/2/2022 17:15	SMPL
CCB021	CCB021	9	0.0021 -	SFAM01.1	PPB	6/2/2022 17:17	SMPL