

LB 120467

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : DM Instrument ID : Konelab

6/2/2022 16:09

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.603	0.0	0.077	
ICB1	0.158	0.0	0.002	
CCV1	242.791	0.0	0.195	
CCB1	-0.054	0.0	0.001	
PB145286BL	-0.186	0.0	0.001	
N3102-13	-0.295	0.0	0.001	
N3102-13DUP	-0.162	0.0	0.001	
CCV2	254.280	0.0	0.204	
CCB2	-0.072	0.0	0.001	

N	9
Mean	65.674
SD	108.2466
CV%	164.83

Aquakem v. 7.2AQ1

Results from time period:

Thu Jun 02 15:19:26 2022

Thu Jun 02 15:23:19 2022

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.3405	µg/l	6/2/2022 11:08:28	
5.0PPBCN	A	Total CN	P	4.0277	µg/l	6/2/2022 11:08:29	
10PPBCN	A	Total CN	P	9.2608	µg/l	6/2/2022 11:08:30	
50PPBCN	A	Total CN	P	50.2353	µg/l	6/2/2022 11:08:31	
100PPBCN	A	Total CN	P	100.3319	µg/l	6/2/2022 11:08:32	
250PPBCN	A	Total CN	P	253.1003	µg/l	6/2/2022 11:08:33	
500PPBCN	A	Total CN	P	498.3845	µg/l	6/2/2022 11:08:34	
ICV1	S	Total CN	P	94.6026	µg/l	6/2/2022 15:19:26	
ICB1	S	Total CN	P	0.1577	µg/l	6/2/2022 15:19:29	
CCV1	S	Total CN	P	242.791	µg/l	6/2/2022 15:19:30	
CCB1	S	Total CN	P	-0.0543	µg/l	6/2/2022 15:19:32	
PB145286BL	S	Total CN	P	-0.1856	µg/l	6/2/2022 15:19:35	
N3102-13	S	Total CN	P	-0.2952	µg/l	6/2/2022 15:19:36	
N3102-13DUP	S	Total CN	P	-0.1618	µg/l	6/2/2022 15:23:17	
CCV2	S	Total CN	P	254.2805	µg/l	6/2/2022 15:23:18	
CCB2	S	Total CN	P	-0.0724	µg/l	6/2/2022 15:23:19	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

6/2/2022 11:11


Test Total CN

Accepted 6/2/2022 11:11

Factor 1256
Bias 0.001

Coeff. of det. 0.999932

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.3405	0.0000	-
2	5.0PPBCN	0.005	4.0277	5.0000	-18.4
3	10PPBCN	0.009	9.2608	10.0000	-7.4
4	50PPBCN	0.041	50.2353	50.0000	0.5
5	100PPBCN	0.081	100.3319	100.0000	0.3
6	250PPBCN	0.203	253.1003	250.0000	1.2
7	500PPBCN	0.398	498.3845	500.0000	-0.3

12 6/2/22