



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB120471

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/03/2022	08:20
2	CAL2	1	Water	NA	NA	20.7	7.00	06/03/2022	08:21
3	CAL3	1	Water	NA	NA	20.7	10.03	06/03/2022	08:22
4	ICV	1	Water	NA	NA	20.8	6.99	06/03/2022	08:25
5	CCV1	1	Water	NA	NA	20.6	2.01	06/03/2022	08:26
6	N3142-03	1	Water	NA	NA	22.1	6.54	06/03/2022	08:30
7	N3175-20	1	Water	NA	NA	24.3	4.97	06/03/2022	08:33
8	N3175-21	1	Water	NA	NA	24.7	6.88	06/03/2022	08:36
9	N3181-03	1	Water	NA	NA	24.7	6.53	06/03/2022	08:40
10	N3181-04	1	Water	NA	NA	23.2	5.74	06/03/2022	08:42
11	N3181-04DUP	1	Water	NA	NA	23.3	5.75	06/03/2022	08:43
12	CCV2	1	Water	NA	NA	20.7	12.03	06/03/2022	08:45

WORKLIST(Hardcopy Internal Chain)

VB 12047

WorkList Name : ph n3142 w WorkList ID : 160199 Department : Wet-Chemistry Date : 06-03-2022 08:12:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/01/2022	Water	N3142-03	pH	Cool 4 deg C	PSEG03	N11	40022	06/01/2022	9040C
06/02/2022	Water	N3175-20	pH	Cool 4 deg C	PSEG03	N11	AUD-1712	06/02/2022	9040C
06/02/2022	Water	N3175-21	pH	Cool 4 deg C	PSEG03	N11	AUD-1713	06/02/2022	9040C
06/02/2022	Water	N3181-03	pH	Cool 4 deg C	PSEG03	P11	TRE-1279	06/02/2022	9040C
06/02/2022	Water	N3181-04	pH	Cool 4 deg C	PSEG03	P11	TRE-31702	06/02/2022	9040C

Date/Time 06/03/22 08:15
 Raw Sample Received by: 20(WC)
 Raw Sample Relinquished by: KS(WC)

Date/Time 06/03/22 11:00
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: 20(WC)