



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB120472

Slope : 98.4

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/03/2022	08:20
2	CAL2	1	Water	NA	NA	20.7	7.00	06/03/2022	08:21
3	CAL3	1	Water	NA	NA	20.7	10.03	06/03/2022	08:22
4	ICV	1	Water	NA	NA	20.8	6.99	06/03/2022	08:25
5	CCV1	1	Water	NA	NA	20.8	2.02	06/03/2022	08:47
6	N3142-01	1	Solid	20.02	20	22.5	5.83	06/03/2022	08:50
7	N3142-01DUP	1	Solid	20.03	20	22.6	5.84	06/03/2022	08:51
8	CCV2	1	Water	NA	NA	20.7	12.02	06/03/2022	09:00

VR 120472

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n3142 s WorkList ID : 160200 Department : Wet-Chemistry Date : 06-03-2022 08:12:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/01/2022	Solid	N3142-01	pH	Cool 4 deg C	PSEG03	N11	40021	06/01/2022	9045D

Date/Time 06/03/22 08:15
Raw Sample Received by: JP(WC)
Raw Sample Relinquished by: KS(WC)

Date/Time 06/03/22 11:00
Raw Sample Received by: KS(WC)
Raw Sample Relinquished by: JP(WC)