



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : mohammad

Run Number: LB120478

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/03/2022	11:40
2	CAL2	1	Water	NA	NA	20.6	7.00	06/03/2022	11:41
3	CAL3	1	Water	NA	NA	20.6	10.03	06/03/2022	11:42
4	ICV	1	Water	NA	NA	20.7	6.99	06/03/2022	11:45
5	CCV1	1	Water	NA	NA	20.7	2.02	06/03/2022	11:46
6	N3102-13	1	Solid	20.02	20	22.6	6.85	06/03/2022	11:50
7	N3102-13DUP	1	Solid	20.03	20	22.7	6.86	06/03/2022	11:51
8	N3178-01	1	Solid	20.02	20	21.2	6.02	06/03/2022	11:55
9	CCV2	1	Water	NA	NA	20.7	12.03	06/03/2022	12:00

MS 120648

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n3178 WorkList ID : 160219 Department : Wet-Chemistry Date : 06-03-2022 11:32:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/26/2022	Solid	N3102-13	Corrosivity	Cool 4 deg C	AEIC01	N11	DISPOSAL	05/26/2022	9045D
06/02/2022	Solid	N3178-01	Corrosivity	Cool 4 deg C	PSEG03	N11	2329	06/02/2022	9045D

Date/Time 06/05/22 11:35
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 06/05/22 13:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]