



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : mohammad

Run Number: LB120479

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	06/03/2022	11:40
2	CAL2	1	Water	NA	NA	20.6	7.00	06/03/2022	11:41
3	CAL3	1	Water	NA	NA	20.6	10.03	06/03/2022	11:42
4	ICV	1	Water	NA	NA	20.7	6.99	06/03/2022	11:45
5	CCV1	1	Water	NA	NA	20.7	2.02	06/03/2022	11:46
6	N3167-02	1	Water	NA	NA	21.2	6.80	06/03/2022	12:10
7	N3167-02DUP	1	Water	NA	NA	21.3	6.81	06/03/2022	12:11
8	N3178-02	1	Water	NA	NA	20.3	6.85	06/03/2022	12:15
9	N3178-03	1	Water	NA	NA	20.2	8.70	06/03/2022	12:20
10	CCV2	1	Water	NA	NA	20.7	12.03	06/03/2022	12:22

W5 128479

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n3178 W WorkList ID : 160220 Department : Wet-Chemistry Date : 06-03-2022 11:33:02

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Raw Sample Storage Location		
06/02/2022	Water	N3167-02	Corrosivity	Cool 4 deg C	PSEG03	N11	BLUE-DRUM	06/02/2022 9040C
06/02/2022	Water	N3178-02	Corrosivity	Cool 4 deg C	PSEG03	N11	2330-2332	06/02/2022 9040C
06/02/2022	Water	N3178-03	Corrosivity	Cool 4 deg C	PSEG03	N11	2342	06/02/2022 9040C

Date/Time 06/03/22 11:35
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 06/03/22 13:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]