

LB120483


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42732



A= 0.0000e+000

B= 2.3718e-004

C= -4.3532e-002

Rho= 0.9992275

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.046	-0.046	-9	0.000	-9					
0.2	0.200	0.162	-0.038	866	0.0 %	866					-19
2.5	2.500	2.620	0.120	11231	0.0 %	11231					5
5.0	5.000	5.150	0.150	21897	0.0 %	21897					3
7.5	7.500	7.222	-0.278	30633	0.0 %	30633					-4
10.0	10.000	10.092	0.092	42732	0.0 %	42732					1

LB120483 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Jun 2022 13:29:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-9	-	0.0000	7470A	PPB	03 Jun 2022 13:37:54	S	Std
0.2 - 1	502	866	-	0.2000	7470A	PPB	03 Jun 2022 13:40:10	S	Std
2.5 - 1	52.5	11231	-	2.5000	7470A	PPB	03 Jun 2022 13:42:28	S	Std
5.0 - 1	55	21897	-	5.0000	7470A	PPB	03 Jun 2022 13:44:44	S	Std
7.5 - 1	57.5	30633	-	7.5000	7470A	PPB	03 Jun 2022 13:47:00	S	Std
10.0 - 1	510	42732	-	10.0000	7470A	PPB	03 Jun 2022 13:49:17	S	Std
ICV58 - 1	ICV58	15844	3.7143	-	7470A	PPB	03 Jun 2022 13:52:40	U	SMPL
ICB58 - 1	ICB58	-93	-0.0656	-	7470A	PPB	03 Jun 2022 13:54:55	U	SMPL
CCV22 - 1	CCV22	21662	5.0942	-	7470A	PPB	03 Jun 2022 13:57:11	U	SMPL
CCB22 - 1	CCB22	-85	-0.0637	-	7470A	PPB	03 Jun 2022 13:59:26	U	SMPL
CRA - 1	CRA	1016	0.1974	-	7470A	PPB	03 Jun 2022 14:01:42	U	SMPL
HighStd - 1	HighStd	41908	9.8962	-	7470A	PPB	03 Jun 2022 14:03:57	U	SMPL
ChkStd - 1	ChkStd	26974	6.3541	-	7470A	PPB	03 Jun 2022 14:06:12	U	SMPL
PB145312BL - 1	PBW	-131	-0.0746	-	7470A	PPB	03 Jun 2022 14:08:30	U	SMPL
PB145312BS - 1	LCSW	16481	3.8654	-	7470A	PPB	03 Jun 2022 14:10:46	U	SMPL
N3104-02 - 1	WC-6	-101	-0.0675	-	7470A	PPB	03 Jun 2022 14:13:02	U	SMPL
N3104-02DUP - 1	WC-6DUP	-23	-0.0490	-	7470A	PPB	03 Jun 2022 14:15:19	U	SMPL
N3104-02MS - 1	WC-6MS	15366	3.6010	-	7470A	PPB	03 Jun 2022 14:17:36	U	SMPL
N3104-02MSD - 1	WC-6MSD	14546	3.4065	-	7470A	PPB	03 Jun 2022 14:19:53	U	SMPL
N3104-05 - 1	WC-7	-66	-0.0592	-	7470A	PPB	03 Jun 2022 14:22:08	U	SMPL
CCV23 - 1	CCV23	22229	5.2287	-	7470A	PPB	03 Jun 2022 14:24:23	U	SMPL
CCB23 - 1	CCB23	-92	-0.0654	-	7470A	PPB	03 Jun 2022 14:26:38	U	SMPL
N3107-06 - 1	MS-5	-42	-0.0535	-	7470A	PPB	03 Jun 2022 14:28:54	U	SMPL
N3107-07 - 1	MS-6	-30	-0.0506	-	7470A	PPB	03 Jun 2022 14:31:10	U	SMPL
N3109-10 - 1	BB-8	-26	-0.0497	-	7470A	PPB	03 Jun 2022 14:33:25	U	SMPL
N3109-11 - 1	BB-9	-16	-0.0473	-	7470A	PPB	03 Jun 2022 14:35:40	U	SMPL
N3109-12 - 1	BB-10	-29	-0.0504	-	7470A	PPB	03 Jun 2022 14:37:56	U	SMPL
N3109-13 - 1	BB-11	-29	-0.0504	-	7470A	PPB	03 Jun 2022 14:40:12	U	SMPL
PB145213TB - 1	PB145213TB	-22	-0.0488	-	7470A	PPB	03 Jun 2022 14:42:28	U	SMPL
N3104-02LX5 - 1		-69	-0.0599	-	7470A	PPB	03 Jun 2022 14:44:45	U	SMPL
N3104-02A - 1		16815	3.9446	-	7470A	PPB	03 Jun 2022 14:47:02	U	SMPL
CCV24 - 1	CCV24	21911	5.1533	-	7470A	PPB	03 Jun 2022 14:49:17	U	SMPL
CCB24 - 1	CCB24	-96	-0.0663	-	7470A	PPB	03 Jun 2022 14:51:33	U	SMPL