

LB120514


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39006



A= 0.0000e+000

B= 2.5884e-004

C= -1.3177e-003

Rho= 0.9998580

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.009	0.009	40	0.000	40					-
0.05	0.050	0.060	0.010	237	0.0 %	237	0				20
0.2	0.200	0.198	-0.002	770	0.0 %	770					-1
2.5	2.500	2.547	0.047	9845	0.0 %	9845					2
5.0	5.000	4.961	-0.039	19171	0.0 %	19171					-1
7.5	7.500	7.380	-0.120	28516	0.0 %	28516					-2
10.0	10.000	10.095	0.095	39006	0.0 %	39006					1

LB120514 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 05 Jun 2022 08:49:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.0000	245.1	PPB	05 Jun 2022 09:39:12	S	Std
0.05 - 1	50.25	237	-	0.0500	245.1	PPB	05 Jun 2022 09:49:18	S	Std
0.2 - 1	50.2	770	-	0.2000	245.1	PPB	05 Jun 2022 09:51:35	S	Std
2.5 - 1	52.5	9845	-	2.5000	245.1	PPB	05 Jun 2022 09:53:52	S	Std
5.0 - 1	55	19171	-	5.0000	245.1	PPB	05 Jun 2022 09:56:09	S	Std
7.5 - 1	57.5	28516	-	7.5000	245.1	PPB	05 Jun 2022 09:58:25	S	Std
10.0 - 1	510	39006	-	10.0000	245.1	PPB	05 Jun 2022 10:00:41	S	Std
ICV61 - 1	ICV61	15742	4.0734	-245.1		PPB	05 Jun 2022 10:08:53	U	SMPL
ICB61 - 1	ICB61	-49	-0.0140	-245.1		PPB	05 Jun 2022 10:11:08	U	SMPL
CCV38 - 1	CCV38	19020	4.9219	-245.1		PPB	05 Jun 2022 10:13:24	U	SMPL
CCB38 - 1	CCB38	-50	-0.0143	-245.1		PPB	05 Jun 2022 10:15:39	U	SMPL
CRA - 1	CRA	236	0.0598	-245.1		PPB	05 Jun 2022 10:20:33	U	SMPL
HighStd - 1	HighStd	38983	10.0892	-245.1		PPB	05 Jun 2022 10:22:48	U	SMPL
ChkStd - 1	ChkStd	27182	7.0346	-245.1		PPB	05 Jun 2022 10:25:03	U	SMPL
PB145342BL - 1	PBW	-107	-0.0290	-245.1		PPB	05 Jun 2022 10:27:21	U	SMPL
PB145342BS - 1	LCSW	15416	3.9890	-245.1		PPB	05 Jun 2022 10:29:38	U	SMPL
N3123-07 - 1	14B-(1-4)-COMP	890	0.2291	-245.1		PPB	05 Jun 2022 10:31:54	U	SMPL
N3123-07DUP - 1	14B-(1-4)-COMP DUP	922	0.2373	-245.1		PPB	05 Jun 2022 10:34:11	U	SMPL
N3123-07MS - 1	14B-(1-4)-COMP MS	15679	4.0571	-245.1		PPB	05 Jun 2022 10:36:28	U	SMPL
N3123-07MSD - 1	14B-(1-4)-COMP MSD	14975	3.8749	-245.1		PPB	05 Jun 2022 10:38:45	U	SMPL
N3157-07 - 1	14B-(1-4)-COMP	1391	0.3587	-245.1		PPB	05 Jun 2022 10:41:00	U	SMPL
CCV39 - 1	CCV39	19247	4.9806	-245.1		PPB	05 Jun 2022 10:43:17	U	SMPL
CCB39 - 1	CCB39	47	0.0108	-245.1		PPB	05 Jun 2022 10:45:32	U	SMPL
N3172-07 - 1	14B-(1-4)COMP	1775	0.4581	-245.1		PPB	05 Jun 2022 10:47:48	U	SMPL
N3181-03 - 1	TRE-1279	56902	14.7274	-245.1		PPB	05 Jun 2022 10:50:02	U	SMPL
N3181-04 - 1	TRE-31702	181	0.0455	-245.1		PPB	05 Jun 2022 10:56:29	U	SMPL
N3198-07 - 1	14B-(1-4)COMP	1304	0.3362	-245.1		PPB	05 Jun 2022 10:58:44	U	SMPL
N3123-07LX5 - 1		122	0.0303	-245.1		PPB	05 Jun 2022 11:01:00	U	SMPL
N3123-07A - 1		15592	4.0346	-245.1		PPB	05 Jun 2022 11:03:16	U	SMPL
CCV40 - 1	CCV40	18848	4.8774	-245.1		PPB	05 Jun 2022 11:05:33	U	SMPL
CCB40 - 1	CCB40	-9	-0.0036	-245.1		PPB	05 Jun 2022 11:07:50	U	SMPL
N3181-03DLX5 - 1		10671	2.7608	-245.1		PPB	05 Jun 2022 11:34:02	U	SMPL
CCV41 - 1	CCV41	19146	4.9545	-245.1		PPB	05 Jun 2022 11:36:17	U	SMPL
CCB41 - 1	CCB41	11	0.0015	-245.1		PPB	05 Jun 2022 11:38:34	U	SMPL