


LB120575

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.0798e-004 *slope*C= -1.2723e-001 *intercept*

Rho= 0.9999645

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.003	0.003	624	0.000	624					
0.05	0.050					185	0				
0.20	0.200	0.224	0.024	1690	0.0 %	1690					
2.50	2.500	2.492	-0.008	12593	0.0 %	12593					
5.00	5.000	4.987	-0.013	24589	0.0 %	24589					
7.50	7.500	7.447	-0.053	36419	0.0 %	36419					
10.0	10.000	10.047	0.047	48918	0.0 %	48918					

LB120515
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	624	-		0 SFAM01.1	PPB	6/5/2022 12:11	Std
	0.2 S01	1690	-		0.2 SFAM01.1	PPB	6/5/2022 12:13	Std
	2.5 S02	12593	-		2.5 SFAM01.1	PPB	6/5/2022 12:16	Std
	5 S03	24589	-		5 SFAM01.1	PPB	6/5/2022 12:23	Std
	7.5 S04	36419	-		7.5 SFAM01.1	PPB	6/5/2022 12:25	Std
	10 S05	48918	-		10 SFAM01.1	PPB	6/5/2022 12:32	Std
ICV062	ICV062	19073	3.8397 -		SFAM01.1	PPB	6/5/2022 12:35	SMPL
ICB062	ICB062	546	-0.0137 -		SFAM01.1	PPB	6/5/2022 12:37	SMPL
CCV042	CCV042	24746	5.0196 -		SFAM01.1	PPB	6/5/2022 12:40	SMPL
CCB042	CCB042	546	-0.0137 -		SFAM01.1	PPB	6/5/2022 12:42	SMPL
PB145341BL	PBS341	644	0.0067 -		SFAM01.1	PPB	6/5/2022 12:47	SMPL
N2947-03	MDBTA7	1224	0.1273 -		SFAM01.1	PPB	6/5/2022 12:56	SMPL
N2947-04	MDBTA8	879	0.0556 -		SFAM01.1	PPB	6/5/2022 12:58	SMPL
N2947-05	MDBTA9	1118	0.1053 -		SFAM01.1	PPB	6/5/2022 13:00	SMPL
N2947-06	MDBTA9D	1133	0.1084 -		SFAM01.1	PPB	6/5/2022 13:03	SMPL
N2947-07	MDBTA9S	12850	2.5454 -		SFAM01.1	PPB	6/5/2022 13:05	SMPL
N2947-08	MDBTB0	2528	0.3986 -		SFAM01.1	PPB	6/5/2022 13:07	SMPL
N2947-01	MDBTA5	12149	2.3996 -		SFAM01.1	PPB	6/5/2022 13:17	SMPL
N2947-02	MDBTA6	442305	HIGH -		SFAM01.1	PPB	6/5/2022 13:23	
CCV043	CCV043	24069	4.8788 -		SFAM01.1	PPB	6/5/2022 13:25	SMPL
CCB043	CCB043	465	-0.0305 -		SFAM01.1	PPB	6/5/2022 13:27	SMPL
N2947-02DLX10	MDBTA6	61168	12.5948 -		SFAM01.1	PPB	6/5/2022 13:36	SMPL
N2947-02DL2X25	MDBTA6	24753	5.021 -		SFAM01.1	PPB	6/5/2022 13:45	SMPL
CCV044	CCV044	25064	5.0857 -		SFAM01.1	PPB	6/5/2022 13:48	SMPL
CCB044	CCB044	548	-0.0133 -		SFAM01.1	PPB	6/5/2022 13:50	SMPL