


LB120516

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4637e-004

C= -4.5770e-002

Rho= 0.9975783

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.002	0.002	194	0.000	194					2
0.2	0.200	0.203	0.003	1008	0.0 %	1008					-3
2.5	2.500	2.418	-0.082	9999	0.0 %	9999					-1
5.0	5.000	4.943	-0.057	20248	0.0 %	20248					7
7.5	7.500	8.001	0.501	32661	0.0 %	32661					-4
10.0	10.000	9.634	-0.366	39288	0.0 %	39288					

LB120516 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Jun 2022 15:14:40

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type	Type
0.0 - 1	50	194	-	0.0000	7470A	PPB	05 Jun 2022 15:16:20	S	Std
0.2 - 1	50.2	1008	-	0.2000	7470A	PPB	05 Jun 2022 15:18:36	S	Std
2.5 - 1	52.5	9999	-	2.5000	7470A	PPB	05 Jun 2022 15:20:53	S	Std
5.0 - 1	55	20248	-	5.0000	7470A	PPB	05 Jun 2022 15:25:37	S	Std
7.5 - 1	57.5	32661	-	7.5000	7470A	PPB	05 Jun 2022 15:27:53	S	Std
10.0 - 1	510	39288	-	10.0000	7470A	PPB	05 Jun 2022 15:30:09	S	Std
ICV63 - 1	ICV63	15026	3.6563	-	7470A	PPB	05 Jun 2022 15:33:34	U	SMPL
ICB63 - 1	ICB63	129	-0.0140	-	7470A	PPB	05 Jun 2022 15:35:49	U	SMPL
CCV45 - 1	CCV45	20151	4.9189	-	7470A	PPB	05 Jun 2022 15:38:04	U	SMPL
CCB45 - 1	CCB45	160	-0.0063	-	7470A	PPB	05 Jun 2022 15:40:19	U	SMPL
CRA - 1	CRA	935	0.1846	-	7470A	PPB	05 Jun 2022 15:42:34	U	SMPL
HighStd - 1	HighStd	39941	9.7947	-	7470A	PPB	05 Jun 2022 15:44:49	U	SMPL
ChkStd - 1	ChkStd	29821	7.3014	-	7470A	PPB	05 Jun 2022 15:47:04	U	SMPL
PB145343BL - 1	PBW	19	-0.0411	-	7470A	PPB	05 Jun 2022 15:49:21	U	SMPL
PB145343BS - 1	LCSW	14413	3.5052	-	7470A	PPB	05 Jun 2022 15:51:38	U	SMPL
N3189-05 - 1	20220423-SPLP-AMENDED-COMPOSITE-C	1878	0.4169	-	7470A	PPB	05 Jun 2022 15:53:54	U	SMPL
N3189-05DUP - 1	20220423-SPLP-AMENDED-COMPOSITE-CDUP	1894	0.4209	-	7470A	PPB	05 Jun 2022 15:56:11	U	SMPL
N3189-05MS - 1	20220423-SPLP-AMENDED-COMPOSITE-CMS	17390	4.2387	-	7470A	PPB	05 Jun 2022 15:58:28	U	SMPL
N3189-05MSD - 1	20220423-SPLP-AMENDED-COMPOSITE-CMSD	17461	4.2562	-	7470A	PPB	05 Jun 2022 16:00:45	U	SMPL
N3189-06 - 1	20220428-SPLP-AMENDED-COMPOSITE-H	570	0.0947	-	7470A	PPB	05 Jun 2022 16:03:00	U	SMPL
CCV46 - 1	CCV46	20498	5.0044	-	7470A	PPB	05 Jun 2022 16:05:16	U	SMPL
CCB46 - 1	CCB46	166	-0.0049	-	7470A	PPB	05 Jun 2022 16:07:31	U	SMPL
N3189-05A - 1		18406	4.4890	-	7470A	PPB	05 Jun 2022 16:15:39	U	SMPL
PB145343TB - 1	PB145328TB	35	-0.0371	-	7470A	PPB	05 Jun 2022 16:17:54	U	SMPL
N3189-05LX5 - 1		2	-0.0453	-	7470A	PPB	05 Jun 2022 16:26:46	U	SMPL
CCV47 - 1	CCV47	20813	5.0820	-	7470A	PPB	05 Jun 2022 16:29:14	U	SMPL
CCB47 - 1	CCB47	131	-0.0135	-	7470A	PPB	05 Jun 2022 16:31:29	U	SMPL