

TitriSoft Report

Ident	Amount	DateTime	Alkalinity	Final pH (sample)	Initial pH (sample)	slope	offset
LB120517BLW	100	06/06/2022 11:13	0.00	3.33	3.32	102.8	7.08
LB120517BSW	100	06/06/2022 11:21	47.52	4.40	8.10	102.8	7.08
N3155-01	50	06/06/2022 11:26	712.00	4.31	7.63	102.8	7.08
N3166-11	50	06/06/2022 11:33	43.68	4.36	5.55	102.8	7.08
N3166-15	50	06/06/2022 11:38	174.96	4.35	6.92	102.8	7.08
N3166-17	50	06/06/2022 11:42	0.00	4.30	4.33	102.8	7.08
N3166-11DUP	50	06/06/2022 11:45	44.56	4.37	5.84	102.8	7.08

WORKLIST(Hardcopy Internal Chain)

LB120517

WorkList Name : Alkalinity-060222

WorkList ID : 160177

Department : Wet-Chemistry

Date : 06-02-2022 10:43:31

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
06/14/2022	Water	N3155-01	Alkalinity	Cool 4 deg C	INFR01	N11 MONTHLY-OUTFALL-002	06/01/2022	SM2320 B
06/14/2022	Water	N3166-11	Alkalinity	Cool 4 deg C	TETR06	N11 11BR-20220601	06/01/2022	SM2320 B
06/14/2022	Water	N3166-15	Alkalinity	Cool 4 deg C	TETR06	N11 38BR-20220601	06/01/2022	SM2320 B
06/14/2022	Water	N3166-17	Alkalinity	Cool 4 deg C	TETR06	N11 RB07-20220601 LB 120520	06/01/2022	SM2320 B

Date/Time 6/6/22 10:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 6/6/22 12:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]