



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120526

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/06/2022	15:15
2	CAL2	1	Water	NA	NA	20.1	7.00	06/06/2022	15:16
3	CAL3	1	Water	NA	NA	20.2	10.03	06/06/2022	15:17
4	ICV	1	Water	NA	NA	20.3	6.98	06/06/2022	15:20
5	CCV1	1	Water	NA	NA	20.2	2.01	06/06/2022	15:22
6	N3215-01	1	Water	NA	NA	20.6	6.78	06/06/2022	15:25
7	N3215-01DUP	1	Water	NA	NA	20.7	6.79	06/06/2022	15:26
8	N3222-01	1	Water	NA	NA	20.6	6.37	06/06/2022	15:30
9	N3222-02	1	Water	NA	NA	20.5	5.56	06/06/2022	15:32
10	N3222-03	1	Water	NA	NA	20.1	6.40	06/06/2022	15:35
11	CCV2	1	Water	NA	NA	20.3	12.02	06/06/2022	15:36

WORKLIST(Hardcopy Internal Chain)

VB 120526

WorkList Name : PH N3215 WorkList ID : 160297 Department : Wet-Chemistry Date : 06-06-2022 15:02:51

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
06/06/2022	Water	N3215-01	pH	Cool 4 deg C	PSEG03	P11 N48964	06/06/2022	9040C
05/27/2022	Water	N3222-01	pH	Cool 4 deg C	STER09	P11 CELL-A-POOL-WATER(DE-IONIZED)	05/27/2022	9040C
05/27/2022	Water	N3222-02	pH	Cool 4 deg C	STER09	P11 CELL-B/C-POOL-WATER(DE-IONIZED)	05/27/2022	9040C
05/27/2022	Water	N3222-03	pH	Cool 4 deg C	STER09	P11 MINI-CELL-POOL-WATER(DE-IONIZED)	05/27/2022	9040C

Date/Time 06/06/22 15:16
 Raw Sample Received by: JP (SM)
 Raw Sample Relinquished by: JP (SM)

Date/Time 06/06/22 16:35
 Raw Sample Received by: JP (SM)
 Raw Sample Relinquished by: JP (SM)