



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 103 °C 06/06/2022 13:00 TEMP1 OUT: 103 °C 06/06/2022 14:00
TEMP2 IN: 104 °C 06/06/2022 16:00 TEMP2 OUT: 103 °C 06/07/2022 08:00
TEMP3 IN: 104 °C 06/07/2022 08:30 TEMP3 OUT: 104 °C 06/07/2022 10:00
TEMP4 IN: 540 °C 06/07/2022 10:30 TEMP4 OUT: 540 °C 06/07/2022 12:00

Run Number: LB120528

SUPERVISOR: mohammad

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB120528BL	89.0175	89.0175	100	89.0175	89.0175	89.0175	89.0175	0.0000	0
2	N3215-01	92.5059	92.5059	100	92.5537	92.5537	92.5318	92.5318	0.0219	219
3	N3215-01DUP	91.5406	91.5406	100	91.5886	91.5886	91.5671	91.5671	0.0215	215

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

B 120528

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS N3215 WorkList ID : 160299 Department : Wet-Chemistry Date : 06-06-2022 15:04:21

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/09/2022	Water	N3215-01	TVS	Cool 4 deg C	PSEG03	P11	N48964	06/06/2022	SM2540 G

Date/Time 06/06/22 15:10
Raw Sample Received by: JH (coll)
Raw Sample Relinquished by: CP SM

Date/Time 06/06/22 16:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JH (coll)