



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : ketankumar

Run Number: LB120532

Slope : 98.8

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	06/06/2022	15:10
2	CAL2	1	Water	NA	NA	20.8	7.00	06/06/2022	15:11
3	CAL3	1	Water	NA	NA	20.8	10.02	06/06/2022	15:12
4	ICV	1	Water	NA	NA	20.7	6.99	06/06/2022	15:15
5	CCV1	1	Water	NA	NA	20.7	2.01	06/06/2022	15:20
6	N3028-04	1	Solid	20.03	20	20.9	7.50	06/06/2022	15:22
7	N3194-02	1	Solid	20.02	20	21.2	7.17	06/06/2022	15:25
8	N3194-05	1	Solid	20.03	20	20.2	7.80	06/06/2022	15:28
9	N3209-02	1	Solid	20.01	20	20.2	6.50	06/06/2022	15:30
10	N3209-07	1	Solid	20.02	20	21.4	8.56	06/06/2022	15:33
11	N3209-12	1	Solid	20.04	20	21.1	10.89	06/06/2022	15:40
12	N3209-12DUP	1	Solid	20.03	20	21.2	10.90	06/06/2022	15:41
13	CCV2	1	Water	NA	NA	20.7	12.03	06/06/2022	15:43

WB 120532

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n3209 WorkList ID : 160285 Department : Wet-Chemistry Date : 06-06-2022 11:28:48

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
05/24/2022	Solid	N3028-04	Corrosivity	Cool 4 deg C	PSEG03	N11 TP-2-1-5FT	05/24/2022	9045D
06/01/2022	Solid	N3194-02	Corrosivity	Cool 4 deg C	PENN09	N11 TP-1	06/01/2022	9045D
06/01/2022	Solid	N3194-05	Corrosivity	Cool 4 deg C	PENN09	N11 TP-3	06/01/2022	9045D
06/03/2022	Solid	N3209-02	Corrosivity	Cool 4 deg C	ROUX02	P11 WC-14B-4	06/03/2022	9045D
06/03/2022	Solid	N3209-07	Corrosivity	Cool 4 deg C	ROUX02	P11 WC-14B-10	06/03/2022	9045D
06/03/2022	Solid	N3209-12	Corrosivity	Cool 4 deg C	ROUX02	P11 WC-14B-2-1	06/03/2022	9045D

Date/Time 06/06/22 15:00
Raw Sample Received by: 70160C
Raw Sample Relinquished by: KS(wc)

Date/Time 06/06/22 17:00
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: 70160C