


LB 120544

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.2944e-004

C= 3.3007e-002

Rho= 0.9996594

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	-89	0.000	-89				
0.2	0.200	0.197	-0.003	713	0.0 %	713				
2.5	2.500	2.534	0.034	10901	0.0 %	10901				
5.0	5.000	5.041	0.041	21826	0.0 %	21826				
7.5	7.500	7.301	-0.199	31678	0.0 %	31678				
10.0	10.000	10.115	0.115	43940	0.0 %	43940				

LB120544 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Jun 2022 14:10:46					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-89	-	0.0000	7470A	PPB	07 Jun 2022 14:14:22	S	Std
0.2 - 1	512	713	-	0.2000	7470A	PPB	07 Jun 2022 14:16:39	S	Std
2.5 - 1	52-5	10901	-	2.5000	7470A	PPB	07 Jun 2022 14:18:56	S	Std
5.0 - 1	53	21826	-	5.0000	7470A	PPB	07 Jun 2022 14:21:13	S	Std
7.5 - 1	5715	31678	-	7.5000	7470A	PPB	07 Jun 2022 14:23:29	S	Std
10.0 - 1	510	43940	-	10.0000	7470A	PPB	07 Jun 2022 14:37:00	S	Std
ICV65 - 1	ICV65	18285	4.2283	-	7470A	PPB	07 Jun 2022 14:42:33	U	SMPL
ICB65 - 1	ICB65	-139	0.0011	-	7470A	PPB	07 Jun 2022 14:44:47	U	SMPL
CCV52 - 1	CCV52	20839	4.8143	-	7470A	PPB	07 Jun 2022 14:47:03	U	SMPL
CCB52 - 1	CCB52	-151	-0.0016	-	7470A	PPB	07 Jun 2022 14:49:17	U	SMPL
CRA - 1	CRA	691	0.1916	-	7470A	PPB	07 Jun 2022 14:51:33	U	SMPL
HighStd - 1	HighStd	43738	10.0683	-	7470A	PPB	07 Jun 2022 14:53:48	U	SMPL
ChkStd - 1	ChkStd	29190	6.7304	-	7470A	PPB	07 Jun 2022 14:56:03	U	SMPL
PB145388BL - 1	PBW	-154	-0.0023	-	7470A	PPB	07 Jun 2022 14:58:21	U	SMPL
PB145388BS - 1	LCSW	16912	3.9133	-	7470A	PPB	07 Jun 2022 15:00:37	U	SMPL
N3028-04 - 1	TP-2-1-5FT	-18	0.0289	-	7470A	PPB	07 Jun 2022 15:02:53	U	SMPL
N3189-07 - 1	20220399-C-1-TCLP	-32	0.0257	-	7470A	PPB	07 Jun 2022 15:05:10	U	SMPL
N3189-08 - 1	20220400-C-2-TCLP	-38	0.0243	-	7470A	PPB	07 Jun 2022 15:07:27	U	SMPL
N3194-02 - 1	TP-1	-53	0.0208	-	7470A	PPB	07 Jun 2022 15:09:44	U	SMPL
N3194-05 - 1	TP-3	-33	0.0254	-	7470A	PPB	07 Jun 2022 15:11:59	U	SMPL
CCV53 - 1	CCV53	20782	4.8013	-	7470A	PPB	07 Jun 2022 15:14:14	U	SMPL
CCB53 - 1	CCB53	-161	-0.0039	-	7470A	PPB	07 Jun 2022 15:16:30	U	SMPL
N3203-02 - 1	GATE-6-JET-GROUT-SPOILS	-51	0.0213	-	7470A	PPB	07 Jun 2022 15:18:45	U	SMPL
N3209-02 - 1	WC-14B-4	-34	0.0252	-	7470A	PPB	07 Jun 2022 15:21:00	U	SMPL
N3209-05 - 1	WC-14B-4-GR	-16	0.0293	-	7470A	PPB	07 Jun 2022 15:23:15	U	SMPL
N3209-07 - 1	WC-14B-10	-41	0.0236	-	7470A	PPB	07 Jun 2022 15:25:31	U	SMPL
N3209-10 - 1	WC-14B-10-GR	-33	0.0254	-	7470A	PPB	07 Jun 2022 15:27:47	U	SMPL
N3209-12 - 1	WC-14B-2-1	9	0.0351	-	7470A	PPB	07 Jun 2022 15:30:03	U	SMPL
N3209-15 - 1	WC-14B-2-1-GR	-25	0.0273	-	7470A	PPB	07 Jun 2022 15:32:20	U	SMPL
N3209-15DUP - 1	WC-14B-2-1-GRDUP	73	0.0498	-	7470A	PPB	07 Jun 2022 15:34:36	U	SMPL
N3209-15MS - 1	WC-14B-2-1-GRMS	14112	3.2709	-	7470A	PPB	07 Jun 2022 15:36:53	U	SMPL
N3209-15MSD - 1	WC-14B-2-1-GRMSD	14238	3.2998	-	7470A	PPB	07 Jun 2022 15:39:09	U	SMPL
CCV54 - 1	CCV54	20948	4.8393	-	7470A	PPB	07 Jun 2022 15:41:25	U	SMPL
CCB54 - 1	CCB54	-171	-0.0062	-	7470A	PPB	07 Jun 2022 15:43:40	U	SMPL
PB145389BL - 1	PBW	-99	0.0103	-	7470A	PPB	07 Jun 2022 15:45:56	U	SMPL
PB145389BS - 1	LCSW	18008	4.1648	-	7470A	PPB	07 Jun 2022 15:51:06	U	SMPL
N3215-01 - 1	N48964	51	0.0447	-	7470A	PPB	07 Jun 2022 15:53:21	U	SMPL
N3215-01DUP - 1	N48964DUP	68	0.0486	-	7470A	PPB	07 Jun 2022 15:55:37	U	SMPL
N3215-01MS - 1	N48964MS	14157	3.2812	-	7470A	PPB	07 Jun 2022 15:57:52	U	SMPL
N3215-01MSD - 1	N48964MSD	14393	3.3354	-	7470A	PPB	07 Jun 2022 16:00:08	U	SMPL
N3223-01 - 1	M-15328	-188	-0.0101	-	7470A	PPB	07 Jun 2022 16:02:25	U	SMPL
PB145354TB - 1		-50	0.0215	-	7470A	PPB	07 Jun 2022 16:04:42	U	SMPL
CCV55 - 1	CCV55	20912	4.8311	-	7470A	PPB	07 Jun 2022 16:06:59	U	SMPL
CCB55 - 1	CCB55	-136	0.0018	-	7470A	PPB	07 Jun 2022 16:09:15	U	SMPL
N3209-15LX5 - 1		-50	0.0215	-	7470A	PPB	07 Jun 2022 16:11:30	U	SMPL
N3209-15A - 1		17613	4.0742	-	7470A	PPB	07 Jun 2022 16:13:46	U	SMPL
N3215-01LX5 - 1		-151	-0.0016	-	7470A	PPB	07 Jun 2022 16:16:01	U	SMPL
N3215-01A - 1		19057	4.4055	-	7470A	PPB	07 Jun 2022 16:18:17	U	SMPL
CCV56 - 1	CCV56	21300	4.9201	-	7470A	PPB	07 Jun 2022 16:23:22	U	SMPL
CCB56 - 1	CCB56	-163	-0.0044	-	7470A	PPB	07 Jun 2022 16:25:37	U	SMPL