



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120582

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	06/09/2022	08:42
2	CAL2	1	Water	NA	NA	20.8	7.00	06/09/2022	08:43
3	CAL3	1	Water	NA	NA	20.8	10.03	06/09/2022	08:44
4	ICV	1	Water	NA	NA	20.7	6.99	06/09/2022	08:45
5	CCV1	1	Water	NA	NA	20.7	2.01	06/09/2022	08:46
6	N3252-01	1	Water	NA	NA	24.1	4.99	06/09/2022	08:48
7	N3263-01	1	Water	NA	NA	21.4	7.40	06/09/2022	08:50
8	N3263-03	1	Water	NA	NA	20.6	6.98	06/09/2022	08:52
9	N3263-05	1	Water	NA	NA	20.8	7.05	06/09/2022	09:00
10	N3266-01	1	Water	NA	NA	24.5	7.98	06/09/2022	09:02
11	N3266-01DUP	1	Water	NA	NA	24.6	7.99	06/09/2022	09:05
12	CCV2	1	Water	NA	NA	20.7	12.02	06/09/2022	09:10

W3120582

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n3266 WorkList ID : 160378 Department : Wet-Chemistry Date : 06-09-2022 08:30:01

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Raw Sample Storage Location		
06/07/2022	Water	N3252-01	pH	Cool 4 deg C	STER01	N11	POOL-WATER	06/07/2022 9040C
06/08/2022	Water	N3263-01	pH	Cool 4 deg C	PSEG04	P11	DSW002	06/08/2022 9040C
06/08/2022	Water	N3263-03	pH	Cool 4 deg C	PSEG04	P11	DSW001	06/08/2022 9040C
06/08/2022	Water	N3263-05	pH	Cool 4 deg C	PSEG04	P11	RWS003	06/08/2022 9040C
06/08/2022	Water	N3266-01	pH	Cool 4 deg C	PSEG03	K11	OH-WATER-0608	06/08/2022 9040C

Date/Time 06/09/22 08:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 06/09/22 10:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]