



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : mohammad

Run Number: LB120592

Slope : 98.9

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	06/09/2022	14:00
2	CAL2	1	Water	NA	NA	20.9	7.00	06/09/2022	14:01
3	CAL3	1	Water	NA	NA	20.9	10.03	06/09/2022	14:02
4	ICV	1	Water	NA	NA	20.8	6.98	06/09/2022	14:05
5	CCV1	1	Water	NA	NA	20.8	2.01	06/09/2022	14:09
6	N3234-03	1	Solid	20.03	20	21.3	7.56	06/09/2022	14:10
7	N3234-03DUP	1	Solid	20.04	20	21.4	7.57	06/09/2022	14:11
8	N3246-02	1	Solid	20.02	20	24.0	6.49	06/09/2022	14:15
9	N3246-03	1	Solid	20.02	20	23.3	5.61	06/09/2022	14:20
10	CCV2	1	Water	NA	NA	20.9	12.03	06/09/2022	14:22