



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 6/10/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 16:25

In Date: 06/09/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:20

Out Date: 06/10/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB120598

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3226-01	EW4L2	1	0.91	8.92	9.83	6.9	67.2	
N3226-02	EW4L3	2	0.93	8.71	9.64	8.13	82.7	
N3226-03	EW4L4	3	0.95	8.98	9.93	9.09	90.6	
N3226-04	EW4L5	4	0.91	9.03	9.94	9.11	90.8	
N3226-05	EW4L6	5	0.91	8.75	9.66	8.04	81.5	
N3226-06	EW4L7	6	0.96	9.00	9.96	8.63	85.2	
N3226-07	EW4M4	7	0.91	9.07	9.98	9.55	95.3	
N3226-08	VHBLK001	8	1.00	1.00	2.00	2.00	100.0	VHBLK
N3226-09	EW4M9	9	0.91	9.06	9.97	7.79	75.9	
N3226-10	EW4N0	10	0.91	9.03	9.94	7.84	76.7	
N3226-11	EW4N1	11	0.96	9.02	9.98	8.58	84.5	
N3226-12	EW4N2	12	0.91	8.75	9.66	7.78	78.5	
N3226-13	EW4N3	13	0.94	9.01	9.95	8.46	83.5	
N3226-14	EW4N4	14	0.88	8.82	9.7	9.04	92.5	
N3226-15	EW4N5	15	0.89	8.75	9.64	7.55	76.1	
N3226-16	EW4N6	16	0.91	9.07	9.98	7.78	75.7	
N3226-17	EW4N7	17	0.93	9.01	9.94	8.61	85.2	
N3226-18	EW4N8	18	0.93	9.02	9.95	9.17	91.4	
N3226-19	EW4N9	19	0.88	9.08	9.96	8.06	79.1	
N3226-20	EW4P0	20	0.84	8.81	9.65	7.7	77.9	
N3226-21	EW4P0MS	21	0.84	8.81	9.65	7.7	77.9	
N3226-22	EW4P0MSD	22	0.84	8.81	9.65	7.7	77.9	
N3226-23	EW4P1	23	0.91	9.07	9.98	9.24	91.8	
N3226-24	EW4P1MS	24	0.91	9.07	9.98	9.24	91.8	
N3226-25	EW4P1MSD	25	0.91	9.07	9.98	9.24	91.8	
N3299-01	COM-4	26	1.00	1.00	2.00	2.00	100.0	concrete
N3299-07	DEC-7N	27	1.00	1.00	2.00	2.00	100.0	concrete
N3299-08	DEC-7E	28	1.00	1.00	2.00	2.00	100.0	concrete



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 6/10/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 16:25

In Date: 06/09/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:20

Out Date: 06/10/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB120598

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3299-09	SC-2-DEC-W	29	1.00	1.00	2.00	2.00	100.0	concrete
N3299-10	SC-2-DEC-N	30	1.00	1.00	2.00	2.00	100.0	concrete
N3299-11	SC-DEC-S	31	1.00	1.00	2.00	2.00	100.0	concrete

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$