

Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB120603

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	06/09/2022	17:15
2	CAL2	1	Water	NA	NA	20.9	7.00	06/09/2022	17:16
3	CAL3	1	Water	NA	NA	20.9	10.03	06/09/2022	17:17
4	ICV	1	Water	NA	NA	20.8	6.99	06/10/2022	17:20
5	CCV1	1	Water	NA	NA	20.8	2.01	06/09/2022	17:22
6	N3293-08	1	Water	NA	NA	25.4	7.89	06/09/2022	17:24
7	N3293-09	1	Water	NA	NA	25.6	7.39	06/09/2022	17:26
8	N3293-10	1	Water	NA	NA	25.2	5.60	06/09/2022	17:29
9	N3293-13	1	Water	NA	NA	25.3	5.34	06/09/2022	17:30
10	N3293-14	1	Water	NA	NA	25.2	6.28	06/09/2022	17:32
11	N3293-15	1	Water	NA	NA	25.3	4.87	06/09/2022	17:35
12	N3293-16	1	Water	NA	NA	24.8	5.91	06/09/2022	17:36
13	N3293-17	1	Water	NA	NA	24.2	5.89	06/09/2022	17:39
14	N3293-18	1	Water	NA	NA	24.7	8.74	06/09/2022	17:40
15	N3293-19	1	Water	NA	NA	25.1	5.81	06/09/2022	17:42
16	CCV2	1	Water	NA	NA	20.9	12.03	06/09/2022	17:43
17	N3293-19DUP	1	Water	NA	NA	25.2	5.82	06/09/2022	17:44
18	CCV3	1	Water	NA	NA	20.9	2.01	06/09/2022	17:45

WORKLIST(Hardcopy Internal Chain)

VP120603

WorkList Name : ph6922

WorkList ID : 160406

Department : Wet-Chemistry

Date : 06-09-2022 17:06:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
5/09/2022	Water	N3293-08	C pH	Cool 4 deg C	PSEG03	P11	52322-A	06/09/2022	9040C
5/09/2022	Water	N3293-09	C pH	Cool 4 deg C	PSEG03	P11	52322-B	06/09/2022	9040C
5/09/2022	Water	N3293-10	C pH	Cool 4 deg C	PSEG03	P11	52322-C	06/09/2022	9040C
5/09/2022	Water	N3293-13	C pH	Cool 4 deg C	PSEG03	P11	40805	06/09/2022	9040C
5/09/2022	Water	N3293-14	C pH	Cool 4 deg C	PSEG03	P11	40806	06/09/2022	9040C
5/09/2022	Water	N3293-15	C pH	Cool 4 deg C	PSEG03	P11	40807	06/09/2022	9040C
5/09/2022	Water	N3293-16	C pH	Cool 4 deg C	PSEG03	P11	40809	06/09/2022	9040C
5/09/2022	Water	N3293-17	C pH	Cool 4 deg C	PSEG03	P11	40810	06/09/2022	9040C
5/09/2022	Water	N3293-18	C pH	Cool 4 deg C	PSEG03	P11	52322-D	06/09/2022	9040C
5/09/2022	Water	N3293-19	C pH	Cool 4 deg C	PSEG03	P11	30986	06/09/2022	9040C