



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120607

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/10/2022	12:00
2	CAL2	1	Water	NA	NA	20.7	7.00	06/10/2022	12:01
3	CAL3	1	Water	NA	NA	20.7	10.03	06/10/2022	12:05
4	ICV	1	Water	NA	NA	20.6	6.98	06/10/2022	12:06
5	CCV1	1	Water	NA	NA	20.6	2.01	06/10/2022	12:09
6	N3293-21	1	Water	NA	NA	23.7	6.51	06/10/2022	12:10
7	N3309-02	1	Water	NA	NA	20.1	6.83	06/10/2022	12:15
8	N3309-02DUP	1	Water	NA	NA	20.2	6.84	06/10/2022	12:16
9	CCV2	1	Water	NA	NA	20.6	12.03	06/10/2022	12:19

WORKLIST(Hardcopy Internal Chain)

1320607

WorkList Name : PH N3309 WorkList ID : 160434 Department : Wet-Chemistry Date : 06-10-2022 11:41:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/08/2022	Water	N3309-02	pH <u>C</u>	Cool 4 deg C	TETRO6	D11	TT-046-IDWGW-20220608	06/08/2022	9040C

Date/Time 06.10.22 11:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 06.10.22 12:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]