


LB120622 Instrument ID : CV1

Reviewed By: bin
On: 6/13/2022 11:28:25 PM
Inst Id: CV1
LB: LB120622

7470A

μ Abs.:

54075



Linear

A= 0.0000e+000

B= 1.8400e-004

C= -3.8465e-002

Rho= 0.9998674

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p
0.0	0.000	-0.039	-0.039	-2	0.000	-2					-24
0.2	0.200	0.153	-0.047	1043	0.0 %	1043					2
2.5	2.500	2.561	0.061	14128	0.0 %	14128					1
5.0	5.000	5.050	0.050	27656	0.0 %	27656					1
7.5	7.500	7.562	0.062	41308	0.0 %	41308					1
10.0	10.000	9.912	-0.088	54075	0.0 %	54075					-1

LB120622 Instrument ID: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jun 2022 16:44:38

Sample ID	Extended ID	μ Abs	Conc	Std Cond Method	Units	Date	Type	Type
0.0 - 1	50.0	-2	-	0.00007470A	PPB	10 Jun 2022 17:05:49	S	Std
0.2 - 1	51.2	1043	-	0.20007470A	PPB	10 Jun 2022 17:08:06	S	Std
2.5 - 1	52.5	14128	-	2.50007470A	PPB	10 Jun 2022 17:10:23	S	Std
5.0 - 1	55.0	27656	-	5.00007470A	PPB	10 Jun 2022 17:12:40	S	Std
7.5 - 1	57.5	41308	-	7.50007470A	PPB	10 Jun 2022 17:14:56	S	Std
10.0 - 1	59.0	54075	-	10.00007470A	PPB	10 Jun 2022 17:17:12	S	Std
ICV66 - 1	ICV66	21602	3.9364	-7470A	PPB	10 Jun 2022 17:30:15	U	SMPL
ICB66 - 1	ICB66	29	-0.0331	-7470A	PPB	10 Jun 2022 17:32:29	U	SMPL
CCV56 - 1	CCV56	26554	4.8476	-7470A	PPB	10 Jun 2022 17:34:46	U	SMPL
CCB56 - 1	CCB56	14	-0.0359	-7470A	PPB	10 Jun 2022 17:37:01	U	SMPL
CRA - 1	CRA	1148	0.1728	-7470A	PPB	10 Jun 2022 17:39:18	U	SMPL
HighStd - 1	HighStd	53385	9.7846	-7470A	PPB	10 Jun 2022 17:41:33	U	SMPL
ChkStd - 1	ChkStd	38654	7.0740	-7470A	PPB	10 Jun 2022 17:47:17	U	SMPL
PB145440BL - 1	PBW	-156	-0.0672	-7470A	PPB	10 Jun 2022 17:49:32	U	SMPL
PB145440BS - 1	LCSW	24578	4.4840	-7470A	PPB	10 Jun 2022 17:51:50	U	SMPL
N3236-01 - 1	21ST-HOYT-AVE-M-15834	-252	-0.0848	-7470A	PPB	10 Jun 2022 17:54:06	U	SMPL
N3239-01 - 1	MLD-MW-4-GW-062022	82	-0.0234	-7470A	PPB	10 Jun 2022 17:56:30	U	SMPL
N3239-03 - 1	MLD-MW-6-GW-062022	84	-0.0230	-7470A	PPB	10 Jun 2022 17:58:46	U	SMPL
N3239-05 - 1	MLD-MW-5-GW-062022	76	-0.0245	-7470A	PPB	10 Jun 2022 18:01:03	U	SMPL
N3239-07 - 1	MLD-MW-3-GW-062022	126	-0.0153	-7470A	PPB	10 Jun 2022 18:03:19	U	SMPL
CCV57 - 1	CCV57	27725	5.0630	-7470A	PPB	10 Jun 2022 18:05:35	U	SMPL
CCB57 - 1	CCB57	-10	-0.0403	-7470A	PPB	10 Jun 2022 18:07:50	U	SMPL
N3239-07DUP - 1	MLD-MW-3-GW-062022DUP	111	-0.0180	-7470A	PPB	10 Jun 2022 18:10:06	U	SMPL
N3239-09 - 1	MLD-EB-1-DW-062022	129	-0.0147	-7470A	PPB	10 Jun 2022 18:12:22	U	SMPL
N3239-11 - 1	MLD-MW-2-GW-062022	87	-0.0225	-7470A	PPB	10 Jun 2022 18:14:37	U	SMPL
N3239-13 - 1	MLD-DUPLICATE-1-GW-062022	87	-0.0225	-7470A	PPB	10 Jun 2022 18:16:52	U	SMPL
N3239-15 - 1	MLD-MW-1-GW-062022	89	-0.0221	-7470A	PPB	10 Jun 2022 18:19:08	U	SMPL
N3239-19 - 1	N3239-07MS	23070	4.2065	-7470A	PPB	10 Jun 2022 18:21:24	U	SMPL
N3239-20 - 1	N3239-07MSD	22937	4.1820	-7470A	PPB	10 Jun 2022 18:23:40	U	SMPL
N3266-01 - 1	OH-WATER-0608	-32	-0.0444	-7470A	PPB	10 Jun 2022 18:25:58	U	SMPL
N3280-01 - 1	21ST-CASING-4	175	-0.0063	-7470A	PPB	10 Jun 2022 18:28:17	U	SMPL
N3280-02 - 1	21ST-CASING-3	119	-0.0166	-7470A	PPB	10 Jun 2022 18:30:33	U	SMPL
CCV58 - 1	CCV58	28292	5.1674	-7470A	PPB	10 Jun 2022 18:32:48	U	SMPL
CCB58 - 1	CCB58	3	-0.0379	-7470A	PPB	10 Jun 2022 18:35:04	U	SMPL
N3280-03 - 1	33RD-AVE-CASING-10	627	0.0769	-7470A	PPB	10 Jun 2022 18:37:21	U	SMPL
N3239-07LX5 - 1	MLD-MW-3-GW-062022	89	-0.0221	-7470A	PPB	10 Jun 2022 18:39:37	U	SMPL
N3239-07A - 1	MLD-MW-3-GW-062022	23053	4.2034	-7470A	PPB	10 Jun 2022 18:41:52	U	SMPL
CCV59 - 1	CCV59	28170	5.1449	-7470A	PPB	10 Jun 2022 18:44:08	U	SMPL
CCB59 - 1	CCB59	-21	-0.0423	-7470A	PPB	10 Jun 2022 18:46:25	U	SMPL