


LB120647 Instrument ID; CV1

7471B

μ Abs.:

54728



Linear

A= 0.0000e+000

B= 1.8194e-004

C= 1.5980e-002

Rho= 0.9999222

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/0/0
0.0	0.000	0.011	0.011	-28	0.000	-28					-9
0.2	0.200	0.182	-0.018	911	0.0 %	911					1
2.5	2.500	2.536	0.036	13850	0.0 %	13850					1
5.0	5.000	4.927	-0.073	26994	0.0 %	26994					1
7.5	7.500	7.571	0.071	41522	0.0 %	41522					1
10.0	10.000	9.973	-0.027	54728	0.0 %	54728					0

LB120647 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jun 2022 16:49:49

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	-28	-	0.0000	7471B	PPB	13 Jun 2022 17:45:56	S	Std
0.2 - 1	50.2	911	-	0.2000	7471B	PPB	13 Jun 2022 17:48:13	S	Std
2.5 - 1	52.5	13850	-	2.5000	7471B	PPB	13 Jun 2022 17:50:31	S	Std
5.0 - 1	55.0	26994	-	5.0000	7471B	PPB	13 Jun 2022 17:52:47	S	Std
7.5 - 1	57.5	41522	-	7.5000	7471B	PPB	13 Jun 2022 17:55:04	S	Std
10.0 - 1	510.0	54728	-	10.0000	7471B	PPB	13 Jun 2022 17:57:21	S	Std
ICV68 - 1	ICV68	21121	3.8588	-	7471B	PPB	13 Jun 2022 18:03:39	U	SMPL
ICB68 - 1	ICB68	-49	0.0071	-	7471B	PPB	13 Jun 2022 18:05:54	U	SMPL
CCV64 - 1	CCV64	25469	4.6499	-	7471B	PPB	13 Jun 2022 18:08:10	U	SMPL
CCB64 - 1	CCB64	-102	-0.0026	-	7471B	PPB	13 Jun 2022 18:10:25	U	SMPL
CRA - 1	CRA	910	0.1815	-	7471B	PPB	13 Jun 2022 18:12:40	U	SMPL
HighStd - 1	HighStd	54408	9.9152	-	7471B	PPB	13 Jun 2022 18:14:55	U	SMPL
ChkStd - 1	ChkStd	36870	6.7243	-	7471B	PPB	13 Jun 2022 18:17:11	U	SMPL
PB145495BL - 1	PBS	-153	-0.0119	-	7471B	PPB	13 Jun 2022 18:19:29	U	SMPL
PB145495BS - 1	LCSS	21503	3.9283	-	7471B	PPB	13 Jun 2022 18:21:47	U	SMPL
N3233-01 - 1	CR-1	432	0.0946	-	7471B	PPB	13 Jun 2022 18:24:03	U	SMPL
N3233-02 - 1	CR-2	3141	0.5875	-	7471B	PPB	13 Jun 2022 18:26:20	U	SMPL
N3234-02 - 1	RB-21499	503	0.1075	-	7471B	PPB	13 Jun 2022 18:28:37	U	SMPL
N3251-01 - 1	NB-8-060822	519	0.1104	-	7471B	PPB	13 Jun 2022 18:30:54	U	SMPL
N3251-01DUP - 1	NB-8-060822DUP	741	0.1508	-	7471B	PPB	13 Jun 2022 18:33:09	U	SMPL
CCV65 - 1	CCV65	26285	4.7984	-	7471B	PPB	13 Jun 2022 18:35:25	U	SMPL
CCB65 - 1	CCB65	-118	-0.0055	-	7471B	PPB	13 Jun 2022 18:37:40	U	SMPL
N3251-01MS - 1	NB-8-060822MS	21446	3.9180	-	7471B	PPB	13 Jun 2022 18:43:53	U	SMPL
N3251-01MSD - 1	NB-8-060822MSD	21974	4.0140	-	7471B	PPB	13 Jun 2022 18:46:08	U	SMPL
N3267-01 - 1	OR-2-060822	2099	0.3979	-	7471B	PPB	13 Jun 2022 18:48:24	U	SMPL
N3281-01 - 1	129	14709	2.6922	-	7471B	PPB	13 Jun 2022 18:50:41	U	SMPL
N3282-01 - 1	125	1067	0.2101	-	7471B	PPB	13 Jun 2022 18:52:57	U	SMPL
N3295-01 - 1	OR-03-060922	1252	0.2438	-	7471B	PPB	13 Jun 2022 18:55:14	U	SMPL
N3296-01 - 1	OK-03-060922	882	0.1765	-	7471B	PPB	13 Jun 2022 18:57:30	U	SMPL
N3301-03 - 1	WC-SO-20220609	1681	0.3218	-	7471B	PPB	13 Jun 2022 18:59:46	U	SMPL
N3304-01 - 1	WC-14A-1A-10	40118	7.3152	-	7471B	PPB	13 Jun 2022 19:02:04	U	SMPL
N3304-02 - 1	WC-14A-1A-10-G	217221	39.5381	-	7471B	PPB	13 Jun 2022 19:04:19	U	SMPL
CCV66 - 1	CCV66	25680	4.6883	-	7471B	PPB	13 Jun 2022 19:06:37	U	SMPL
CCB66 - 1	CCB66	-196	-0.0197	-	7471B	PPB	13 Jun 2022 19:08:57	U	SMPL
N3304-03 - 1	WC-14A-1A-9	36867	6.7237	-	7471B	PPB	13 Jun 2022 19:11:14	U	SMPL
N3304-04 - 1	WC-14A-1A-9-G	7986	1.4690	-	7471B	PPB	13 Jun 2022 19:13:29	U	SMPL
N3304-05 - 1	WC-14A-1A-6	15979	2.9233	-	7471B	PPB	13 Jun 2022 19:15:48	U	SMPL
N3304-06 - 1	WC-14A-1A-6-G	50604	9.2231	-	7471B	PPB	13 Jun 2022 19:18:03	U	SMPL
N3319-01 - 1	TR-06-061022	1133	0.2221	-	7471B	PPB	13 Jun 2022 19:20:19	U	SMPL
N3327-01 - 1	EW-WMR-COMP-1	1476	0.2845	-	7471B	PPB	13 Jun 2022 19:22:37	U	SMPL
N3327-03 - 1	EW-WMR-COMP-2	777	0.1574	-	7471B	PPB	13 Jun 2022 19:24:54	U	SMPL
N3327-05 - 1	EW-WMR-COMP-3	7400	1.3624	-	7471B	PPB	13 Jun 2022 19:27:11	U	SMPL
CCV67 - 1	CCV67	26142	4.7724	-	7471B	PPB	13 Jun 2022 19:29:28	U	SMPL
CCB67 - 1	CCB67	-109	-0.0039	-	7471B	PPB	13 Jun 2022 19:31:43	U	SMPL
N3251-01LX5 - 1		49	0.0249	-	7471B	PPB	13 Jun 2022 19:34:00	U	SMPL
N3251-01A - 1		21846	3.9907	-	7471B	PPB	13 Jun 2022 19:36:16	U	SMPL
N3204-02DLX10 - 1		21431	3.9152	-	7471B	PPB	13 Jun 2022 19:38:31	U	SMPL
CCV068 - 1	CCV068	26860	4.9030	-	7471B	PPB	13 Jun 2022 19:40:47	U	SMPL
CCB068 - 1	CCB068	-105	-0.0031	-	7471B	PPB	13 Jun 2022 19:43:04	U	SMPL