



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB120660

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	06/14/2022	12:40
2	CAL2	1	Water	NA	NA	20.9	7.00	06/14/2022	12:41
3	CAL3	1	Water	NA	NA	20.9	10.03	06/14/2022	12:42
4	ICV	1	Water	NA	NA	20.9	6.98	06/14/2022	12:45
5	CCV1	1	Water	NA	NA	20.8	2.01	06/14/2022	12:46
6	N3339-03	1	Water	NA	NA	20.1	12.55	06/14/2022	12:50
7	N3339-03DUP	1	Water	NA	NA	20.2	12.56	06/14/2022	12:51
8	CCV2	1	Water	NA	NA	20.6	12.03	06/14/2022	13:00

WORKLIST(Hardcopy Internal Chain)

VB 120660

WorkList Name : corrosivity n3339 WorkList ID : 160533 Department : Wet-Chemistry Date : 06-14-2022 12:31:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/13/2022	Water	N3339-03	Corrosivity	Cool 4 deg C	EART04	D11	WC-GW-20220613	06/13/2022	9040C

Date/Time 06/14/22 12:35
 Raw Sample Received by: JD/60C
 Raw Sample Relinquished by: CR SM

Date/Time 06-14-22 12:00
 Raw Sample Received by: CR SM
 Raw Sample Relinquished by: JD/60C