



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB120682

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/15/2022	10:05
2	CAL2	1	Water	NA	NA	20.7	7.00	06/15/2022	10:06
3	CAL3	1	Water	NA	NA	20.7	10.03	06/15/2022	10:07
4	ICV	1	Water	NA	NA	20.6	6.99	06/15/2022	10:10
5	CCV1	1	Water	NA	NA	20.7	2.01	06/15/2022	10:15
6	N3314-01	1	Water	NA	NA	24.7	9.52	06/15/2022	10:19
7	N3314-01DUP	1	Water	NA	NA	24.8	9.53	06/15/2022	10:20
8	N3321-01	1	Water	NA	NA	21.7	5.05	06/15/2022	10:25
9	CCV2	1	Water	NA	NA	20.8	12.02	06/15/2022	10:26

WORKLIST(Hardcopy Internal Chain)

13 120682

WorkList Name : ph n3314 WorkList ID : 160562 Department : Wet-Chemistry Date : 06-15-2022 09:53:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/10/2022	Water	N3314-01	pH	Cool 4 deg C	PSEG03	D11	ER-HEATER-COMP	06/10/2022	9040C
06/01/2022	Water	N3321-01	pH	Cool 4 deg C	STER15	D11	POOL-WATER	06/01/2022	9040C

Date/Time 06-15-22 10:00
 Raw Sample Received by: CP sm
 Raw Sample Relinquished by: CP sm

Date/Time 06-15-22 11:00
 Raw Sample Received by: CP sm
 Raw Sample Relinquished by: 78(WC)