


LB120686 INSTRUMENT ID: CV1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	-34	0.000	-34				
0.2	0.200	0.170	-0.030	850	0.0 %	850				
2.5	2.500	2.376	-0.124	13069	0.0 %	13069				
5.0	5.000	5.278	0.278	29139	0.0 %	29139				
7.5	7.500	7.432	-0.068	41066	0.0 %	41066				
10.0	10.000	9.933	-0.067	54913	0.0 %	54913				

LB120686 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Jun 2022 12:55:36

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-34	-	0.0000	7471B	PPB	15 Jun 2022 14:13:19	S	Std
0.2 - 1	50.2	850	-	0.2000	7471B	PPB	15 Jun 2022 14:15:36	S	Std
2.5 - 1	52.5	13069	-	2.5000	7471B	PPB	15 Jun 2022 14:17:53	S	Std
5.0 - 1	53.0	29139	-	5.0000	7471B	PPB	15 Jun 2022 14:20:09	S	Std
7.5 - 1	53.5	41066	-	7.5000	7471B	PPB	15 Jun 2022 14:22:26	S	Std
10.0 - 1	54.0	54913	-	10.0000	7471B	PPB	15 Jun 2022 14:24:44	S	Std
ICV76 - 1	ICV76	22069	4.0017	-	7471B	PPB	15 Jun 2022 14:33:02	U	SMPL
ICB76 - 1	ICB76	-70	0.0037	-	7471B	PPB	15 Jun 2022 14:35:17	U	SMPL
CCV01 - 1	CCV01	29037	5.2600	-	7471B	PPB	15 Jun 2022 14:37:33	U	SMPL
CCB01 - 1	CCB01	-87	0.0007	-	7471B	PPB	15 Jun 2022 14:39:48	U	SMPL
CRA - 1	CRA	848	0.1695	-	7471B	PPB	15 Jun 2022 14:42:05	U	SMPL
HighStd - 1	HighStd	54904	9.9312	-	7471B	PPB	15 Jun 2022 14:44:20	U	SMPL
ChkStd - 1	ChkStd	35603	6.4457	-	7471B	PPB	15 Jun 2022 14:46:35	U	SMPL
PB145496BL - 1	PBS	-130	-0.0071	-	7471B	PPB	15 Jun 2022 14:48:53	U	SMPL
PB145496BS - 1	LCSS	23439	4.2491	-	7471B	PPB	15 Jun 2022 14:51:10	U	SMPL
N3293-05 - 1	30873-76	3373	0.6255	-	7471B	PPB	15 Jun 2022 14:53:27	U	SMPL
N3293-06 - 1	40804	201	0.0527	-	7471B	PPB	15 Jun 2022 14:55:44	U	SMPL
N3293-07 - 1	40808	15481	2.8120	-	7471B	PPB	15 Jun 2022 14:58:01	U	SMPL
N3299-01 - 1	COM-4	209	0.0541	-	7471B	PPB	15 Jun 2022 15:00:17	U	SMPL
N3299-03 - 1	SC-15	377	0.0844	-	7471B	PPB	15 Jun 2022 15:02:34	U	SMPL
CCV02 - 1	CCV02	28965	5.2470	-	7471B	PPB	15 Jun 2022 15:04:50	U	SMPL
CCB02 - 1	CCB02	-89	0.0003	-	7471B	PPB	15 Jun 2022 15:07:05	U	SMPL
N3299-04 - 1	SC-16	204	0.0532	-	7471B	PPB	15 Jun 2022 15:09:22	U	SMPL
N3299-05 - 1	SC-17	277	0.0664	-	7471B	PPB	15 Jun 2022 15:11:37	U	SMPL
N3299-06 - 1	SC-18	43	0.0241	-	7471B	PPB	15 Jun 2022 15:13:53	U	SMPL
N3302-01 - 1	I-1	1299	0.2509	-	7471B	PPB	15 Jun 2022 15:16:08	U	SMPL
N3302-02 - 1	I-2	1344	0.2591	-	7471B	PPB	15 Jun 2022 15:18:24	U	SMPL
N3324-01 - 1	BT-SPOILS-07	3009	0.5597	-	7471B	PPB	15 Jun 2022 15:20:41	U	SMPL
N3328-01 - 1	BU-02-061022	1132	0.2208	-	7471B	PPB	15 Jun 2022 15:22:57	U	SMPL
N3329-01 - 1	PALMYRA-RT-73-COMP-1	46	0.0247	-	7471B	PPB	15 Jun 2022 15:25:14	U	SMPL
N3329-03 - 1	PALMYRA-RT-73-COMP-2	146	0.0427	-	7471B	PPB	15 Jun 2022 15:27:31	U	SMPL
N3329-05 - 1	PALMYRA-RT-73-COMP-3	402	0.0890	-	7471B	PPB	15 Jun 2022 15:29:47	U	SMPL
CCV03 - 1	CCV03	29003	5.2539	-	7471B	PPB	15 Jun 2022 15:32:02	U	SMPL
CCB03 - 1	CCB03	-75	0.0028	-	7471B	PPB	15 Jun 2022 15:34:18	U	SMPL
N3334-01 - 1	WC-14A-1A-7	1471	0.2820	-	7471B	PPB	15 Jun 2022 15:36:35	U	SMPL
N3334-07 - 1	WC-14A-1A-8	10936	1.9912	-	7471B	PPB	15 Jun 2022 15:38:50	U	SMPL
N3334-13 - 1	WC-14A-1A-11	21603	3.9175	-	7471B	PPB	15 Jun 2022 15:41:06	U	SMPL
N3334-19 - 1	WC-14A-1A-5	63213	11.4317	-	7471B	PPB	15 Jun 2022 15:43:22	U	SMPL
N3334-19DUP - 1	WC-14A-1A-5DUP	62026	11.2173	-	7471B	PPB	15 Jun 2022 15:45:39	U	SMPL
N3334-19MS - 1	WC-14A-1A-5MS	73652	13.3168	-	7471B	PPB	15 Jun 2022 15:47:58	U	SMPL
N3334-19MSD - 1	WC-14A-1A-5MSD	72614	13.1293	-	7471B	PPB	15 Jun 2022 15:53:51	U	SMPL
CCV04 - 1	CCV04	29518	5.3469	-	7471B	PPB	15 Jun 2022 15:57:42	U	SMPL
CCB04 - 1	CCB04	-179	-0.0160	-	7471B	PPB	15 Jun 2022 15:59:58	U	SMPL
N3334-19LX5 - 1		14157	2.5729	-	7471B	PPB	15 Jun 2022 16:02:16	U	SMPL
N3334-19A - 1		102907	18.5998	-	7471B	PPB	15 Jun 2022 16:15:09	U	SMPL
N3334-19DLX2 - 1		21413	3.8832	-	7471B	PPB	15 Jun 2022 16:23:17	U	SMPL
N3334-19DUPDLX5 - 1		20167	3.6582	-	7471B	PPB	15 Jun 2022 16:34:56	U	SMPL
N3334-19LDLX25 - 1		4310	0.7947	-	7471B	PPB	15 Jun 2022 16:37:11	U	SMPL
N3334-19MSDLX5 - 1		24136	4.3750	-	7471B	PPB	15 Jun 2022 16:40:45	U	SMPL
N3334-19MSDDLX5 - 1		24166	4.3804	-	7471B	PPB	15 Jun 2022 16:43:01	U	SMPL
N3334-19ADLX5 - 1		24501	4.4409	-	7471B	PPB	15 Jun 2022 16:45:18	U	SMPL
CCV05 - 1	CCV05	28197	5.1083	-	7471B	PPB	15 Jun 2022 16:48:59	U	SMPL
CCB05 - 1	CCB05	-107	-0.0030	-	7471B	PPB	15 Jun 2022 16:51:15	U	SMPL