



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB120687

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/15/2022	10:20
2	N3314-01		1	15	100.00	>212.0	>212.0	06/15/2022	10:45
3	N3339-03		1	12	100.00	>212.0	>212.0	06/15/2022	11:10
4	N3339-03DUP		1	12	100.00	>212.0	>212.0	06/15/2022	11:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 120687

WorkList Name : FP61422 WorkList ID : 160549 Department : Wet-Chemistry Date : 06-14-2022 17:46:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/15/2022	Water	N3314-01	Flash Point	Cool 4 deg C	PSEG03	D11	ER-HEATER-COMP	06/10/2022	1010B
06/27/2022	Water	N3339-03	Flash Point	Cool 4 deg C	EART04	D11	WC-GW-20220613	06/13/2022	1010B

Date/Time 06/15/22 09:55
 Raw Sample Received by: KS(WC)
 Raw Sample Relinquished by: _____

Date/Time 06/15/22 11:40
 Raw Sample Received by: _____
 Raw Sample Relinquished by: KS(WC)