

L13 12068P

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : i2 Instrument ID : Konelab

6/15/2022 13:24

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.828	0.0	0.074	
ICB1	1.149	0.0	0.001	
CCV1	240.757	0.0	0.189	
CCB1	0.914	0.0	0.001	
PB145569BL	0.419	0.0	0.001	
N3314-01	0.720	0.0	0.001	
N3339-03	0.861	0.0	0.001	
N3339-03DUP	0.867	0.0	0.001	
CCV2	241.221	0.0	0.189	
CCB2	1.002	0.0	0.001	

N	10
Mean	58.174
SD	100.6195
CV%	172.96

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 15 09:37:25 2022

Wed Jun 15 13:18:17 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.3722	µg/l	6/15/2022 9:37:25	
5.OPPBCN	A	Total CN	P	5.114	µg/l	6/15/2022 9:37:26	
10PPBCN	A	Total CN	P	10.0267	µg/l	6/15/2022 9:37:27	
50PPBCN	A	Total CN	P	48.7652	µg/l	6/15/2022 9:37:28	
100PPBCN	A	Total CN	P	99.0392	µg/l	6/15/2022 9:37:29	
250PPBCN	A	Total CN	P	252.7376	µg/l	6/15/2022 9:37:30	
500PPBCN	A	Total CN	P	498.9452	µg/l	6/15/2022 9:37:31	
ICV1	S	Total CN	P	93.8277	µg/l	6/15/2022 13:11:31	
ICB1	S	Total CN	P	1.1489	µg/l	6/15/2022 13:11:32	
CCV1	S	Total CN	P	240.7566	µg/l	6/15/2022 13:11:33	
CCB1	S	Total CN	P	0.9136	µg/l	6/15/2022 13:11:34	
PB145569BL	S	Total CN	P	0.4188	µg/l	6/15/2022 13:11:35	
N3314-01	S	Total CN	P	0.72	µg/l	6/15/2022 13:11:36	
N3339-03	S	Total CN	P	0.8605	µg/l	6/15/2022 13:11:37	
N3339-03DUP	S	Total CN	P	0.867	µg/l	6/15/2022 13:11:38	
CCV2	S	Total CN	P	241.2207	µg/l	6/15/2022 13:11:39	
CCB2	S	Total CN	P	1.0022	µg/l	6/15/2022 13:11:40	

CHEMTECH CONSULTING GROUP INC

284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : GM

Instrument ID : Konelab

6/15/2022 9:45


Test Total CN

Accepted 6/15/2022 9:45

Factor 1277
Bias 0

Coeff. of det. 0.999945

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.3722	0.0000	
2	5.0PPBCN	0.004	5.1140	5.0000	2.3
3	10PPBCN	0.008	10.0267	10.0000	6.3
4	50PPBCN	0.039	48.7652	50.0000	-2.5
5	100PPBCN	0.078	99.0392	100.0000	-1.0
6	250PPBCN	0.198	252.7376	250.0000	1.1
7	500PPBCN	0.391	498.9452	500.0000	-0.2

GM
6/15/22