

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: ASTM COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB120695

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP96871
COD calibration std. 100 ppm	WP96868
COD calibration std. 10 ppm	WP96866
COD calibration std. 150 ppm	WP96869
COD calibration std. 50 ppm	WP96867
COD calibration std. 0 ppm	WP96865
COD ICV-LCS std, 50ppm	WP97294
COD CCV std, 50ppm	WP97293
COD Digestion Vials Low Level 0-150Mg/L	W2832

Temp In (C): 148	Date In: 06/15/2022	Time In: 09:50
Temp Out (C): 150	Date Out: 06/15/2022	Time Out: 11:50

Intercept: 0.5982

Slope: 0.9936

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.602		05/19/2022	13:35
2	CAL2	10	1	Water	11.000	10.469	-4.7	05/19/2022	13:35
3	CAL3	50	1	Water	50.000	49.72	0.6	05/19/2022	13:36
4	CAL4	100	1	Water	101.000	101.049	-1	05/19/2022	13:36
5	CAL5	150	1	Water	149.000	149.358	0.4	05/19/2022	13:37



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: ASTM COD

SUPERVISOR REVIEW BY: jignesh

Run Number: LB120695

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	51.000	50.726	05/19/2022	13:37
2	ICB		NA	NA	1	Water	0.000	-0.602	05/19/2022	13:38
3	CCV1	50	NA	NA	1	Water	51.000	50.726	06/15/2022	15:20
4	CCB1		NA	NA	1	Water	0.000	-0.602	06/15/2022	15:20
5	LB120695BL		NA	NA	1	Water	1.000	0.404	06/15/2022	15:21
6	LB120695BS	50	NA	NA	1	Water	49.000	48.714	06/15/2022	15:21
7	N3334-04		NA	NA	1	Water	95.000	95.010	06/15/2022	15:22
8	N3334-10		NA	NA	1	Water	51.000	50.726	06/15/2022	15:22
9	N3334-16		NA	NA	10	Water	18.000	17.514	06/15/2022	15:23
10	N3334-22		NA	NA	1	Water	57.000	56.765	06/15/2022	15:23
11	N3334-22DUP		NA	NA	1	Water	57.000	56.765	06/15/2022	15:24
12	N3334-22MS	50	NA	NA	1	Water	104.000	104.068	06/15/2022	15:24
13	N3334-22MSD	50	NA	NA	1	Water	105.000	105.074	06/15/2022	15:25
14	CCV2	50	NA	NA	1	Water	50.000	49.720	06/15/2022	15:25
15	CCB2		NA	NA	1	Water	1.000	0.404	06/15/2022	15:26