



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: mohammad

Run Number: LB120699

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP97418
Residual chlorine ICV-LCS, 0.4PPM	WP97417
Chlorine Calibration std, 0.1ppm	WP97412
Chlorine Calibration std, 0.2ppm	WP97413
Chlorine Calibration std, 0.8ppm	WP97415
Chlorine Calibration std, 0.0ppm	WP97411
Chlorine Calibration std, 1.6ppm	WP97416
Residual Chlorine Calibration and CCV std, 0	WP97414
Total Chlorine Powder Pillows	W2759

Intercept: 0.0049

Slope: 1.0164

Regression: 0.999988

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		06/15/2022	13:20
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	-3	06/15/2022	13:23
3	CAL3	0.2	1	0.000	0.210	0.210	0.20	-1	06/15/2022	13:26
4	CAL4	0.4	1	0.000	0.410	0.410	0.40	0.3	06/15/2022	13:29
5	CAL5	0.8	1	0.000	0.820	0.820	0.80	-0.3	06/15/2022	13:32
6	CAL6	1.6	1	0.000	1.630	1.630	1.60	0.1	06/15/2022	13:36

Analysis Method: SM4500 Cl GANALYST: IwonaParameter: Residual ChlorineSUPERVISOR REVIEW BY: mohammadRun Number: LB120699

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4000	0.4000	0.3890	06/15/2022	13:39
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	06/15/2022	13:42
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.3970	06/15/2022	13:45
4	CCB1				1	0.0000	0.0100	0.0100	0.0020	06/15/2022	13:48
5	LB120699BL	50	50		1	0.0000	0.0100	0.0100	0.0020	06/15/2022	13:51
6	LB120699BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.4070	06/15/2022	13:54
7	N3214-09	50	50		1	0.0000	0.0500	0.0500	0.0440	06/15/2022	13:57
8	CCV2			0.4	1	0.0000	0.4100	0.4100	0.3970	06/15/2022	14:01
9	CCB2				1	0.0000	0.0000	0.0000	-0.0080	06/15/2022	14:04