

LB120439


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

43137



A= 0.0000e+000

B= 2.3290e-004

C= -9.6976e-004

Rho= 0.9994118

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.0	0.000	0.054	0.054	236	0.000	236					
0.2	0.200	0.213	0.013	919	0.0 %	919					
2.5	2.500	2.341	-0.159	10054	0.0 %	10054					
5.0	5.000	5.203	0.203	22343	0.0 %	22343					
7.5	7.500	7.344	-0.156	31536	0.0 %	31536					
10.0	10.000	10.046	0.046	43137	0.0 %	43137					

LB120439 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Jun 2022 10:59:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	236	-	0.0000	7470A	PPB	01 Jun 2022 11:14:28	S	Std
0.2 - 1	502	919	-	0.2000	7470A	PPB	01 Jun 2022 11:16:45	S	Std
2.5 - 1	52.5	10054	-	2.5000	7470A	PPB	01 Jun 2022 11:19:02	S	Std
5.0 - 1	55	22343	-	5.0000	7470A	PPB	01 Jun 2022 11:21:18	S	Std
7.5 - 1	57.5	31536	-	7.5000	7470A	PPB	01 Jun 2022 11:23:35	S	Std
10.0 - 1	51	43137	-	10.0000	7470A	PPB	01 Jun 2022 11:25:51	S	Std
ICV55 - 1	ICV55	17633	4.1058	-7470A	PPB	01 Jun 2022 11:28:34	U	SMPL	
ICB55 - 1	ICB55	128	0.0288	-7470A	PPB	01 Jun 2022 11:30:48	U	SMPL	
CCV07 - 1	CCV07	22079	5.1413	-7470A	PPB	01 Jun 2022 11:33:03	U	SMPL	
CCB07 - 1	CCB07	132	0.0298	-7470A	PPB	01 Jun 2022 11:35:18	U	SMPL	
CRA - 1	CRA	849	0.1968	-7470A	PPB	01 Jun 2022 11:37:34	U	SMPL	
HighStd - 1	HighStd	42067	9.7965	-7470A	PPB	01 Jun 2022 11:39:49	U	SMPL	
ChkStd - 1	ChkStd	30341	7.0655	-7470A	PPB	01 Jun 2022 11:45:21	U	SMPL	
PB145240BL - 1	PBW	-166	-0.0396	-7470A	PPB	01 Jun 2022 11:47:37	U	SMPL	
PB145240BS - 1	LCSW	14954	3.4818	-7470A	PPB	01 Jun 2022 11:52:27	U	SMPL	
N3112-24 - 1	FB02-20220527	-53	-0.0133	-7470A	PPB	01 Jun 2022 11:54:43	U	SMPL	
N3114-01 - 1	220519081-01	47	0.0100	-7470A	PPB	01 Jun 2022 11:57:00	U	SMPL	
N3114-01DUP - 1	220519081-01DUP	35	0.0072	-7470A	PPB	01 Jun 2022 11:59:17	U	SMPL	
N3114-01MS - 1	220519081-01MS	14617	3.4034	-7470A	PPB	01 Jun 2022 12:01:33	U	SMPL	
N3114-01MSD - 1	220519081-01MSD	15031	3.4998	-7470A	PPB	01 Jun 2022 12:03:49	U	SMPL	
CCV08 - 1	CCV08	22048	5.1341	-7470A	PPB	01 Jun 2022 12:06:04	U	SMPL	
CCB08 - 1	CCB08	127	0.0286	-7470A	PPB	01 Jun 2022 12:08:19	U	SMPL	
N3115-01 - 1	220518043-03	413	0.0952	-7470A	PPB	01 Jun 2022 12:10:35	U	SMPL	
N3116-01 - 1	220518044-01	910	0.2110	-7470A	PPB	01 Jun 2022 12:12:50	U	SMPL	
N3117-01 - 1	220519080-01	114	0.0256	-7470A	PPB	01 Jun 2022 12:15:06	U	SMPL	
N3118-01 - 1	220509078-01	235	0.0538	-7470A	PPB	01 Jun 2022 12:17:22	U	SMPL	
N3121-01 - 1	220518061-01	70	0.0153	-7470A	PPB	01 Jun 2022 12:19:38	U	SMPL	
N3124-02 - 1	220518045-03	8503	1.9794	-7470A	PPB	01 Jun 2022 12:21:54	U	SMPL	
N3124-08 - 1	220518045-02	4358	1.0140	-7470A	PPB	01 Jun 2022 12:24:11	U	SMPL	
N3126-01 - 1	220510019-02	886	0.2054	-7470A	PPB	01 Jun 2022 12:26:29	U	SMPL	
N3127-01 - 1	220511094-02	46	0.0097	-7470A	PPB	01 Jun 2022 12:28:46	U	SMPL	
N3128-01 - 1	220511011-02	48	0.0102	-7470A	PPB	01 Jun 2022 12:31:01	U	SMPL	
CCV09 - 1	CCV09	21832	5.0837	-7470A	PPB	01 Jun 2022 12:33:17	U	SMPL	
CCB09 - 1	CCB09	120	0.0270	-7470A	PPB	01 Jun 2022 12:35:32	U	SMPL	
N3129-01 - 1	220523012-01	52	0.0111	-7470A	PPB	01 Jun 2022 12:37:48	U	SMPL	
N3130-01 - 1	220518071-02	66	0.0144	-7470A	PPB	01 Jun 2022 12:40:04	U	SMPL	
N3131-01 - 1	220510015-02	356	0.0819	-7470A	PPB	01 Jun 2022 12:42:20	U	SMPL	
N3132-01 - 1	220509007-02	77	0.0170	-7470A	PPB	01 Jun 2022 12:44:36	U	SMPL	
N3133-01 - 1	220510019-05	80	0.0177	-7470A	PPB	01 Jun 2022 12:46:52	U	SMPL	
N3134-01 - 1	220509011-03	305	0.0701	-7470A	PPB	01 Jun 2022 12:49:08	U	SMPL	
N3135-01 - 1	220510079-02	68	0.0149	-7470A	PPB	01 Jun 2022 12:51:25	U	SMPL	
N3136-01 - 1	220510063-01	83	0.0184	-7470A	PPB	01 Jun 2022 12:53:42	U	SMPL	
N3114-01LX5 - 1		-1	-0.0012	-7470A	PPB	01 Jun 2022 12:55:59	U	SMPL	
N3114-01A - 1		17951	4.1799	-7470A	PPB	01 Jun 2022 12:58:14	U	SMPL	
CCV10 - 1	CCV10	21892	5.0977	-7470A	PPB	01 Jun 2022 13:00:29	U	SMPL	
CCB10 - 1	CCB10	117	0.0263	-7470A	PPB	01 Jun 2022 13:02:45	U	SMPL	