



Analytical Summary Report

Analysis Method: 9045D
Parameter: Corrosivity
Run Number: LB120731
BalanceID: WC SC-4

Analyst By : Guerda
Supervisor Review By : jignesh
Slope : 98.9
pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.00	06/17/2022	08:17
2	CAL2	1	Water	NA	NA	20.7	7.00	06/17/2022	08:19
3	CAL3	1	Water	NA	NA	20.8	10.00	06/17/2022	08:21
4	ICV	1	Water	NA	NA	20.3	6.99	06/17/2022	08:23
5	CCV1	1	Water	NA	NA	20.2	2.00	06/17/2022	08:25
6	N3301-04	1	Solid	20.02	20	22.4	8.40	06/17/2022	08:27
7	N3301-04DUP	1	Solid	20.02	20	22.5	8.41	06/17/2022	08:29
8	N3373-14	1	Solid	20.04	20	20.7	11.35	06/17/2022	08:31
9	CCV2	1	Water	NA	NA	20.2	12.00	06/17/2022	08:33

WB 120731

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrsivity n3301 S WorkList ID : 160626 Department : Wet-Chemistry Date : 06-17-2022 08:12:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/09/2022	Solid	N3301-04	Corrosivity	Cool 4 deg C	EART04	WC-SO-20220609		06/09/2022	9045D
06/15/2022	Solid	N3373-14	Corrosivity	Cool 4 deg C	PENN02	D11	DS-7WC	06/15/2022	9045D

Date/Time 06.17.22 08:12
 Raw Sample Received by: JLW
 Raw Sample Relinquished by: JLW

Date/Time 06.17.22 12:00
 Raw Sample Received by: JLW
 Raw Sample Relinquished by: JLW