



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB120759

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	06/20/2022	11:55
2	N3398-07		1	12	100.00	>212.0	>212.0	06/20/2022	12:20
3	N3398-07DUP		1	12	100.00	>212.0	>212.0	06/20/2022	12:45
4	N3398-10		1	15	100.00	>212.0	>212.0	06/20/2022	13:10
5	N3398-12		1	17	100.00	>212.0	>212.0	06/20/2022	13:35
6	N3398-13		1	18	100.00	>212.0	>212.0	06/20/2022	14:40
7	N3398-14		1	16	100.00	>212.0	>212.0	06/20/2022	15:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB | 20759

WorkList Name : fp62022 WorkList ID : 160701 Department : Wet-Chemistry Date : 06-20-2022 11:17:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/22/2022	Water	N3398-07	Flash Point	Cool 4 deg C	PSEG03	D11	ORA-501-03	06/17/2022	1010B
06/22/2022	Water	N3398-10	Flash Point	Cool 4 deg C	PSEG03	D11	ORA-520-521	06/17/2022	1010B
06/22/2022	Water	N3398-12	Flash Point	Cool 4 deg C	PSEG03	D11	ORA-021	06/17/2022	1010B
06/22/2022	Water	N3398-13	Flash Point	Cool 4 deg C	PSEG03	D11	ORA-001	06/17/2022	1010B
06/22/2022	Water	N3398-14	Flash Point	Cool 4 deg C	PSEG03	D11	ORA-519	06/17/2022	1010B

Date/Time 06/20/22 11:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: RA (Ext. Lab)

Date/Time 06/20/22 15:15
 Raw Sample Received by: RA (Ext. Lab)
 Raw Sample Relinquished by: KS (WC)