

LB120761


SFAM01.1

INSTRUMENT 11) : CV 1

Linear

 μ Abs.:

42060



A= 0.0000e+000

B= 2.3994e-004 slope

C= 1.7456e-002 $y = 10765 \cdot x$

Rho= 0.9997496

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01)
0.00	0.000	0.009	0.009	-37	0.000	-37					=
0.05	0.050					185	0				10
0.20	0.200	0.219	0.019	838	0.0 %	838					1
2.50	2.500	2.515	0.015	10408	0.0 %	10408					0
5.00	5.000	5.016	0.016	20834	0.0 %	20834					-2
7.50	7.500	7.332	-0.168	30487	0.0 %	30487					1
10.0	10.000	10.109	0.109	42060	0.0 %	42060					

LB120761
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-37	-		0 SFAM01.1	PPB	6/20/2022 14:02	Std
	0.2 S01	838	-		0.2 SFAM01.1	PPB	6/20/2022 14:05	Std
	2.5 S02	10408	-		2.5 SFAM01.1	PPB	6/20/2022 14:07	Std
	5 S03	20834	-		5 SFAM01.1	PPB	6/20/2022 14:09	Std
	7.5 S04	30487	-		7.5 SFAM01.1	PPB	6/20/2022 14:12	Std
	10 S05	42060	-		10 SFAM01.1	PPB	6/20/2022 14:14	Std
ICV001	ICV001	16778	4.0432	-	SFAM01.1	PPB	6/20/2022 14:17	SMPL
ICB001	ICB001	-66	0.0016	-	SFAM01.1	PPB	6/20/2022 14:20	SMPL
CCV001	CCV001	20779	5.0032	-	SFAM01.1	PPB	6/20/2022 14:22	SMPL
CCB001	CCB001	-66	0.0016	-	SFAM01.1	PPB	6/20/2022 14:24	SMPL
PB145607BL	PBW607	-11	0.0148	-	SFAM01.1	PPB	6/20/2022 14:26	SMPL
N3151-01	JNLQ8	-30	0.0103	-	SFAM01.1	PPB	6/20/2022 14:29	SMPL
N3151-02	JNLQ9	-68	0.0011	-	SFAM01.1	PPB	6/20/2022 14:31	SMPL
N3151-03	JNLR0	-52	0.005	-	SFAM01.1	PPB	6/20/2022 14:33	SMPL
N3151-04	JNLR1	-47	0.0062	-	SFAM01.1	PPB	6/20/2022 14:35	SMPL
N3151-05	JNLR2	-35	0.0091	-	SFAM01.1	PPB	6/20/2022 14:38	SMPL
N3151-06	JNLR3	-40	0.0079	-	SFAM01.1	PPB	6/20/2022 14:40	SMPL
N3151-07	JNLR4	-31	0.01	-	SFAM01.1	PPB	6/20/2022 14:42	SMPL
N3151-08	JNLR5	-38	0.0083	-	SFAM01.1	PPB	6/20/2022 14:45	SMPL
N3151-09	JNLR5D	-56	0.004	-	SFAM01.1	PPB	6/20/2022 14:47	SMPL
N3151-10	JNLR5S	3524	0.863	-	SFAM01.1	PPB	6/20/2022 14:49	SMPL
PB145608BL	PBW608	-18	0.0131	-	SFAM01.1	PPB	6/20/2022 14:51	SMPL
N3145-01	JNLQ0	-3	0.0167	-	SFAM01.1	PPB	6/20/2022 14:54	SMPL
N3145-02	JNLQ1	-48	0.0059	-	SFAM01.1	PPB	6/20/2022 14:56	SMPL
N3145-03	JNLQ2	-57	0.0038	-	SFAM01.1	PPB	6/20/2022 14:58	SMPL
N3145-04	JNLQ3	-41	0.0076	-	SFAM01.1	PPB	6/20/2022 15:00	SMPL
N3145-05	JNLQ4	-33	0.0095	-	SFAM01.1	PPB	6/20/2022 15:03	SMPL
N3145-06	JNLQ5	-40	0.0079	-	SFAM01.1	PPB	6/20/2022 15:05	SMPL
N3145-07	JNLQ6	3	0.0182	-	SFAM01.1	PPB	6/20/2022 15:07	SMPL
N3145-08	JNLQ7	-57	0.0038	-	SFAM01.1	PPB	6/20/2022 15:09	SMPL
CCV002	CCV002	20500	4.9362	-	SFAM01.1	PPB	6/20/2022 15:12	SMPL
CCB002	CCB002	-71	0.0004	-	SFAM01.1	PPB	6/20/2022 15:14	SMPL
N3145-09	JNLQ7D	-39	0.0081	-	SFAM01.1	PPB	6/20/2022 15:16	SMPL
N3145-10	JNLQ7S	3584	0.8774	-	SFAM01.1	PPB	6/20/2022 15:19	SMPL
CCV003	CCV003	20732	4.9919	-	SFAM01.1	PPB	6/20/2022 15:21	SMPL
CCB003	CCB003	-83	-0.0025	-	SFAM01.1	PPB	6/20/2022 15:23	SMPL