



PERCENT SOLID

Supervisor: mohammad
Analyst: ketankumar
Date: 6/22/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 06/21/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 105
Time OUT: 09:02
Out Date: 06/22/2022
Weight Check 1.0g: .00
Weight Check 10g: .00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB120768

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3412-01	NB-7-062022	1	0.91	8.97	9.88	9.43	95.0	
N3412-02	NB-7-062022-EPH-2	2	0.91	8.97	9.88	9.43	95.0	
N3413-01	EO-01-062022	3	0.93	8.88	9.81	9.39	95.3	
N3413-02	EO-01-062022-EPH-2	4	0.93	8.88	9.81	9.39	95.3	
N3420-01	MMC-CON-0620	5	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N3421-01	SB-1-0-2-20220620	6	0.92	8.65	9.57	8.53	88.0	
N3421-02	SB-1-7-9-20220620	7	0.94	8.22	9.16	8.38	90.5	
N3421-03	SB-1-21-23-20220620	8	0.91	8.67	9.58	8.16	83.6	
N3421-04	SB-3-0-2-20220620	9	0.91	8.59	9.5	8.44	87.7	
N3421-05	SB-3-9-11-20220620	10	0.92	9.06	9.98	9.32	92.7	
N3421-06	SB-3-23-25-20220620	11	0.93	8.73	9.66	9.16	94.3	
N3422-03	WC-14A-1A-12	12	0.91	8.81	9.72	8.52	86.4	
N3422-05	WC-14A-1A-12-GR	13	0.92	8.89	9.81	7.83	77.7	
N3422-09	WC-14A-1A-13	14	0.98	8.99	9.97	8.81	87.1	
N3422-11	WC-14A-1A-13-GR	15	0.93	8.67	9.6	8.37	85.8	
N3424-01	331-0621	18	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N3424-03	OH-CON-0621	19	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N3427-01	TR-04-062122	16	0.94	8.84	9.78	9.43	96.0	
N3427-02	TR-04-062122-EPH-2	17	0.94	8.84	9.78	9.43	96.0	
N3428-01	CL-01-062122	20	0.92	8.93	9.85	9.46	95.6	
N3428-02	CL-01-062122-EPH-2	21	0.93	8.89	9.82	9.5	96.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$