

LB120770


7470A

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

36329



A= 0.0000e+000

B= 2.7617e-004

C= -3.3294e-002

Rho= 0.9998993

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.052	0.052	308	0.000	308					11
0.2	0.200	0.221	0.021	921	0.0 %	921					4
2.5	2.500	2.388	-0.112	8769	0.0 %	8769					0
5.0	5.000	5.013	0.013	18271	0.0 %	18271					0
7.5	7.500	7.527	0.027	27375	0.0 %	27375					0
10.0	10.000	10.000	-0.000	36329	0.0 %	36329					0

LB120770 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 21 Jun 2022 10:14:46							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	308	-	0.0000	7470A	PPB	21 Jun 2022 10:41:32	S	Std
0.2 - 1	52.2	921	-	0.2000	7470A	PPB	21 Jun 2022 10:43:49	S	Std
2.5 - 1	52.5	8769	-	2.5000	7470A	PPB	21 Jun 2022 10:46:06	S	Std
5.0 - 1	55	18271	-	5.0000	7470A	PPB	21 Jun 2022 10:48:23	S	Std
7.5 - 1	57.5	27375	-	7.5000	7470A	PPB	21 Jun 2022 10:50:39	S	Std
10.0 - 1	51.8	36329	-	10.0000	7470A	PPB	21 Jun 2022 10:52:56	S	Std
ICV02 - 1	ICV02	15144	4.1490	-	7470A	PPB	21 Jun 2022 10:56:28	U	SMPL
ICB02 - 1	ICB02	270	0.0413	-	7470A	PPB	21 Jun 2022 10:58:42	U	SMPL
CCV04 - 1	CCV04	19177	5.2627	-	7470A	PPB	21 Jun 2022 11:00:57	U	SMPL
CCB04 - 1	CCB04	267	0.0404	-	7470A	PPB	21 Jun 2022 11:03:12	U	SMPL
CRA - 1	CRA	946	0.2280	-	7470A	PPB	21 Jun 2022 11:08:05	U	SMPL
HighStd - 1	HighStd	36496	10.0456	-	7470A	PPB	21 Jun 2022 11:10:20	U	SMPL
ChkStd - 1	ChkStd	26482	7.2801	-	7470A	PPB	21 Jun 2022 11:12:35	U	SMPL
PB145705BL - 1	PBW	-114	-0.0648	-	7470A	PPB	21 Jun 2022 11:14:53	U	SMPL
PB145705BS - 1	LCSW	14229	3.8963	-	7470A	PPB	21 Jun 2022 11:19:46	U	SMPL
N3373-14 - 1	DS-7WC	821	0.1934	-	7470A	PPB	21 Jun 2022 11:22:02	U	SMPL
N3373-14DUP - 1	DS-7WCDUP	690	0.1573	-	7470A	PPB	21 Jun 2022 11:24:18	U	SMPL
N3373-14MS - 1	DS-7WCMS	14895	4.0802	-	7470A	PPB	21 Jun 2022 11:29:00	U	SMPL
N3373-14MSD - 1	DS-7WCMSD	14856	4.0694	-	7470A	PPB	21 Jun 2022 11:31:16	U	SMPL
PB145706BL - 1	PBW	309	0.0520	-	7470A	PPB	21 Jun 2022 11:33:31	U	SMPL
CCV05 - 1	CCV05	19376	5.3177	-	7470A	PPB	21 Jun 2022 11:35:47	U	SMPL
CCB05 - 1	CCB05	245	0.0344	-	7470A	PPB	21 Jun 2022 11:38:02	U	SMPL
PB145706BS - 1	LCSW	15043	4.1211	-	7470A	PPB	21 Jun 2022 11:40:18	U	SMPL
N3378-01 - 1	D2484	948	0.2285	-	7470A	PPB	21 Jun 2022 11:42:33	U	SMPL
N3378-01DUP - 1	D2484DUP	908	0.2175	-	7470A	PPB	21 Jun 2022 11:44:48	U	SMPL
N3378-01MS - 1	D2484MS	14557	3.9868	-	7470A	PPB	21 Jun 2022 11:47:03	U	SMPL
N3378-01MSD - 1	D2484MSD	14546	3.9838	-	7470A	PPB	21 Jun 2022 11:49:19	U	SMPL
PB145707BL - 1	PBW	327	0.0570	-	7470A	PPB	21 Jun 2022 11:51:35	U	SMPL
PB145707BS - 1	LCSW	14900	4.0816	-	7470A	PPB	21 Jun 2022 11:53:51	U	SMPL
N3346-01 - 1	PZ-1S	-102	-0.0615	-	7470A	PPB	21 Jun 2022 11:56:08	U	SMPL
N3346-01DUP - 1	PZ-1SDUP	-54	-0.0482	-	7470A	PPB	21 Jun 2022 11:58:25	U	SMPL
N3346-01MS - 1	PZ-1SMS	16538	4.5339	-	7470A	PPB	21 Jun 2022 12:00:40	U	SMPL
CCV06 - 1	CCV06	19511	5.3550	-	7470A	PPB	21 Jun 2022 12:02:56	U	SMPL
CCB06 - 1	CCB06	266	0.0402	-	7470A	PPB	21 Jun 2022 12:05:12	U	SMPL
N3346-01MSD - 1	PZ-1SMSD	16626	4.5582	-	7470A	PPB	21 Jun 2022 12:07:28	U	SMPL
N3346-02 - 1	PZ-1D	-115	-0.0651	-	7470A	PPB	21 Jun 2022 12:09:43	U	SMPL
N3346-06 - 1	PZ-1S	-71	-0.0529	-	7470A	PPB	21 Jun 2022 12:11:59	U	SMPL
N3346-07 - 1	PZ-1D	-63	-0.0507	-	7470A	PPB	21 Jun 2022 12:14:14	U	SMPL
N3363-15 - 1	NWB-1803	246	0.0346	-	7470A	PPB	21 Jun 2022 12:16:30	U	SMPL
N3368-01 - 1	MW-109	-56	-0.0488	-	7470A	PPB	21 Jun 2022 12:18:46	U	SMPL
N3368-02 - 1	MW-108	-63	-0.0507	-	7470A	PPB	21 Jun 2022 12:21:03	U	SMPL
N3368-03 - 1	MW-110	-72	-0.0532	-	7470A	PPB	21 Jun 2022 12:23:21	U	SMPL
N3368-04 - 1	MW-107	-63	-0.0507	-	7470A	PPB	21 Jun 2022 12:25:38	U	SMPL
N3368-05 - 1	MW-106	-64	-0.0510	-	7470A	PPB	21 Jun 2022 12:27:53	U	SMPL
CCV07 - 1	CCV07	18708	5.1332	-	7470A	PPB	21 Jun 2022 12:30:10	U	SMPL
CCB07 - 1	CCB07	266	0.0402	-	7470A	PPB	21 Jun 2022 12:32:27	U	SMPL
N3368-06 - 1	MW-101	-67	-0.0518	-	7470A	PPB	21 Jun 2022 12:34:44	U	SMPL
N3368-07 - 1	MW-102	-61	-0.0501	-	7470A	PPB	21 Jun 2022 12:36:59	U	SMPL
N3368-08 - 1	MW-104	-81	-0.0557	-	7470A	PPB	21 Jun 2022 12:39:14	U	SMPL
N3368-09 - 1	MW-105	-63	-0.0507	-	7470A	PPB	21 Jun 2022 12:41:30	U	SMPL
N3368-10 - 1	MW-103	-62	-0.0504	-	7470A	PPB	21 Jun 2022 12:43:45	U	SMPL
N3398-10 - 1	ORA-520-521	11011	3.0076	-	7470A	PPB	21 Jun 2022 12:46:02	U	SMPL
N3398-13 - 1	ORA-001	3612	0.9642	-	7470A	PPB	21 Jun 2022 12:48:19	U	SMPL
N3398-14 - 1	ORA-519	251	0.0360	-	7470A	PPB	21 Jun 2022 12:50:35	U	SMPL
PB145645TB - 1	PB145645TB	379	0.0714	-	7470A	PPB	21 Jun 2022 12:55:36	U	SMPL
PB145609TB - 1	PB145609TB	204	0.0230	-	7470A	PPB	21 Jun 2022 12:57:51	U	SMPL
CCV08 - 1	CCV08	18706	5.1327	-	7470A	PPB	21 Jun 2022 13:00:07	U	SMPL
CCB08 - 1	CCB08	249	0.0355	-	7470A	PPB	21 Jun 2022 13:02:22	U	SMPL
N3373-14LX5 - 1		148	0.0076	-	7470A	PPB	21 Jun 2022 13:04:39	U	SMPL
N3373-14A - 1		15068	4.1280	-	7470A	PPB	21 Jun 2022 13:06:55	U	SMPL
N3378-01LX5 - 1		136	0.0043	-	7470A	PPB	21 Jun 2022 13:09:10	U	SMPL
N3378-01A - 1		15573	4.2674	-	7470A	PPB	21 Jun 2022 13:11:26	U	SMPL
N3346-01LX5 - 1		-83	-0.0562	-	7470A	PPB	21 Jun 2022 13:13:41	U	SMPL
N3346-01A - 1		16528	4.5312	-	7470A	PPB	21 Jun 2022 13:15:58	U	SMPL
CCV09 - 1	CCV09	19277	5.2903	-	7470A	PPB	21 Jun 2022 13:18:14	U	SMPL
CCB09 - 1	CCB09	143	0.0062	-	7470A	PPB	21 Jun 2022 13:20:32	U	SMPL