



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120771

Analyst By : Guerda

Supervisor Review By : Iwona

Slope : 99.8

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	06/21/2022	11:00
2	CAL2	1	Water	NA	NA	20.3	7.00	06/21/2022	11:02
3	CAL3	1	Water	NA	NA	20.8	10.00	06/21/2022	11:04
4	ICV	1	Water	NA	NA	20.2	7.01	06/21/2022	11:06
5	CCV1	1	Water	NA	NA	20.2	2.00	06/21/2022	11:08
6	N3398-02	1	Water	NA	NA	21.5	6.94	06/21/2022	11:10
7	N3398-02DUP	1	Water	NA	NA	21.5	6.94	06/21/2022	11:12
8	N3398-03	1	Water	NA	NA	22.0	6.64	06/21/2022	11:14
9	N3398-04	1	Water	NA	NA	21.8	7.10	06/21/2022	11:16
10	N3398-05	1	Water	NA	NA	21.0	7.08	06/21/2022	11:18
11	N3398-06	1	Water	NA	NA	21.8	6.78	06/21/2022	11:20
12	N3398-08	1	Water	NA	NA	21.5	7.26	06/21/2022	11:22
13	N3398-09	1	Water	NA	NA	21.5	7.29	06/21/2022	11:22
14	N3398-12	1	Water	NA	NA	21.6	7.10	06/21/2022	11:24
15	N3398-13	1	Water	NA	NA	21.9	5.47	06/21/2022	11:26
16	CCV2	1	Water	NA	NA	20.2	12.03	06/21/2022	11:30
17	N3398-14	1	Water	NA	NA	21.7	8.48	06/21/2022	11:33
18	N3411-01	1	Water	NA	NA	21.0	7.86	06/21/2022	11:35
19	N3411-02	1	Water	NA	NA	20.9	8.09	06/21/2022	11:37
20	N3411-03	1	Water	NA	NA	20.9	8.35	06/21/2022	11:39

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.2	2.02	06/21/2022	11:40

WORKLIST(Hardcopy Internal Chain)

LB120771

WorkList Name : PH-062122 WorkList ID : 160740 Department : Wet-Chemistry Date : 06-21-2022 10:24:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/17/2022	Water	N3398-02	✓ pH	Cool 4 deg C	PSEG03	D11	ORA-501	06/17/2022	9040C
06/17/2022	Water	N3398-03	C pH	Cool 4 deg C	PSEG03	D11	ORA-502	06/17/2022	9040C
06/17/2022	Water	N3398-04	✓ pH	Cool 4 deg C	PSEG03	D11	ORA-503	06/17/2022	9040C
06/17/2022	Water	N3398-05	✓ pH	Cool 4 deg C	PSEG03	D11	ORA-505	06/17/2022	9040C
06/17/2022	Water	N3398-06	C pH	Cool 4 deg C	PSEG03	D11	ORA-506	06/17/2022	9040C
06/17/2022	Water	N3398-08	C pH	Cool 4 deg C	PSEG03	D11	ORA-520	06/17/2022	9040C
06/17/2022	Water	N3398-09	C pH	Cool 4 deg C	PSEG03	D11	ORA-521	06/17/2022	9040C
06/17/2022	Water	N3398-12	C pH	Cool 4 deg C	PSEG03	D11	ORA-021	06/17/2022	9040C
06/17/2022	Water	N3398-13	✓ pH	Cool 4 deg C	PSEG03	D11	ORA-001	06/17/2022	9040C
06/17/2022	Water	N3398-14	✓ pH	Cool 4 deg C	PSEG03	D11	ORA-519	06/17/2022	9040C
06/17/2022	Water	N3411-01	✓ pH	Cool 4 deg C	STER07	D11	CELL-A	06/17/2022	9040C
06/17/2022	Water	N3411-02	✓ pH	Cool 4 deg C	STER07	D11	CELL-B/C	06/17/2022	9040C
06/17/2022	Water	N3411-03	✓ pH	Cool 4 deg C	STER07	D11	CELL-D	06/17/2022	9040C

Date/Time 6/21/22 1030
 Raw Sample Received by: SPH (UC)
 Raw Sample Relinquished by: CP SM

Date/Time 6/21/22 12:03
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: SPH (UC)