

7471B


LB120778

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53946



A= 0.0000e+000

B= 1.8298e-004

C= -1.5126e-002

Rho= 0.9996655

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	-0.020	-0.020	-24	0.000	-24					-
0.2	0.200	0.189	-0.011	1116	0.0 %	1116					-6
2.5	2.500	2.474	-0.026	13603	0.0 %	13603					-1
5.0	5.000	5.020	0.020	27519	0.0 %	27519					0
7.5	7.500	7.680	0.180	42053	0.0 %	42053					2
10.0	10.000	9.856	-0.144	53946	0.0 %	53946					-1

LB120778 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Jun 2022 14:01:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-24	-	0.0000	7471B	PPB	21 Jun 2022 14:05:07	S	Std
0.2 - 1	50.2	1116	-	0.2000	7471B	PPB	21 Jun 2022 14:07:24	S	Std
2.5 - 1	52.5	13603	-	2.5000	7471B	PPB	21 Jun 2022 14:09:42	S	Std
5.0 - 1	55	27519	-	5.0000	7471B	PPB	21 Jun 2022 14:11:59	S	Std
7.5 - 1	57.5	42053	-	7.5000	7471B	PPB	21 Jun 2022 14:14:16	S	Std
10.0 - 1	57.6	53946	-	10.0000	7471B	PPB	21 Jun 2022 14:19:09	S	Std
ICV03 - 1	ICV03	22043	4.0184	-	7471B	PPB	21 Jun 2022 14:25:21	U	SMPL
ICB03 - 1	ICB03	-90	-0.0316	-	7471B	PPB	21 Jun 2022 14:27:36	U	SMPL
CCV11 - 1	CCV11	26972	4.9203	-	7471B	PPB	21 Jun 2022 14:29:51	U	SMPL
CCB11 - 1	CCB11	-87	-0.0310	-	7471B	PPB	21 Jun 2022 14:32:06	U	SMPL
CRA - 1	CRA	1119	0.1896	-	7471B	PPB	21 Jun 2022 14:34:23	U	SMPL
HighStd - 1	HighStd	55041	10.0565	-	7471B	PPB	21 Jun 2022 14:36:38	U	SMPL
ChkStd - 1	ChkStd	37772	6.8965	-	7471B	PPB	21 Jun 2022 14:38:53	U	SMPL
PB145719BL - 1	PBS	-136	-0.0400	-	7471B	PPB	21 Jun 2022 14:41:11	U	SMPL
PB145719BS - 1	LCSS	22036	4.0171	-	7471B	PPB	21 Jun 2022 14:46:32	U	SMPL
N3371-01 - 1	VNJ-211	2977	0.5296	-	7471B	PPB	21 Jun 2022 14:48:48	U	SMPL
N3373-12 - 1	DS-7D	2879	0.5117	-	7471B	PPB	21 Jun 2022 14:51:06	U	SMPL
N3373-13 - 1	DS-7WC	3503	0.6259	-	7471B	PPB	21 Jun 2022 14:53:23	U	SMPL
N3383-01 - 1	RB-03-(0-3.5)	4191	0.7518	-	7471B	PPB	21 Jun 2022 14:55:40	U	SMPL
N3383-01DUP - 1	RB-03-(0-3.5)DUP	4669	0.8392	-	7471B	PPB	21 Jun 2022 14:57:56	U	SMPL
CCV12 - 1	CCV12	27927	5.0951	-	7471B	PPB	21 Jun 2022 15:00:11	U	SMPL
CCB12 - 1	CCB12	-101	-0.0336	-	7471B	PPB	21 Jun 2022 15:02:27	U	SMPL
N3383-02 - 1	N3383-01MS	0	-0.0151	-	7471B	PPB	21 Jun 2022 15:04:43	U	SMPL
N3383-02 - 1	N3383-01MS	25087	4.5754	-	7471B	PPB	21 Jun 2022 15:06:43	U	SMPL
N3383-03 - 1	N3383-01MSD	25550	4.6601	-	7471B	PPB	21 Jun 2022 15:08:59	U	SMPL
N3383-04 - 1	RB-01-(10-11.5)	630	0.1002	-	7471B	PPB	21 Jun 2022 15:11:16	U	SMPL
N3383-05 - 1	RB-02-(15-16)	9024	1.6361	-	7471B	PPB	21 Jun 2022 15:13:34	U	SMPL
N3383-06 - 1	RB-04-(3-5)	5179	0.9325	-	7471B	PPB	21 Jun 2022 15:15:50	U	SMPL
N3383-07 - 1	RB-04-(10-15)	9	-0.0135	-	7471B	PPB	21 Jun 2022 15:18:06	U	SMPL
N3384-01 - 1	WC-3	10363	1.8811	-	7471B	PPB	21 Jun 2022 15:20:22	U	SMPL
N3390-01 - 1	PT-2-061722	8628	1.5637	-	7471B	PPB	21 Jun 2022 15:22:39	U	SMPL
N3397-01 - 1	OR-2-061722	5511	0.9933	-	7471B	PPB	21 Jun 2022 15:24:56	U	SMPL
N3398-01 - 1	ORA-503-505	4307	0.7730	-	7471B	PPB	21 Jun 2022 15:27:11	U	SMPL
CCV13 - 1	CCV13	28277	5.1591	-	7471B	PPB	21 Jun 2022 15:29:27	U	SMPL
CCB13 - 1	CCB13	-102	-0.0338	-	7471B	PPB	21 Jun 2022 15:31:42	U	SMPL
N3412-01 - 1	NB-7-062022	1678	0.2919	-	7471B	PPB	21 Jun 2022 15:33:59	U	SMPL
N3413-01 - 1	EO-01-062022	1027	0.1728	-	7471B	PPB	21 Jun 2022 15:36:14	U	SMPL
N3414-01 - 1	SU-04-062022	1786	0.3117	-	7471B	PPB	21 Jun 2022 15:38:30	U	SMPL
N3383-01LX5 - 1		696	0.1122	-	7471B	PPB	21 Jun 2022 15:40:45	U	SMPL
N3383-01A - 1		23263	4.2416	-	7471B	PPB	21 Jun 2022 15:43:02	U	SMPL
CCV14 - 1	CCV14	27901	5.0903	-	7471B	PPB	21 Jun 2022 15:45:18	U	SMPL
CCB14 - 1	CCB14	-124	-0.0378	-	7471B	PPB	21 Jun 2022 15:47:36	U	SMPL