


LB120781

SFAM01.1

INSTRUMENT ID: CY1

Linear



A= 0.0000e+000

B= 2.3490e-004 *slope*C= -2.6613e-002 *-1.51e-5*

Rho= 0.9997950

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%d</i>
0.00	0.000	0.062	0.062	377	0.000	377					
0.05	0.050					185	0				27
0.20	0.200	0.254	0.054	1193	0.0 %	1193					-3
2.50	2.500	2.422	-0.078	10425	0.0 %	10425					-3
5.00	5.000	4.871	-0.129	20851	0.0 %	20851					1
7.50	7.500	7.546	0.046	32240	0.0 %	32240					0
10.0	10.000	10.045	0.045	42875	0.0 %	42875					

LB120781
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	377	-		0 SFAM01.1	PPB	6/21/2022 16:00	Std
	0.2 S01	1193	-		0.2 SFAM01.1	PPB	6/21/2022 16:03	Std
	2.5 S02	10425	-		2.5 SFAM01.1	PPB	6/21/2022 16:06	Std
	5 S03	20851	-		5 SFAM01.1	PPB	6/21/2022 16:09	Std
	7.5 S04	32240	-		7.5 SFAM01.1	PPB	6/21/2022 16:11	Std
	10 S05	42875	-		10 SFAM01.1	PPB	6/21/2022 16:13	Std
ICV004	ICV004	17374	4.0545 -		SFAM01.1	PPB	6/21/2022 16:16	SMPL
ICB004	ICB004	337	0.0525 -		SFAM01.1	PPB	6/21/2022 16:18	SMPL
CCV015	CCV015	20929	4.8895 -		SFAM01.1	PPB	6/21/2022 16:21	SMPL
CCB015	CCB015	330	0.0509 -		SFAM01.1	PPB	6/21/2022 16:23	SMPL
PB145618BL	PBW618	-105	-0.0513 -		SFAM01.1	PPB	6/21/2022 16:28	SMPL
N3230-01	GBGY9	701	0.138 -		SFAM01.1	PPB	6/21/2022 16:30	SMPL
N3230-02	GBGZ0	679	0.1329 -		SFAM01.1	PPB	6/21/2022 16:32	SMPL
N3230-03	GBGZ0D	694	0.1364 -		SFAM01.1	PPB	6/21/2022 16:34	SMPL
N3230-04	GBGZ0S	4345	0.994 -		SFAM01.1	PPB	6/21/2022 16:42	SMPL
N3230-05	GBGZ1	674	0.1317 -		SFAM01.1	PPB	6/21/2022 16:44	SMPL
PB145619BL	PBW619	345	0.0544 -		SFAM01.1	PPB	6/21/2022 16:47	SMPL
N3231-09	GBGX4	601	0.1146 -		SFAM01.1	PPB	6/21/2022 16:49	SMPL
N3231-10	GBGX4D	451	0.0793 -		SFAM01.1	PPB	6/21/2022 16:51	SMPL
N3231-11	GBGX4S	4512	1.0332 -		SFAM01.1	PPB	6/21/2022 16:53	SMPL
N3231-12	GBGX5	347	0.0549 -		SFAM01.1	PPB	6/21/2022 16:56	SMPL
N3231-13	GBGX6	382	0.0631 -		SFAM01.1	PPB	6/21/2022 16:58	SMPL
N3231-16	GBGY2	21658	5.0608 -		SFAM01.1	PPB	6/21/2022 17:00	SMPL
PB145620BL	PBW620	-34	-0.0346 -		SFAM01.1	PPB	6/21/2022 17:02	SMPL
N3232-01	GBGX7	355	0.0568 -		SFAM01.1	PPB	6/21/2022 17:05	SMPL
N3232-02	GBGX7D	360	0.0579 -		SFAM01.1	PPB	6/21/2022 17:07	SMPL
N3232-03	GBGX7S	4424	1.0126 -		SFAM01.1	PPB	6/21/2022 17:09	SMPL
N3232-04	GBGX8	340	0.0533 -		SFAM01.1	PPB	6/21/2022 17:11	SMPL
N3232-05	GBGX9	366	0.0594 -		SFAM01.1	PPB	6/21/2022 17:14	SMPL
PB145621BL	PBW621	42	-0.0167 -		SFAM01.1	PPB	6/21/2022 17:16	SMPL
CCV016	CCV016	21261	4.9675 -		SFAM01.1	PPB	6/21/2022 17:18	SMPL
CCB016	CCB016	347	0.0549 -		SFAM01.1	PPB	6/21/2022 17:21	SMPL
N3229-06	GBGY6	960	0.1989 -		SFAM01.1	PPB	6/21/2022 17:23	SMPL
N3229-07	GBGY7	933	0.1925 -		SFAM01.1	PPB	6/21/2022 17:25	SMPL
N3229-08	GBGY7D	635	0.1225 -		SFAM01.1	PPB	6/21/2022 17:27	SMPL
N3229-09	GBGY7S	4429	1.0137 -		SFAM01.1	PPB	6/21/2022 17:30	SMPL
N3229-10	GBGY8	880	0.1801 -		SFAM01.1	PPB	6/21/2022 17:32	SMPL
N3229-13	GBGZ4	21762	5.0852 -		SFAM01.1	PPB	6/21/2022 17:34	SMPL
CCV017	CCV017	21073	4.9234 -		SFAM01.1	PPB	6/21/2022 17:45	SMPL
CCB017	CCB017	324	0.0495 -		SFAM01.1	PPB	6/21/2022 17:48	SMPL