


LB120788

7471B

INSTRUMENT 111: CV1

 μ Abs.:

53860



Linear

A= 0.0000e+000

B= 1.8418e-004

C= 6.5903e-002

Rho= 0.9997445

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/0/1
0.0	0.000	0.066	0.066	2	0.000	2					-
0.2	0.200	0.247	0.047	981	0.0 %	981					24
2.5	2.500	2.378	-0.122	12552	0.0 %	12552					-5
5.0	5.000	4.912	-0.088	26313	0.0 %	26313					-2
7.5	7.500	7.611	0.111	40966	0.0 %	40966					1
10.0	10.000	9.986	-0.014	53860	0.0 %	53860					0

LB120788 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Jun 2022 13:53:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	2	-	0.0000	7471B	PPB	22 Jun 2022 14:14:17	S	Std
0.2 - 1	50.2	981	-	0.2000	7471B	PPB	22 Jun 2022 14:16:33	S	Std
2.5 - 1	52.5	12552	-	2.5000	7471B	PPB	22 Jun 2022 14:18:51	S	Std
5.0 - 1	55	26313	-	5.0000	7471B	PPB	22 Jun 2022 14:21:08	S	Std
7.5 - 1	57.5	40966	-	7.5000	7471B	PPB	22 Jun 2022 14:23:24	S	Std
10.0 - 1	516	53860	-	10.0000	7471B	PPB	22 Jun 2022 14:25:41	S	Std
ICV05 - 1	ICV05	21057	3.9442	-	7471B	PPB	22 Jun 2022 14:28:26	U	SMPL
ICB05 - 1	ICB05	-100	0.0475	-	7471B	PPB	22 Jun 2022 14:30:40	U	SMPL
CCV18 - 1	CCV18	26429	4.9337	-	7471B	PPB	22 Jun 2022 14:32:57	U	SMPL
CCB18 - 1	CCB18	-84	0.0504	-	7471B	PPB	22 Jun 2022 14:35:11	U	SMPL
CRA - 1	CRA	877	0.2274	-	7471B	PPB	22 Jun 2022 14:37:28	U	SMPL
HighStd - 1	HighStd	54052	10.0213	-	7471B	PPB	22 Jun 2022 14:39:43	U	SMPL
ChkStd - 1	ChkStd	37742	7.0173	-	7471B	PPB	22 Jun 2022 14:41:58	U	SMPL
PB145747BL - 1	PBS	-160	0.0364	-	7471B	PPB	22 Jun 2022 14:44:16	U	SMPL
PB145747BS - 1	LCSS	21173	3.9656	-	7471B	PPB	22 Jun 2022 14:49:10	U	SMPL
N3421-01 - 1	SB-1-0-2-20220620	333	0.1272	-	7471B	PPB	22 Jun 2022 14:51:25	U	SMPL
N3421-02 - 1	SB-1-7-9-20220620	473	0.1530	-	7471B	PPB	22 Jun 2022 14:53:44	U	SMPL
N3421-04 - 1	SB-3-0-2-20220620	7014	1.3578	-	7471B	PPB	22 Jun 2022 14:56:00	U	SMPL
N3421-05 - 1	SB-3-9-11-20220620	4211	0.8415	-	7471B	PPB	22 Jun 2022 14:58:18	U	SMPL
N3422-01 - 1	WC-14A-1A-12	143059	26.4149	-	7471B	PPB	22 Jun 2022 15:00:33	U	SMPL
CCV19 - 1	CCV19	26443	4.9362	-	7471B	PPB	22 Jun 2022 15:02:48	U	SMPL
CCB19 - 1	CCB19	-147	0.0388	-	7471B	PPB	22 Jun 2022 15:05:06	U	SMPL
N3422-07 - 1	WC-14A-1A-13	30609	5.7035	-	7471B	PPB	22 Jun 2022 15:07:23	U	SMPL
N3427-01 - 1	TR-04-062122	1165	0.2805	-	7471B	PPB	22 Jun 2022 15:09:39	U	SMPL
N3428-01 - 1	CL-01-062122	3484	0.7076	-	7471B	PPB	22 Jun 2022 15:11:56	U	SMPL
N3428-01DUP - 1	CL-01-062122DUP	3698	0.7470	-	7471B	PPB	22 Jun 2022 15:14:12	U	SMPL
N3428-01MS - 1	CL-01-062122MS	26185	4.8887	-	7471B	PPB	22 Jun 2022 15:16:28	U	SMPL
N3428-01MSD - 1	CL-01-062122MSD	25487	4.7602	-	7471B	PPB	22 Jun 2022 15:18:44	U	SMPL
N3428-01LX5 - 1		693	0.1935	-	7471B	PPB	22 Jun 2022 15:21:02	U	SMPL
N3428-01A - 1		24769	4.6279	-	7471B	PPB	22 Jun 2022 15:23:20	U	SMPL
N3422-01DLX5 - 1		30601	5.7021	-	7471B	PPB	22 Jun 2022 15:25:38	U	SMPL
CCV20 - 1	CCV20	26347	4.9186	-	7471B	PPB	22 Jun 2022 15:27:54	U	SMPL
CCB20 - 1	CCB20	-107	0.0462	-	7471B	PPB	22 Jun 2022 15:30:11	U	SMPL