


LB120798

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.9609e-004

C= 1.0635e-002

Rho= 0.9998139

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.005	-0.005	-53	0.000	-53					-
0.2	0.200	0.181	-0.019	575	0.0 %	575					-10
2.5	2.500	2.585	0.085	8693	0.0 %	8693					3
5.0	5.000	4.882	-0.118	16453	0.0 %	16453					-2
7.5	7.500	7.589	0.089	25596	0.0 %	25596					1
10.0	10.000	9.968	-0.032	33630	0.0 %	33630					0

LB120798 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 23 Jun 2022 11:09:05					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-53	-	0.0000	7470A	PPB	23 Jun 2022 11:11:07	S	Std
0.2 - 1	50.2	575	-	0.2000	7470A	PPB	23 Jun 2022 11:13:23	S	Std
2.5 - 1	52.5	8693	-	2.5000	7470A	PPB	23 Jun 2022 11:15:40	S	Std
5.0 - 1	55	16453	-	5.0000	7470A	PPB	23 Jun 2022 11:20:28	S	Std
7.5 - 1	57.5	25596	-	7.5000	7470A	PPB	23 Jun 2022 11:25:19	S	Std
10.0 - 1	59.1	33630	-	10.0000	7470A	PPB	23 Jun 2022 11:30:08	S	Std
ICV06 - 1	ICV06	14539	4.3155	-	7470A	PPB	23 Jun 2022 11:32:52	U	SMPL
ICB06 - 1	ICB06	-119	-0.0246	-	7470A	PPB	23 Jun 2022 11:35:07	U	SMPL
CCV21 - 1	CCV21	16394	4.8647	-	7470A	PPB	23 Jun 2022 11:37:22	U	SMPL
CCB21 - 1	CCB21	-102	-0.0196	-	7470A	PPB	23 Jun 2022 11:39:37	U	SMPL
CRA - 1	CRA	591	0.1856	-	7470A	PPB	23 Jun 2022 11:41:52	U	SMPL
HighStd - 1	HighStd	33537	9.9406	-	7470A	PPB	23 Jun 2022 11:44:07	U	SMPL
ChkStd - 1	ChkStd	24209	7.1787	-	7470A	PPB	23 Jun 2022 11:48:54	U	SMPL
PB145752BL - 1	PBW	-77	-0.0122	-	7470A	PPB	23 Jun 2022 11:51:09	U	SMPL
PB145752BS - 1	LCSW	14365	4.2640	-	7470A	PPB	23 Jun 2022 11:55:57	U	SMPL
N3416-01 - 1	TOTE-1	2390	0.7183	-	7470A	PPB	23 Jun 2022 11:58:13	U	SMPL
N3422-02 - 1	WC-14A-1A-12	-41	-0.0015	-	7470A	PPB	23 Jun 2022 12:00:30	U	SMPL
N3422-02DUP - 1	WC-14A-1A-12DUP	-2	0.0100	-	7470A	PPB	23 Jun 2022 12:02:47	U	SMPL
N3422-02MS - 1	WC-14A-1A-12MS	14395	4.2728	-	7470A	PPB	23 Jun 2022 12:05:04	U	SMPL
N3422-02MSD - 1	WC-14A-1A-12MSD	14568	4.3241	-	7470A	PPB	23 Jun 2022 12:09:57	U	SMPL
CCV22 - 1	CCV22	16532	4.9056	-	7470A	PPB	23 Jun 2022 12:12:12	U	SMPL
CCB22 - 1	CCB22	-114	-0.0231	-	7470A	PPB	23 Jun 2022 12:14:27	U	SMPL
N3422-08 - 1	WC-14A-1A-13	-32	0.0012	-	7470A	PPB	23 Jun 2022 12:16:42	U	SMPL
N3422-08DUP - 1	WC-14A-1A-13DUP	-13	0.0068	-	7470A	PPB	23 Jun 2022 12:18:58	U	SMPL
N3422-08MS - 1	WC-14A-1A-13MS	15006	4.4537	-	7470A	PPB	23 Jun 2022 12:21:13	U	SMPL
N3422-08MSD - 1	WC-14A-1A-13MSD	14777	4.3859	-	7470A	PPB	23 Jun 2022 12:23:29	U	SMPL
PB145753BL - 1	PBW	-94	-0.0172	-	7470A	PPB	23 Jun 2022 12:25:45	U	SMPL
PB145753BS - 1	LCSW	13677	4.0602	-	7470A	PPB	23 Jun 2022 12:28:02	U	SMPL
N3384-03 - 1	WC-4	-49	-0.0039	-	7470A	PPB	23 Jun 2022 12:30:18	U	SMPL
N3384-03DUP - 1	WC-4DUP	-4	0.0095	-	7470A	PPB	23 Jun 2022 12:32:34	U	SMPL
N3384-03MS - 1	WC-4MS	14798	4.3922	-	7470A	PPB	23 Jun 2022 12:45:20	U	SMPL
N3384-03MSD - 1	WC-4MSD	15305	4.5423	-	7470A	PPB	23 Jun 2022 12:47:36	U	SMPL
CCV23 - 1	CCV23	16471	4.8875	-	7470A	PPB	23 Jun 2022 12:49:51	U	SMPL
CCB23 - 1	CCB23	-110	-0.0219	-	7470A	PPB	23 Jun 2022 12:52:06	U	SMPL
N3416-05 - 1	TOTE-1	25499	7.5606	-	7470A	PPB	23 Jun 2022 12:54:23	U	SMPL
PB145717TB - 1	PB145717TB	-159	-0.0364	-	7470A	PPB	23 Jun 2022 12:56:38	U	SMPL
PB145718TB - 1	PB145718TB	-54	-0.0054	-	7470A	PPB	23 Jun 2022 12:58:55	U	SMPL
N3422-02LX5 - 1		-78	-0.0125	-	7470A	PPB	23 Jun 2022 13:01:10	U	SMPL
N3422-02A - 1		13804	4.0978	-	7470A	PPB	23 Jun 2022 13:03:26	U	SMPL
N3422-08LX5 - 1		-124	-0.0261	-	7470A	PPB	23 Jun 2022 13:05:42	U	SMPL
N3422-08A - 1		14375	4.2669	-	7470A	PPB	23 Jun 2022 13:07:58	U	SMPL
N3384-03LX5 - 1		-50	-0.0042	-	7470A	PPB	23 Jun 2022 13:10:15	U	SMPL
N3384-03A - 1		15204	4.5124	-	7470A	PPB	23 Jun 2022 13:16:22	U	SMPL
CCV24 - 1	CCV24	16515	4.9005	-	7470A	PPB	23 Jun 2022 13:18:37	U	SMPL
CCB24 - 1	CCB24	-102	-0.0196	-	7470A	PPB	23 Jun 2022 13:20:53	U	SMPL